

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017 & KS Q ISO/IEC 17025:2017

HCT Co., Ltd.  
 74, 578-gil, Seoicheon-ro, Majang-myeon, Icheon-si, Gyungki-do, Republic of Korea  
 Phone : 82-31-645-6300, Fax : 82-31-645-6969, e-mail : hct.qm@hct.co.kr

CALIBRATION

Valid To : Jan. 07, 2026.

Accreditation No : KC00-000

In recognition of the successful completion of the KOLAS evaluation process,  
 accreditation is granted to this laboratory to perform the following calibrations

Field Code	Item of Calibration	on-site	Field Code	Item of Calibration	on-site	Field Code	Item of Calibration	on-site
102.	Linear dimension		10405	Optical parallels	N	202. Force		
			10406	Parallel blocks	N	20202	Force measuring devices	N
10201	Balls	N	10407	Precision surface plates	Y	20203	Tension/compression testing machines	Y
10203	Electrical/mechanical comparators	Y	10409	Roundness measurement instruments	Y		20204	Push-pull gauges
10206	Dial/cylinder gauge testers	N	10412	Straight edges	N	203. Torque		
10207	Doctor blades	N	10413	Straight rules	N			
10209	End bars	N	10415	Test bars	N	20302	Torque measuring devices	N
10210	Extensometers, linear displacement transducers	Y	10416	Spherometers	N	20303	Torque wrenches/drivers	N
10211	Filler gauges	N	105. Complex geometry			204. Pressure		
10212	Film applicators	N	10503	Contact coordinate measuring machines	Y	20401	Altimeters	Y
10213	Gap gauges	N		10504	Non-contact coordinate measuring machines	Y	20402	Manometers
10214	Gauge blocks, by comparison	N	10505	Gauge block accessories	N	20406	Absolute pressure gauges	Y
10216	Height gauges/measuring machines	Y		10511	Measuring microscopes, profile projectors	Y	20408	Compound pressure gauges
10219	Linear scales	N	10512	Micro measuring	Y	20409	Differential pressure	Y
10220	Standard measuring machines	Y		10517	Stylus type roughness	Y	20411	Gauge pressure gauges
10223	Electronic micrometers	N	10525	Thread plug gauges	N	20412	Pressure transducers/transmitters	Y
10224	Height micrometers, Riser blocks	N		10529	V-blocks, box blocks	N	20413	Dial type vacuum gauges
10225	Laser scan micrometers	Y	106. Various dimensional			20414	Water Depth meters	Y
10227	Standard tape rules, peripheral gauges	N				206. Volume		
10228	Cylindrical plug/pin gauges, thread measuring wire gauges	N	10601	Inside/outside/gear tooth calipers, caliper gauges	Y	20601	Volumetric glasswares	N
						20602	Pycnometers	N
10229	Radius gauges	N	10603	Cylinder/bore gauges	Y	20605	Concrete air content	N
10230	Cylindrical ring gauges	N	10604	Depth gauges, depth micrometers	Y	20606	Piston type volume meters	N
10231	Step blocks	N				207. Density		
10232	Step gauges	N	10605	Dial/digital gauges	Y	20702	Liquid density meters	N
10233	Taper thickness gauges	N	10609	Micro indicators, Test indicators	Y	20704	Salinity meters	N
10234	Ultrasonic thickness gauges	Y		10610	Micrometer heads	N	20705	Sucrose meters
10235	Ultrasonic/coating thickness specimens	N	10611	3-point micrometers	Y	20707	Chloride meters	N
10236	Coating thickness testers	Y	10612	Inside micrometers	Y	208. Viscosity		
10237	Torque arms	N	10613	Outside micrometers	Y	20802	Dynamic viscometers; rotational, etc.	N
10238	Width measuring specimens	N	10615	Particle counters	Y			
103.	Angle		10617	Standard sieves	N	209. Fluid flow		
			10620	Welding gauges	N	20901	Anemometers; hot-wire	N
10304	Bevel protractors	N	10622	Particle dilution Systems	Y	20902	Anemometers; pitot tube,	N
10311	Plate/square/electric levels	N	201. Mass			20908	Gas flowmeters; differential pressure	Y
10317	Sine bars, plates, tables, centers	N	20105	Counter beam balances	Y	20909	Liquid flowmeters; differential pressure	N
10318	Squareness testers, right angle testers	N	20106	Dial platform scale	Y			
10319	Cylindrical squares	N	20108	Direct reading balances	Y	20910	Liquid flowmeters; electromagnetic	N
10320	Precision squares	N	20109	Electric balances	Y			
104.	Form		20113	Spring scale balances	Y	20911	Gas flowmeters; thermal mass, etc.	Y
			20114	Trip balances	N			
			10401	Form testers	Y	20116	Weights	Y
10404	Optical flats	N				Coriolis, etc.		

Field Code	Item of Calibration	on-site	Field Code	Item of Calibration	on-site	Field Code	Item of Calibration	on-site
20914	Gas flowmeters; positive displacement	Y	40217	Impedance bridges/LCR	Y	40606	Attenuator calibrators	Y
						40607	RF power meter calibrators	Y
20915	Liquid flowmeters; positive displacement	N	403. AC voltage, current & power			40608	EMC transducers; current probes, absorbing clamps, etc.	Y
			40301	AC ammeters	Y			
20916	Gas flowmeters; turbine	Y	40302	Clamp ammeters/voltmeters	Y			
20917	Liquid flowmeters; turbine	N	40303	AC voltage/current calibrators	Y	40610	Coaxial directional coupler/splitters	Y
20918	Gas flowmeters; ultrasonic	Y						
20919	Liquid flowmeters;	N	40304	Wattmeter calibrators	Y	40611	Waveguide directional couplers	N
20920	Gas flowmeters; variable area	Y	40305	AC current shunts	Y			
			40310	Power factor meters	Y	40613	Electrostatic discharge generators	N
20921	Liquid flowmeters; variable	N	40311	AC power meters	Y			
20922	Gas flowmeters; vortex	Y	40312	AC power supplies	Y	40614	EMC receivers	Y
20923	Liquid flowmeters; vortex	N	40313	Puncture/safety testers	Y	40615	RF filters	Y
20925	Anemometers; vane, etc.	N	40314	Power recorders	Y	40616	RF impedance meters	Y
			40318	AC voltmeters	Y	40617	RF impulse generators	Y
301. Time & frequency						40618	Line impedance networks; LISN, CDN, ISN, etc.	Y
30102	Frequency standards	N	404. Other DC & LF measurements					
30103	General frequency sources	Y	40401	LF amplifiers	Y			
30104	Frequency meters/counters	Y	40402	DC/LF attenuators	Y	40619	Coaxial standard mismatches	Y
30105	Time interval sources	Y	40403	Multimeter calibrators	Y	40621	Mobile communication test sets	Y
30106	Time interval meters/stop watches/timers	Y	40404	Oscilloscope calibrators	Y			
			40406	Video signal generators	Y	40622	Modulation meters	Y
			40407	Audio distortion analyzers /meters	Y	40623	Network analyzers	Y
302. Velocity & revolution						40624	Noise figure meters	Y
30201	Standard RPM generators	Y	40408	LF filters	Y	40625	Noise generators	Y
30202	Contact type tachometers	Y	40409	LF/audio signal analyzers	Y	40626	Noise impulse simulators	Y
30203	Photo tachometers /stroboscopes	Y	40410	Line frequency meters	Y	40635	RF power meters	Y
			40411	Function generators	Y	40636	Diode power sensors	Y
30205	Wow-flutter generators	Y	40412	Genescopes	Y	40637	Thermocouple power sensors	Y
30206	Wow-flutter meters	Y	40413	AC/DC high voltage	Y	40638	Pulse generators	Y
			40416	Leakage current testers	Y	40639	Radar test sets	Y
401. DC voltage& current			40417	Electronic AC/DC loads	Y	40640	RF signal generators	Y
40101	DC ammeters	Y	40419	Analogue/digital	Y	40641	RF spectrum analyzers	Y
40102	Transconductance amplifiers	Y	40420	Noise meters	Y	40642	RF speed guns	Y
40103	DC voltage/current calibrators	Y	40421	Oscilloscopes	Y	40643	Surge generators	Y
			40422	LF phase meters	Y	40645	RF terminations	Y
40104	Electrical temperature calibrators	Y	40423	Random wave generators	Y	40646	Coaxial thermistor mounts	Y
			40424	Voltage/current recorders	Y	40648	Transmission trouble tester	Y
40105	DC current shunts	Y	40425	Relay test sets	Y	40650	RF voltmeters	Y
40106	Galvanometers/null detectors	Y	40426	LF signal generators	Y	40651	Vector voltmeters	Y
40108	DC power supplies	Y	40427	LF spectrum analyzers	Y	40652	Field strength meters	Y
40110	DC voltage dividers	N	40429	Sweep generators	Y	40653	AM/FM test sources	Y
40111	DC voltage standards	Y	40432	Transistor curve tracers	Y	40654	Dip simulators	Y
40112	DC voltmeters	Y	40433	Waveform analyzers	Y			
40113	Static/Ionic voltmeters	N	40434	AC/DC high voltage	Y	407. Field strength & antenna		
			40435	AC/DC high voltage probes	Y	40701	Microwave leakage monitors	N
402. Resistance, capacitance inductance			40436	Logic analyzers	Y	40702	Probes	N
40201	Capacitance bridges /indicators	Y	40437	Telephone testers	Y	40703	Dipole antennas	N
			40438	Video signal analyzers	Y	40704	Loop antennas	N
40202	Decade capacitors	Y				40705	Monopole antennas	N
40204	Standard capacitors	Y	405. Low frequency electric & magnetic field			40707	Horn antennas	N
40205	Earth testers	Y	40503	Flux meters	N			
40208	Inductors	Y	40504	Flux sources	N	501. Contact thermometry		
40210	Insulation testers	Y	40508	Magnetometers	N	50101	Temperature generators; ovens, furnaces, isothermal liquid baths, ice-point baths,	Y
40211	Q-meters	Y	40510	Reference/standard magnets	N			
40213	Resistance bridges & similar instruments	Y	406. Radio frequency measurement					
40214	Resistance meters	Y	40601	RF amplifiers	Y	50102	dry-block calibrators	Y
40215	Resistors	Y	40602	Coaxial attenuators	Y		Temperature indicators /recorders/controllers, temperature calibrators	
40216	Electrical conductivity meters	N	40603	Waveguide attenuators	N			
			40605	Burst pulse generators	Y			

Field Code	Item of Calibration	on-site	Field Code	Item of Calibration	on-site	Field Code	Item of Calibration	on-site
50103	Glass thermometers; liquid-in-glass, Beckmann	N	701. Photometry			901. Chemical analysis		
						90101	Breath Alcohol Analyzer	N
50104	Resistance thermometers; SPRT, IPRT, thermistors, etc.	Y	70101	Illuminance meters	Y	90102	Environmental air quality monitoring instruments	N
			70102	Luminance meters	Y			
			70103	Total luminous flux meters	Y	90103	Gas analyzers	N
50105	Thermal expansion thermometers; bimetal, gas or liquid type	Y	70104	Luminous intensity meters	Y	90104	Exhaust gas test instrument	N
			702. Property of detectors & sources					
50106	Thermocouples; noble metal, base metal, pure metal, special type, etc.	Y	70202	Color temperature meters	Y			
			70203	Color temperature standard lamps	Y			
			70204	Colorimeters; source color	Y			
50107	Temperature transducers	Y	70209	Total luminous flux standard lamps	N			
502. Non contact thermometry			70213	Display color analyzers; luminance, chromaticity, white balance, etc.	N			
50204	Standard radiation thermometers	N	70214	Luminous intensity standard lamps	N			
50205	Thermal image apparatus	N	70215	Spectral irradiance standard lamps	N			
50206	Blackbody furnaces	N	70216	Total spectral radiant flux standard lamps	N			
503. Humidity			70217	Luminance standard sources	N			
50301	Dew-point hygrometers; chilled mirror, alumina thin film, etc.	N	70218	Spectral radiance standard sources	N			
50302	Relative humidity hygrometers; polymer thin film, hair, etc.	Y	70219	UV irradiance meters	N			
50304	Temperature humidity recorders; hygrothermograph, etc.	N	70220	Spectral irradiance meters	N			
			70221	Total spectral radiant flux meters	Y			
50305	Transducers; dew-point/relative humidity	Y	70222	Spectral radiance meters	N			
			70223	Spectral radiant intensity meters	N			
50306	Humidity generators; two-pressure, two-temperature, flow mixing humidity generator, constant temperature and humidity chamber, etc.	Y	703. Property of materials					
			70301	Colorimeters; object color	Y			
			70321	Reflectance meters	N			
			704. Fiber optics					
601. Sound in air			70402	Broadband light sources	N			
			70410	Optical attenuators	N			
60102	Sound calibrators	N	70413	Optical loss testers	N			
60104	Microphones	N	70415	Optical multimeters	N			
60106	Sound level meters	Y	70417	Optical spectrum analyzers	N			
603. Vibration			70418	Optical time domain reflectometers; OTDR	N			
60301	Vibration calibrators	Y	70430	ASE light sources	N			
60302	Vibration transducers	N	70433	Optical power stabilized lasers and LDs	N			
60303	Vibration measuring instruments	N						

Note

1. This laboratory provides calibration services in permanent standard laboratory and at on-site.
2. Laboratory conducts on-site calibration should meet requirements of KOLAS-SR-007.
3. On-site calibration is allowed to items with marking 'Y', not allowed to items with marking 'N'.
4. Measurement uncertainty normally is quoted as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k=2$ . It expresses the lowest uncertainty of measurement that can be provided by accredited calibration laboratories in normal conditions.
5. Due to the calibration environment such as reference standards or customers' facilities, it is note that uncertainty of measurement on a calibration certificate may be expressed larger than measurement uncertainty on scope of accreditation in general.

102. Linear dimension

Measured Quantity Instrument or Gauge	Field Code	Range	CMC (The Confidence Level is about 95 %)	Comments
Balls	10201	(0 ~ 100) mm	$\sqrt{0.27^2+(0.010 \times I_0)^2}$ μm (unit of $I_0$ : mm)	Gauge blocks, Standard measuring machines /HCT-CS-223-10201
Electrical/mechanical comparators	10203	(0 ~ 5) mm	0.12 μm	Gauge blocks /HCT-CS-334-10203
Dual/cylinder gauge testers	10206	(0 ~ 25) mm (25 ~ 100) mm	0.21 μm 0.25 μm	Gauge blocks, Electronic micrometers /HCT-CS-001-10206
Doctor blades	10207	(0 ~ 10) mm	2.3 μm	Precision surface plates, Precision surface plates, Electronic micrometers /HCT-CS-335-10207
End bars	10209	(0 ~ 500) mm (500 ~ 1 000) mm	$\sqrt{0.13^2+(0.002 \times I_0)^2}$ μm $\sqrt{0.16^2+(0.002 \times I_0)^2}$ μm (unit of $I_0$ : mm)	Gauge blocks, Electronic micrometers /HCT-CS-183-10209
Extensometers, linear displacement transducers	10210	(0 ~ 50) mm (50 ~ 100) mm (100 ~ 1 000) mm	$\sqrt{0.14^2+(0.002 \times I_0)^2}$ μm $\sqrt{0.78^2+(0.002 \times I_0)^2}$ μm $\sqrt{7.8^2+(0.002 \times I_0)^2}$ μm (unit of $I_0$ : mm)	Gauge blocks /HCT-CS-184-10210
Filter gauges	10211	(0.01 ~ 5) mm	0.33 μm	Standard measuring machines /HCT-CS-002-10211
Film applicators	10212	(0 ~ 10) mm	2.3 μm	Height micrometers Precision surface plates Electronic micrometers /HCT-CS-336-10212
Gap gauges	10213	(1 ~ 300) mm	$\sqrt{3.1^2+(0.005 \times I_0)^2}$ μm (unit of $I_0$ : mm)	Height micrometers, Electronic micrometers /HCT-CS-003-10213
Gap gauges/measuring machines	10214	(0.5 ~ 100) mm	$\sqrt{71^2+(1.3 \times I_0)^2}$ nm (unit of $I_0$ : mm)	Gauge block comparators, Gauge blocks /HCT-CS-254-10214
Height gauges/measuring machines	10216	(0 ~ 1 000) mm	$\sqrt{1.2^2+(0.003 \times I_0)^2}$ μm (unit of $I_0$ : mm)	Gauge blocks /HCT-CS-005-10216
Linear scales	10219	(0 ~ 2 000) mm	$\sqrt{1.5^2+(0.0014 \times I_0)^2}$ μm (unit of $I_0$ : mm)	Lasor interferometers /HCT-CS-325-10219
Standard measuring machines	10220	(0 ~ 500) mm	$\sqrt{0.25^2+(0.0021 \times I_0)^2}$ μm (unit of $I_0$ : mm)	Gauge blocks, Long gauge blocks /HCT-CS-224-10220
Electronic micrometers	10223	(0 ~ 0.02) mm (0.02 ~ 0.2) mm (0.2 ~ 2) mm	0.08 μm 0.16 μm 0.76 μm	Gauge blocks /HCT-CS-006-10223
Height micrometers, Riser blocks Height micrometers Blocks Head Riser blocks	10224	(0 ~ 610) mm (0 ~ 30) mm (0 ~ 600) mm	$\sqrt{1.2^2+(0.003 \times I_0)^2}$ μm (unit of $I_0$ : mm) $\sqrt{1.3^2+(0.003 \times I_0)^2}$ μm (unit of $I_0$ : mm) $\sqrt{1.2^2+(0.003 \times I_0)^2}$ μm (unit of $I_0$ : mm)	Gauge blocks, Electronic micrometers /HCT-CS-007-10224
Laser scan micrometers	10225	(0.1 ~ 55) mm	0.56 μm	Pin gauges /HCT-CS-282-10225

102. Linear dimension

Measured Quantity Instrument or Gauge	Field Code	Range	CMC (The Confidence Level is about 95 %)	Comments
Standard tape rules, peripheral gauges	10227	(0 ~ 13) m (13 ~ 26) m (26 ~ 39) m (39 ~ 50) m	$\sqrt{0.26^2+(0.002 \times I_0)^2}$ mm $\sqrt{0.46^2+(0.002 \times I_0)^2}$ mm $\sqrt{0.67^2+(0.002 \times I_0)^2}$ mm $\sqrt{0.88^2+(0.002 \times I_0)^2}$ mm (unit of $I_0$ : mm)	Lasor interferometers, Tape measure calibration system /HCT-CS-241-10227
Cylindrical plug/pin gauges, thread measuring wire gauges	10228	(0.1 ~ 100) mm	$\sqrt{0.29^2+(0.003 \times I_0)^2}$ μm (unit of $I_0$ : mm)	Standard measuring machines /HCT-CS-008-10228
Radius gauges	10229	(0 ~ 50) mm	3.0 μm	Profile projectors /HCT-CS-225-10229
Cylindrical ring gauges	10230	(1 ~ 100) mm	$\sqrt{0.59^2+(0.003 \times I_0)^2}$ μm (unit of $I_0$ : mm)	Standard measuring machines Standard ring gauges /HCT-CS-226-10230
Step blocks	10231	(0 ~ 200) μm	0.33 μm	Electronic micrometers, Precision surface plates /HCT-CS-337-10231
Step gauges	10232	(0 ~ 1 000) mm	$\sqrt{1.4^2+(0.005 \times I_0)^2}$ μm (unit of $I_0$ : mm)	Gauge blocks Electronic micrometers /HCT-CS-009-10232
Taper thickness gauges	10233	(0.1 ~ 60) mm	0.03 mm	Profile projectors /HCT-CS-242-10233
Ultrasonic thickness gauges	10234	(0 ~ 100) mm	3.6 μm	Ultrasonic thickness specimens /HCT-CS-243-10234
Ultrasonic/coating thickness specimens coating Ultrasonic	10235	(0 ~ 8) mm (0 ~ 100) mm	1.4 μm $\sqrt{1.4^2+(0.002 \times I_0)^2}$ μm (unit of $I_0$ : mm)	Gauge blocks Standard measuring machines Electronic micrometers /HCT-CS-227-10235
Coating thickness testers	10236	(0 ~ 0.25) mm (0.25 ~ 1.05) mm (1.05 ~ 3.7) mm (3.7 ~ 7.9) mm	1.5 μm 2.0 μm 6.9 μm 6.9 μm	Coating thickness specimens /HCT-CS-228-10236
Torque arms Touque Arm Wire	10237	(0 ~ 500) mm (0 ~ 5) mm	$\sqrt{3.1^2+(0.009 \times I_0)^2}$ μm (unit of 10 : mm) 0.7 μm	Contact coordinate measuring machines, Standard measuring machine /HCT-CS-287-10237
Width measuring specimens	10238	(0 ~ 1 000) mm	$\sqrt{1.7^2+(0.007 \times I_0)^2}$ μm (unit of $I_0$ : mm)	Contact coordinate measuring machines /HCT-CS-338-10238

103. Angle

Measured Quantity Instrument or Gauge	Field Code	Range	CMC (The Confidence Level is about 95 %)	Comments
Bevel protractors Angle accuracy Angle of accessories	10304	0° ~ 90° 90° ~ 360° 0° ~ 360°	1.3´ 2.0´ 2.3´	Angle gauge blocks, Precision surface plates, Profile projectors /HCT-CS-251-10304
Plate/square/electric levels Angle Squareness Flatness	10311	±200´´ ±1 000´´ ±2 000´´ (0 ~ 300) mm 300 mm × 60 mm	0.3´´ 0.5´´ 0.9´´ 2.3 µm 1.0 µm	Fine angle generators, Electronic micrometers, Squareness testers, Precision surface plates /HCT-CS-252-10311
Sine bars, plates, tables, centers Center distance of both rollers Flatness Parallelism between rollers Parallelism between the measuring plane and the roller	10317	(50 ~ 200) mm (50 ~ 200) mm (50 ~ 200) mm (50 ~ 200) mm	0.82 µm 0.24 µm 1.0 µm 0.27 µm	Standard measuring machines, Angle gauge blocks, Gauge blocks, Optical flats, Electronic micrometers, Precision surface plates /HCT-CS-326-10317
Squareness testers, right angle testers Squareness	10318	(0 ~ 400) mm	1.6 µm	Cylindrical squares, Precision surface plates, Electronic micrometers /HCT-CS-327-10318
Cylindrical squares Squareness Straightness	10319	(0 ~ 400) mm (0 ~ 400) mm	$\sqrt{2.0^2+(0.001 2 \times I_0)^2}$ µm (unit of $I_0$ : mm) 2.0 µm	Squareness testers, right angle testers Cylindrical squares Electronic micrometers Standard measuring machines /HCT-CS-328-10319
Precision squares Squareness Parallelism Straightness	10320	(0 ~ 450) mm (0 ~ 450) mm (0 ~ 450) mm	$\sqrt{2.0^2+0.003^2 \times I_0^2}$ µm ( $I_0$ : height, unit : mm) 1.2 µm 2.9 µm	Cylindrical squares, Squareness testers, right angle testers, Precision surface plates /HCT-CS-278-10320

104. Form

Measured Quantity Instrument or Gauge	Field Code	Range	CMC (The Confidence Level is about 95 %)	Comments
Form testers Z-axis X-axis Angle	10401	(0 ~ 60) mm (0 ~ 200) mm (0 ~ 180) °	0.16 μm $\sqrt{0.57^2+(0.002 \times I_0)^2}$ μm (unit of $I_0$ : mm) 1.3 ′	Form standard specimens, Angle gauge blocks, Standard scales. Gauge blocks, /HCT-CS-284-10401
Optical flats	10404	∅ (10 ~ 130) mm	0.06 μm	Optical flats, Monochromatic light sources /HCT-CS-229-10404
Optical parallels Flatness Parallelism	10405	∅ (10 ~ 30) mm ∅ (10 ~ 30) mm	0.061 μm 0.080 μm	Optical flats, Monochromatic light sources, Gauge block comparators /HCT-CS-230-10405
Parallel blocks Parallelism Flatness Difference of both blocks	10406	(0 ~ 1 000) mm (0 ~ 1 000) mm (0 ~ 1 000) mm	1.2 μm 1.2 μm 1.8 μm	Electronic micrometers, Precision surface plates, Test bars /HCT-CS-285-10406
Precision surface plates Area	10407	(0 ~ 2 500) cm <sup>2</sup> (2 500 ~ 5 000) cm <sup>2</sup> (5 000 ~ 10 000) cm <sup>2</sup> (10 000 ~ 15 000) cm <sup>2</sup> (15 000 ~ 30 000) cm <sup>2</sup> (30 000 ~ 60 000) cm <sup>2</sup>	1.8 μm 2.2 μm 2.6 μm 2.8 μm 3.9 μm 4.7 μm	Electric levels /HCT-CS-010-10407
Roundness measurement instruments Accuracy of detector Rotation accuracy of circumference direction Rotation accuracy of shaft direction Straightness	10409	(0 ~ 1 000) μm (0 ~ 360) ° (0 ~ 360) ° (0 ~ 300) mm	0.23 μm 16 nm 16 nm 1.3 μm	Roundness magnification specimens, Cylindrical squares Standard hemispheres, Optical flats, /HCT-CS-279-10409
Straight edges Straightness Parallelism	10412	(0 ~ 2 000) mm (0 ~ 2 000) mm	6.0 μm 5.9 μm	Electronic micrometers, Precision surface plates /HCT-CS-329-10412
Straight rules	10413	(0 ~ 3 000) mm	$\sqrt{0.3^2+(0.002 \times I_0)^2}$ mm (unit of $I_0$ : m)	Laser interferometer, Tape measure calibration system /HCT-CS-244-10413
Test bars Roundness Cylindricity Essentric	10415	(0 ~ 100) mm (0 ~ 100) mm (0 ~ 100) mm	0.062 μm 0.26 μm 0.51 μm	Roundness measurement instruments, Precision surface plates Electronic micrometers /HCT-CS-330-10415
Spherometers	10416	(0 ~ 10) mm	0.14 μm	Gauge blocks, Optical flats /HCT-CS-340-10416

105. Complex geometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Contact coordinate measuring machines Accuracy Straightness Squareness	10503	(0 ~ 600) mm (0 ~ 600) mm (0 ~ 600) mm	$\sqrt{0.53^2+(0.002 \times I_0)^2}$ μm (unit of $I_0$ : mm) 2.1 μm 0.8 ′	Step gauges, Precision squares, Straight edges /HCT-CS-011-10503
Non-contact coordinate measuring machines Accuracy	10504	(0 ~ 1 000) mm	$\sqrt{0.43^2+(0.003 \times I_0)^2}$ μm (unit of $I_0$ : mm)	Standard scales /HCT-CS-012-10504
Gauge block accessories Flatness(close-contact plane) Parallelism (Parallel jaw) Round type jaw A type Parallel jaw Base block Center point	10505	(0 ~ 50) mm (0 ~ 150) mm (0 ~ 50) mm (0 ~ 50) mm (0 ~ 50) mm (0 ~ 20) mm	0.03 μm 1.2 μm $\sqrt{0.26^2+(0.004 \times I_0)^2}$ μm $\sqrt{0.35^2+(0.004 \times I_0)^2}$ μm $\sqrt{0.65^2+(0.004 \times I_0)^2}$ μm (unit of $I_0$ : mm) 0.28 μm	Angle gauge blocks, Precision surface plates, Profile projectors /HCT-CS-308-10505
Measuring microscopes, profile projectors Length accuracy Right angle accuracy Magnification accuracy Squareness accuracy	10511	(0 ~ 300) mm (0 ~ 360) ° ×2 ~ ×100 (0 ~ 300) mm	$\sqrt{0.45^2+(0.003 \times I_0)^2}$ μm (unit of $I_0$ : mm) 1.7 ′ $3.2 \times 10^{-4}$ 3.6 μm	Standard scales, Precision squares /HCT-CS-013-10511
Micro measuring microscopes	10512	(0 ~ 50) mm	2.7 μm	Standard scales /HCT-CS-014-10512
Stylus type roughness testers Rsm Pt Ra Rz	10517	(0 ~ 120) μm (0 ~ 10) μm (0 ~ 1) μm (1 ~ 3) μm (0 ~ 3) μm (3 ~ 12) μm	1.4 μm 0.070 μm 0.015 μm 0.045 μm 0.052 μm 0.15 μm	Roughness standard/ comparison specimens /HCT-CS-295-10517
Thread plug gauges External diameter Effective diameter Pitch Half angle	10525	(1 ~ 100) mm (1 ~ 100) mm (0.2 ~ 6) mm (0 ~ 45) °	0.48 μm 1.6 μm 1.2 μm 1.8 ′	Standard measuring machines, Thread measuring wire gauges, Projectors /HCT-CS-016-10525
V-blocks, box blocks Flatness Parallelism Gradient Difference of both part Squareness	10529	(0 ~ 150) mm (0 ~ 150) mm (0 ~ 150) mm (0 ~ 150) mm (0 ~ 150) mm	1.0 μm 1.2 μm 0.6 μm 1.1 μm $\sqrt{2.0^2+0.003^2 \times I_0^2}$ μm ( $I_0$ : height, unit : mm)	Pin gauges, Electronic micrometers, Precision surface plates, Test bars /HCT-CS-283-10529



106. Various dimensional

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Inside/outside/gear tooth calipers, caliper gauges	10601	(0 ~ 150) mm (150 ~ 1 500) mm	$\sqrt{3.8^2+(0.007 \times I_0)^2}$ μm $\sqrt{7.6^2+(0.007 \times I_0)^2}$ μm (unit of $I_0$ : mm)	Gauge blocks /HCT-CS-017-10601
Cylinder/bore gauges Cylinder gauges Bore gauges	10603	(0 ~ 2) mm (0 ~ 2) mm	0.78 μm 0.76 μm	Dial gauge testers, Gauge blocks /HCT-CS-019-10603
Depth gauges, depth micrometers	10604	(0 ~ 300) mm (300 ~ 1 000) mm	$\sqrt{0.9^2+(0.004 \times I_0)^2}$ μm $\sqrt{7.2^2+(0.004 \times I_0)^2}$ μm (unit of $I_0$ : mm)	Gauge blocks /HCT-CS-020-10604
Dial/digital gauges	10605	(0 ~ 50) mm (50 ~ 150) mm	$\sqrt{0.16+(0.002 \times I_0)^2}$ μm $\sqrt{0.93+(0.002 \times I_0)^2}$ μm (unit of $I_0$ : mm)	Gauge blocks /HCT-CS-021-10605
Micro indicators, Test indicators	10609	(0 ~ 2) mm	0.33 μm	Dial gauge testers / HCT-CS-022-10609
Micrometer heads	10610	(0 ~ 50) mm	0.61 μm	Gauge blocks, Electronic micrometers / HCT-CS-023-10610
3-point micrometers	10611	(1 ~ 200) mm	1.9 μm	Standard ring gauges /HCT-CS-231-10611
Inside micrometers	10612	(5 ~ 300) mm (300 ~ 1 500) mm	$\sqrt{1.6^2+(0.005 \times I_0)^2}$ μm $\sqrt{2.3^2+(0.005 \times I_0)^2}$ μm (unit of $I_0$ : mm)	Gauge blocks / HCT-CS-026-10612
Outside micrometers	10613	(0 ~ 25) mm (25 ~ 100) mm (100 ~ 1 500) mm	$\sqrt{0.36^2+(0.004 \times I_0)^2}$ μm $\sqrt{1.4^2+(0.004 \times I_0)^2}$ μm $\sqrt{2.5^2+(0.004 \times I_0)^2}$ μm (unit of $I_0$ : mm)	Gauge blocks / HCT-CS-027-10613
Particle Counters Airbone particle counter Laser reference voltage Flow rate Threshold voltage Counting efficiency CPC OPC Liquid particle counter Laser reference voltage Flow rate Threshold voltage	10615	(0 ~ 10) V (0 ~ 100) L/min (0 ~ 10) V (0 ~ 1.0) μm (0.1 ~ 1.0) μm (0 ~ 10) V (0 ~ 25) mL/min (25 ~ 300) mL/min (0 ~ 10) V	$5.4 \times 10^{-4}$ $2.3 \times 10^{-2}$ $5.4 \times 10^{-4}$ 3.0 % 4.7 % $5.4 \times 10^{-4}$ $8.1 \times 10^{-2}$ $5.0 \times 10^{-2}$ $5.4 \times 10^{-4}$	Certified reference material (CRM), Particle counters, Flowmeters / HCT-CS-028-10615         / HCT-CS-029-10615
Standard sieves Sieve opening Wire rod diameter	10617	(0.01 ~ 8) mm (0.01 ~ 125) mm	1.7 μm 2.6 μm	Profile projectors /HCT-CS-232-10617

106. Various dimensional

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Welding gauges Height/depth measuring scale Thick measuring scale Rule measuring scale Angle measuring scale Taper measuring scale	10620	(0 ~ 100) mm (0 ~ 16) mm (0 ~ 50) mm (0 ~ 90) ° (0 ~ 7) mm	0.009 mm 0.009 mm 0.096 mm 0.13 ° 0.096 mm	Welding gauges /HCT-CS-246-10620
Particle dilution Systems PCRF	10622	(30 ~ 100) nm	$8.2 \times 10^{-2}$	ELECTRICAL PARTICLE SIZER, CPC/HCT-CS-256-10622

201. Mass and related quantities

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Counter beam balances	20105	(0 ~ 311) g (311 ~ 2 610) g (2.61 ~ 20) kg	10 mg 82 mg 0.82 g	Standard weights /HCT-CS-233-20105
Dial platform scale balances	20106	(0 ~ 30) kg (30 ~ 60) kg (60 ~ 100) kg	0.06 kg 0.12 kg 0.29 kg	Standard weights /HCT-CS-309-20106
Dial reading balances	20108	(0 ~ 160) g	0.19 mg	Weights /HCT-CS-031-20108
Electric balances	20109	(0 ~ 2) g (2 ~ 5) g (5 ~ 20) g (20 ~ 50) g (50 ~ 100) g (100 ~ 200) g (200 ~ 500) g (0.5 ~ 1) kg (1 ~ 2) kg (2 ~ 5) kg (5 ~ 10) kg (10 ~ 30) kg (30 ~ 50) kg (50 ~ 100) kg (100 ~ 200) kg (200 ~ 300) kg (300 ~ 600) kg	0.032 mg 0.054 mg 0.063 mg 0.11 mg 0.14 mg 0.20 mg 0.49 mg 0.94 mg 1.8 mg 4.6 mg 9.4 mg 25 mg 0.44 g 1.6 g 3.4 g 5.3 g 11 g	Standard weights /HCT-CS-032-20109
Platform scale balances	20112	(0 ~ 20) kg (20 ~ 200) kg	1.2 g 58 g	Standard weights /HCT-CS-234-20112
Spring scale balances	20113	(0 ~ 1) kg (1 ~ 2) kg (2 ~ 5) kg (5 ~ 10) kg (10 ~ 30) kg (30 ~ 50) kg	2.9 g 5.8 g 12 g 29 g 58 g 0.12 kg	Standard weights /HCT-CS-235-20113
Trip balances	20114	(0 ~ 500) g	0.12 g	Standard Weights /HCT-CS-356-20114
Weights F1 class	20116	1 mg 2 mg 5 mg 10 mg 20 mg 50 mg 100 mg 200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg	4.6 µg 4.6 µg 4.6 µg 5.1 µg 5.1 µg 5.8 µg 6.6 µg 7.4 µg 9.1 µg 11 µg 14 µg 17 µg 21 µg 27 µg 36 µg 55 µg 0.11 mg 0.29 mg 0.55 mg	Standard weights, Mass comparators /HCT-CS-033-20116

201. Mass and related quantities

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
F1 class	20116	2 kg 5 kg 10 kg 20 kg	1.1 mg 2.9 mg 5.5 mg 11 mg	



203. Torque

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Torque measuring devices	20302	(0.005 ~ 100) N · m	$7.1 \times 10^{-3}$	Weights, Torque arm /HCT-CS-036-20302
Torque wrenches/drivers	20303			Torque calibration machines /HCT-CS-037-20303
Clockwise		(0.06 ~ 0.6) N · m (0.6 ~ 1) N · m (1 ~ 2.5) N · m (2.5 ~ 5) N · m (5 ~ 10) N · m (10 ~ 25) N · m (25 ~ 50) N · m (50 ~ 100) N · m (100 ~ 250) N · m (250 ~ 500) N · m (500 ~ 1 000) N · m (1 000 ~ 2 000) N · m	$1.0 \times 10^{-2}$ $1.1 \times 10^{-2}$ $8.7 \times 10^{-3}$ $4.2 \times 10^{-3}$ $5.2 \times 10^{-3}$ $4.9 \times 10^{-3}$ $2.8 \times 10^{-3}$ $7.9 \times 10^{-3}$ $3.3 \times 10^{-3}$ $2.4 \times 10^{-3}$ $5.9 \times 10^{-3}$ $9.8 \times 10^{-3}$	
Counterclockwise		(0.06 ~ 0.6) N · m (0.6 ~ 1) N · m (1 ~ 2.5) N · m (2.5 ~ 5) N · m (5 ~ 10) N · m (10 ~ 25) N · m (25 ~ 50) N · m (50 ~ 100) N · m (100 ~ 250) N · m (250 ~ 500) N · m (500 ~ 1 000) N · m (1 000 ~ 2 000) N · m	$1.0 \times 10^{-2}$ $7.3 \times 10^{-3}$ $9.4 \times 10^{-3}$ $4.7 \times 10^{-3}$ $6.0 \times 10^{-3}$ $5.6 \times 10^{-3}$ $4.0 \times 10^{-3}$ $8.0 \times 10^{-3}$ $3.5 \times 10^{-3}$ $2.4 \times 10^{-3}$ $4.4 \times 10^{-3}$ $1.0 \times 10^{-2}$	

204. Pressure

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Altimeters	20401	(0 ~ 15) km	12 m	Pressure calibrators (PM500-BG200K) / HCT-CS-357-20401
Manometers	20402	(0 ~ 22) kPa	$7.3 \times 10^{-3}$	Pressure calibrators / HCT-CS-344-20402
Absolute pressure gauges	20406	(0.005 ~ 7) MPa abs.	$6.0 \times 10^{-4}$	Pressure calibrators (PM500-BG200K), (PACE 6000) / HCT-CS-255-20406
Compound pressure gauges	20408	(-0.095 ~ 7) MPa	$2.9 \times 10^{-4}$	Pressure calibrators (PACE 6000) / HCT-CS-215-20408
Differential pressure gauges pneumatic	20409	(0 ~ 7) MPa	$5.8 \times 10^{-4}$	Pressure calibrators (PACE 6000) / HCT-CS-215-20408
Gauge pressure gauges	20411	(0 ~ 500) kPa (0.5 ~ 10) MPa (10 ~ 100) MPa (100 ~ 200) MPa	$1.2 \times 10^{-4}$ $2.0 \times 10^{-4}$ $8.1 \times 10^{-5}$ $7.1 \times 10^{-5}$	Pneumatic pressure balances(PDPG-P), Hydraulic pressure balances(PG7302) / HCT-CS-039-20411
Pressure transducers/transmitters	20412	(-95 ~ 0) kPa (0 ~ 500) kPa (0.5 ~ 10) MPa (10 ~ 100) MPa (100 ~ 200) MPa (0.005 ~ 7) MPa abs.	$2.8 \times 10^{-3}$ $3.1 \times 10^{-3}$ $3.4 \times 10^{-3}$ $3.3 \times 10^{-3}$ $2.8 \times 10^{-3}$ $3.5 \times 10^{-3}$	Pressure calibrators (PM500-BG200K), (PACE 6000) Pneumatic pressure balances(PDPG-P), Hydraulic pressure balances(PG7302) / HCT-CS-169-20413
Dial type vacuum gauges	20413	(-95 ~ 0) kPa	$2.0 \times 10^{-2}$	Pressure calibrators / HCT-CS-216-20413
Water Depth meters	20414	(0 ~ 2) MPa	$1.2 \times 10^{-2}$	Pressure calibrators /HCT-CS-245-20414

206. Volume

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Volumetric glasswares	20601	(0.1 ~ 2) ml (2 ~ 10) ml (10 ~ 25) ml (25 ~ 100) ml (100 ~ 250) ml (250 ~ 500) ml (500 ~ 1 000) ml (1 000 ~ 2 000) ml (2 000 ~ 5 000) ml	2.4 μl 4.9 μl 13 μl 26 μl 47 μl 70 μl 0.14 ml 0.22 ml 0.43 ml	Electric balances, Weights, Pure water /HCT-CS313-20601
Pycnometers	20602	(0 ~ 50) ml (50 ~ 100) ml (100 ~ 500) ml	3.2 μl 7.0 μl 30 μl	Electric balances, Weights, Pure water /HCT-CS313-20601
Concrete air content meters	20605	(0 ~ 10) %	0.01 %	Electric balances, Weights, Pure water /HCT-CS-314-20605
Piston type volume meters	20606	(0.1 ~ 5) μl (5 ~ 10) μl (10 ~ 20) μl (20 ~ 50) μl (50 ~ 100) μl (0.1 ~ 0.2) ml (0.2 ~ 0.5) ml (0.5 ~ 1) ml (1 ~ 2) ml (2 ~ 5) ml (5 ~ 10) ml (10 ~ 20) ml	18 nl 26 nl 36 nl 73 nl 0.12 μl 0.23 μl 0.53 μl 1.2 μl 2.3 μl 5.3 μl 12 μl 23 μl	Electric balances, Weights, Pure water /HCT-CS-315-20606



207.Density

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Liquid density meters	20702	(0 ~ 2) g/cm <sup>3</sup>	0.000 14 g/cm <sup>3</sup>	Density standard materials /HCT-CS-351-20702
Salinity meters	20704	(0 ~ 26) %	0.012 %	NaCl /HCT-CS-352-20704
Sucrose meters	20705	(0 ~ 80) %	0.032 %	Sucrose /HCT-CS-294-20705
Chloride meters	20707	(0 ~ 1.5) %	0.002 4 %	Chloride ion standard solution /HCT-CS-353-20707

208. Viscosity

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Dynamic viscometers: rotaional, etc. Rotational viscometers	20802	(2.5 ~ 200 000) mPa · s	$1.7 \times 10^{-2}$	Viscosity standard /HCT-CS-288-20802

209. Fluid flow

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Anemometers; hot-wire	20901	(0.1 ~ 2) m/s (2 ~ 55) m/s	$5.8 \times 10^{-2}$ $4.4 \times 10^{-3}$	Lasor Doppler / HCT-CS-272-20901
Anemometers; pitot tube, etc.	20902	(0.1 ~ 2) m/s (2 ~ 55) m/s	$5.8 \times 10^{-2}$ $4.4 \times 10^{-3}$	Lasor Doppler / HCT-CS-273-20902
Gas flowmeters; differential pressure	20908	(0.001 8 ~ 260) m <sup>3</sup> /h	$2.5 \times 10^{-3}$	Sonic nozzle / HCT-CS-312-20908
Liquid flowmeters; differential pressure	20909	(0.005 ~ 50) m <sup>3</sup> /h	$2.4 \times 10^{-3}$	Reference flowmeter / HCT-CS-313-20909
Liquid flowmeter; electromagnetic	20910	(0.005 ~ 50) m <sup>3</sup> /h	$2.4 \times 10^{-3}$	Reference flowmeter / HCT-CS-313-20909
Gas flowmeters; thermal mass, etc.	20911	(0.001 8 ~ 260) m <sup>3</sup> /h	$2.5 \times 10^{-3}$	Sonic nozzle / HCT-CS-312-20908
Liquid flowmeters; Coriolis, etc.	20912	(0.005 ~ 50) m <sup>3</sup> /h	$2.4 \times 10^{-3}$	Reference flowmeter / HCT-CS-313-20909
Gas flowmeters; positive displacement	20914	(0.001 8 ~ 260) m <sup>3</sup> /h	$2.5 \times 10^{-3}$	Sonic nozzle / HCT-CS-312-20908
Liquid flowmeters; positive displacement	20915	(0.005 ~ 50) m <sup>3</sup> /h	$2.4 \times 10^{-3}$	Reference flowmeter / HCT-CS-313-20909
Gas flowmeters; turbine	20916	(0.001 8 ~ 260) m <sup>3</sup> /h	$2.5 \times 10^{-3}$	Sonic nozzle / HCT-CS-312-20908
Liquid flowmeters; turbine	20917	(0.005 ~ 50) m <sup>3</sup> /h	$2.4 \times 10^{-3}$	Reference flowmeter / HCT-CS-313-20909
Gas flowmeters; ultrasonic	20918	(0.001 8 ~ 260) m <sup>3</sup> /h	$2.5 \times 10^{-3}$	Sonic nozzle / HCT-CS-312-20908
Liquid flowmeters; ultrasonic	20919	(0.005 ~ 50) m <sup>3</sup> /h	$2.4 \times 10^{-3}$	Reference flowmeter / HCT-CS-313-20909
Gas flowmeters; variable area	20920	(0.001 8 ~ 260) m <sup>3</sup> /h	$2.5 \times 10^{-3}$	Sonic nozzle / HCT-CS-312-20908
Liquid flowmeters; variable area	20921	(0.005 ~ 50) m <sup>3</sup> /h	$2.4 \times 10^{-3}$	Reference flowmeter / HCT-CS-313-20909
Gas flowmeters; vortex	20922	(0.001 8 ~ 260) m <sup>3</sup> /h	$2.5 \times 10^{-3}$	Sonic nozzle / HCT-CS-312-20908
Liquid flowmeters; vortex	20923	(0.005 ~ 50) m <sup>3</sup> /h	$2.4 \times 10^{-3}$	Reference flowmeter / HCT-CS-313-20909
Anemometers; vane, etc.	20925	(0.1 ~ 2) m/s (2 ~ 55) m/s	$5.8 \times 10^{-2}$ $4.4 \times 10^{-3}$	Lasor Doppler / HCT-CS-274-20925

301. Time/frequency

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Frequency standards Timebase Frequency	30102	100 kHz ~ 10 MHz	$2.4 \times 10^{-12}$	Atomic clock /HCT-CS-040-30102
General frequency sources Output frequency	30103	100 kHz ~ 100 MHz	$2.7 \times 10^{-12}$	Atomic clock /HCT-CS-041-30103
Frequency meters/counters Input Frequency	30104	0.1 Hz 0.1 Hz ~ 3 GHz (3 ~ 40) GHz	$6.2 \times 10^{-11}$ $6.2 \times 10^{-12}$ 0.58 Hz	Standard frequency, General frequency sources HCT-CS-042-30104
Timebase Frequency		100 kHz ~ 10 MHz	$2.7 \times 10^{-12}$	
Time interval sources Time interval	30105	(1 ~ 10) ns (10 ~ 100) ns 100 ns ~ 1 μs (1 ~ 10) μs (10 ~ 100) μs 100 μs ~ 1 ms (1 ~ 10) ms (10 ~ 100) ms 100 ms ~ 1 s	$6.2 \times 10^{-7}$ $6.2 \times 10^{-7}$ $6.2 \times 10^{-7}$ $6.2 \times 10^{-7}$ $6.2 \times 10^{-7}$ $6.2 \times 10^{-7}$ $6.2 \times 10^{-7}$ $6.2 \times 10^{-7}$ $6.2 \times 10^{-7}$	Frequency counters /HCT-CS-043-30105
Frequency		(1 ~ 10) Hz (10 ~ 100) Hz 100 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (1 ~ 10) MHz (10 ~ 100) MHz 100 MHz ~ 1 GHz	$6.2 \times 10^{-7}$ $6.2 \times 10^{-7}$ $6.2 \times 10^{-7}$ $6.2 \times 10^{-7}$ $6.2 \times 10^{-7}$ $6.2 \times 10^{-7}$ $6.2 \times 10^{-7}$ $6.2 \times 10^{-7}$ $6.2 \times 10^{-7}$	
Time interval meters/Stop watches /Timers Relative time difference	30106	day month	$2.8 \times 10^{-7}$ $1.3 \times 10^{-8}$	Atomic clock /HCT-CS-044-30106
Timer		(1 ~ 60) s (60 ~ 6 000) s (6 000 ~ 86 400) s	$6.2 \times 10^{-6}$ $6.2 \times 10^{-5}$ $6.2 \times 10^{-5}$	

302. Velocity & revolution

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Standard RPM generators Revolution velocity measurement  Centrifuge	30201	(1 ~ 1 000) min <sup>-1</sup> (1 000 ~ 100 000) min <sup>-1</sup>  100 min <sup>-1</sup> (100 ~ 900) min <sup>-1</sup> (900 ~ 1 000) min <sup>-1</sup> (1 000 ~ 3 000) min <sup>-1</sup> (3 000 ~ 6 000) min <sup>-1</sup> (6 000 ~ 10 000) min <sup>-1</sup> (10 000 ~ 20 000) min <sup>-1</sup> (20 000 ~ 30 000) min <sup>-1</sup> (30 000 ~ 40 000) min <sup>-1</sup> (40 000 ~ 50 000) min <sup>-1</sup> (50 000 ~ 60 000) min <sup>-1</sup> (60 000 ~ 70 000) min <sup>-1</sup> (80 000 ~ 90 000) min <sup>-1</sup> (90 000 ~ 99 000) min <sup>-1</sup>	6.2×10 <sup>-2</sup> min <sup>-1</sup> 6.2×10 <sup>-1</sup> min <sup>-1</sup>  0.12 min <sup>-1</sup> 1.1 min <sup>-1</sup> 1.2 min <sup>-1</sup> 1.4 min <sup>-1</sup> 1.9 min <sup>-1</sup> 2.6 min <sup>-1</sup> 4.8 min <sup>-1</sup> 7.1 min <sup>-1</sup> 10 min <sup>-1</sup> 12 min <sup>-1</sup> 14 min <sup>-1</sup> 17 min <sup>-1</sup> 21 min <sup>-1</sup> 23 min <sup>-1</sup>	Atomic clock /HCT-CS-045-30201
Contact type tachometers Revolution velocity measurement	30202	(1 ~ 4 000) min <sup>-1</sup>	6.2×10 <sup>-2</sup> min <sup>-1</sup>	Atomic clock /HCT-CS-046-30202
Photo tachometers/stroboscopes Photo-tachometer  Stroboscopic tachometer	30203	1 min <sup>-1</sup> (1 ~ 300) min <sup>-1</sup> (300 ~ 6 000) min <sup>-1</sup> (6 000 ~ 100 000) min <sup>-1</sup>  1 min <sup>-1</sup> (1 ~ 300) min <sup>-1</sup> (300 ~ 6 000) min <sup>-1</sup> (6 000 ~ 100 000) min <sup>-1</sup>	0.006 2 min <sup>-1</sup> 6.2×10 <sup>-3</sup> min <sup>-1</sup> 6.2×10 <sup>-2</sup> min <sup>-1</sup> 6.2×10 <sup>-1</sup> min <sup>-1</sup>  0.006 2 min <sup>-1</sup> 6.2×10 <sup>-3</sup> min <sup>-1</sup> 6.2×10 <sup>-2</sup> min <sup>-1</sup> 6.2×10 <sup>-1</sup> min <sup>-1</sup>	Atomic clock /HCT-CS-047-30203
Wow-flutter generators Wow-flutter Deflection  Frequency  Level	30205	(0.01 ~ 3) %  0.1 Hz ~ 99.99 kHz  100 Hz ~ 100 kHz 100 mV  100 Hz ~ 100 kHz (100 mV ~ 1 V)  100 Hz ~ 100 kHz (1 V ~ 10 V)	6.2×10 <sup>-3</sup>  6.2×10 <sup>-4</sup>  1.3×10 <sup>-3</sup>  1.1×10 <sup>-3</sup>  1.1×10 <sup>-3</sup>	Atomic clock /HCT-CS-049-30205
Wow-flutter meters Wow-flutter Deflection  Input frequency  Output frequency  CCIR PULSE	30206	0.01 % 0.03 % 0.1 % 0.3 % 1 % 3 %  10 Hz 99.99 kHz  3.00 kHz 3.15 kHz  10 ms ~ 100 ms	2.4×10 <sup>-4</sup> 4.6×10 <sup>-4</sup> 1.6×10 <sup>-3</sup> 4.6×10 <sup>-3</sup> 1.5×10 <sup>-2</sup> 4.5×10 <sup>-2</sup>  0.58 Hz 5.8 Hz  0.58 Hz 0.58 Hz  1.5×10 <sup>-2</sup>	Wow-flutter generators /HCT-CS-050-30206

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC ammeters DC Current	40101	(+) 10 pA (10 ~ 40) pA (40 ~ 100) pA 100 pA ~ 1 nA (1 ~ 4) nA (4 ~ 10) nA (10 ~ 40) nA 40 nA ~ 1 μA (1 ~ 4) μA (4 ~ 8) μA (8 ~ 10) μA (10 ~ 40) μA (40 ~ 80) μA (80 ~ 100) μA (100 ~ 400) μA (400 ~ 800) μA (0.8 ~ 1) mA (1 ~ 4) mA (4 ~ 8) mA (8 ~ 10) mA (10 ~ 40) mA (40 ~ 80) mA (80 ~ 100) mA (100 ~ 400) mA (400 ~ 800) mA (0.8 ~ 1) A (1 ~ 4) A (4 ~ 8) A (8 ~ 10) A (10 ~ 40) A (40 ~ 80) A (80 ~ 100) A  (-) -10 pA (-10 ~ -40) pA (-40 ~ -100) pA -100 pA ~ -1 nA (-1 ~ -4) nA (-4 ~ -10) nA (-10 ~ -40) nA -40 nA ~ -1 μA (-1 ~ -4) μA (-4 ~ -8) μA (-8 ~ -10) μA (-10 ~ -40) μA (-40 ~ -80) μA (-80 ~ -100) μA (-100 ~ -400) μA (-400 ~ -800) μA (-0.8 ~ -1) mA (-1 ~ -4) mA (-4 ~ -8) mA (-8 ~ -10) mA (-10 ~ -40) mA (-40 ~ -80) mA (-80 ~ -100) mA (-100 ~ -400) mA (-400 ~ -800) mA	7.1 fA $4.0 \times 10^{-4}$ $3.6 \times 10^{-4}$ $1.2 \times 10^{-4}$ $3.1 \times 10^{-5}$ $2.7 \times 10^{-5}$ $3.3 \times 10^{-5}$ $2.8 \times 10^{-5}$ $1.8 \times 10^{-3}$ $9.3 \times 10^{-4}$ $7.4 \times 10^{-4}$ $2.3 \times 10^{-4}$ $1.4 \times 10^{-4}$ $1.2 \times 10^{-4}$ $7.3 \times 10^{-5}$ $5.5 \times 10^{-5}$ $5.1 \times 10^{-5}$ $6.5 \times 10^{-5}$ $5.0 \times 10^{-5}$ $4.8 \times 10^{-5}$ $7.8 \times 10^{-5}$ $6.4 \times 10^{-5}$ $8.6 \times 10^{-5}$ $3.8 \times 10^{-5}$ $1.9 \times 10^{-5}$ $1.3 \times 10^{-4}$ $2.1 \times 10^{-4}$ $1.4 \times 10^{-4}$ $1.2 \times 10^{-4}$ $8.9 \times 10^{-4}$ $4.5 \times 10^{-4}$ $3.7 \times 10^{-4}$  7.1 fA $4.0 \times 10^{-4}$ $3.6 \times 10^{-4}$ $1.2 \times 10^{-4}$ $3.1 \times 10^{-5}$ $2.7 \times 10^{-5}$ $3.3 \times 10^{-5}$ $2.7 \times 10^{-5}$ $1.8 \times 10^{-3}$ $9.3 \times 10^{-4}$ $7.4 \times 10^{-4}$ $2.3 \times 10^{-4}$ $1.4 \times 10^{-4}$ $1.2 \times 10^{-4}$ $7.3 \times 10^{-5}$ $5.5 \times 10^{-5}$ $5.1 \times 10^{-5}$ $6.5 \times 10^{-5}$ $5.0 \times 10^{-5}$ $4.8 \times 10^{-5}$ $7.8 \times 10^{-5}$ $6.4 \times 10^{-5}$ $8.6 \times 10^{-5}$ $3.8 \times 10^{-5}$ $1.9 \times 10^{-5}$	Current calibrators Multimeter calibrators / HCT-CS-051-40101

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC Current	40101	(-0.8 ~ -1) A (-1 ~ -4) A (-4 ~ -8) A (-8 ~ -10) A (-10 ~ -40) A (-40 ~ -80) A (-80 ~ -100) A	$1.3 \times 10^{-4}$ $2.1 \times 10^{-4}$ $1.4 \times 10^{-4}$ $1.2 \times 10^{-4}$ $8.9 \times 10^{-4}$ $4.5 \times 10^{-4}$ $3.7 \times 10^{-4}$	
Transconductance amplifiers DC Current	40102	0 $\mu$ A (+) (0 ~ 1) mA (1 ~ 10) mA 10 mA ~ 1 A (1 ~ 10) A (10 ~ 100) A  (-) (0 ~ -1) mA (-1 ~ -10) mA -10 mA ~ -1 A (-1 ~ -10) A (-10 ~ -100) A	1.2 nA  $2.8 \times 10^{-5}$ $2.5 \times 10^{-5}$ $2.8 \times 10^{-5}$ $3.5 \times 10^{-5}$ $4.5 \times 10^{-5}$  $2.8 \times 10^{-5}$ $2.5 \times 10^{-5}$ $2.8 \times 10^{-5}$ $3.5 \times 10^{-5}$ $4.5 \times 10^{-5}$	Digital multimeters, Current shunts, Multimeter calibrators HCT-CS-052-40102
AC Current		50 Hz 100 $\mu$ A (100 ~ 400) $\mu$ A 400 $\mu$ A ~ 1 mA (1 ~ 4) mA (4 ~ 10) mA (10 ~ 40) mA (40 ~ 100) mA (100 ~ 400) mA 400 mA ~ 1 A (1 ~ 4) A (4 ~ 10) A (10 ~ 40) A (40 ~ 100) A  (50 ~ 100) Hz 100 $\mu$ A (100 ~ 400) $\mu$ A 400 $\mu$ A ~ 1 mA (1 ~ 4) mA (4 ~ 10) mA (10 ~ 40) mA (40 ~ 100) mA (100 ~ 400) mA 400 mA ~ 1 A (1 ~ 4) A (4 ~ 10) A (10 ~ 40) A (40 ~ 100) A  (100 Hz ~ 1 kHz) 100 $\mu$ A (100 ~ 400) $\mu$ A (400 ~ 800) $\mu$ A 800 $\mu$ A ~ 1 mA (1 ~ 4) mA (4 ~ 10) mA	20 nA  $1.9 \times 10^{-4}$ $1.5 \times 10^{-4}$ $1.8 \times 10^{-4}$ $1.4 \times 10^{-4}$ $1.8 \times 10^{-4}$ $1.4 \times 10^{-4}$ $1.8 \times 10^{-4}$ $1.4 \times 10^{-4}$ $1.8 \times 10^{-4}$ $1.4 \times 10^{-4}$ $1.9 \times 10^{-4}$ $1.4 \times 10^{-4}$ $2.0 \times 10^{-4}$ $1.6 \times 10^{-4}$  20 nA  $1.8 \times 10^{-4}$ $1.5 \times 10^{-4}$ $1.7 \times 10^{-4}$ $1.4 \times 10^{-4}$ $1.7 \times 10^{-4}$ $1.4 \times 10^{-4}$ $1.8 \times 10^{-4}$ $1.3 \times 10^{-4}$ $1.8 \times 10^{-4}$ $1.4 \times 10^{-4}$ $2.0 \times 10^{-4}$ $1.6 \times 10^{-4}$  17 nA  $1.6 \times 10^{-4}$ $1.3 \times 10^{-4}$ $1.0 \times 10^{-4}$ $1.5 \times 10^{-4}$ $1.1 \times 10^{-4}$	

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC Current	40102	(10 ~ 40) mA (40 ~ 100) mA (100 ~ 400) mA 400 mA ~ 1 A (1 ~ 4) A (4 ~ 10) A (10 ~ 40) A (40 ~ 100) A  (1 ~ 10) kHz 100 μA (100 ~ 400) μA 400 μA ~ 4 mA (4 ~ 10) mA (10 ~ 40) mA (40 ~ 100) mA (100 ~ 400) mA 400 mA ~ 1 A (1 ~ 4) A (4 ~ 10) A (10 ~ 40) A (40 ~ 100) A  (10 ~ 100) kHz 100 μA (100 ~ 400) μA (400 ~ 800) μA (0.8 ~ 1) mA (1 ~ 4) mA (4 ~ 8) mA (8 ~ 10) mA (10 ~ 40) mA (40 ~ 80) mA (80 ~ 100) mA (100 ~ 400) mA (400 ~ 800) mA (0.8 ~ 1) A (1 ~ 4) A (4 ~ 8) A (8 ~ 10) A (10 ~ 40) A (40 ~ 80) A (80 ~ 100) A	$1.5 \times 10^{-4}$ $1.1 \times 10^{-4}$ $1.5 \times 10^{-4}$ $1.1 \times 10^{-4}$ $1.6 \times 10^{-4}$ $1.2 \times 10^{-4}$ $1.8 \times 10^{-4}$ $1.5 \times 10^{-4}$  21 nA $1.9 \times 10^{-4}$ $1.7 \times 10^{-4}$ $1.4 \times 10^{-4}$ $1.8 \times 10^{-4}$ $1.4 \times 10^{-4}$ $1.8 \times 10^{-4}$ $1.3 \times 10^{-4}$ $2.0 \times 10^{-4}$ $1.7 \times 10^{-4}$ $2.8 \times 10^{-4}$ $2.6 \times 10^{-4}$  0.11 μA $1.4 \times 10^{-3}$ $9.7 \times 10^{-4}$ $9.0 \times 10^{-4}$ $1.4 \times 10^{-3}$ $9.6 \times 10^{-4}$ $8.8 \times 10^{-4}$ $1.4 \times 10^{-3}$ $9.6 \times 10^{-4}$ $8.8 \times 10^{-4}$ $1.4 \times 10^{-3}$ $9.6 \times 10^{-4}$ $8.8 \times 10^{-4}$ $1.4 \times 10^{-3}$ $9.7 \times 10^{-4}$ $8.9 \times 10^{-4}$ $1.4 \times 10^{-3}$ $1.0 \times 10^{-3}$ $9.2 \times 10^{-4}$	
DC Voltage/Current Calibrator DC Current	40103	0 pA (±) (0 ~ 1) pA (1 ~ 10) pA (10 ~ 100) pA (0.1 ~ 1) nA (1 ~ 10) nA (10 ~ 100) nA (0.1 ~ 1) μA (1 ~ 10) μA (10 ~ 100) μA (0.1 ~ 1) mA (1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 10) A (10 ~ 100) A	26 fA  $1.6 \times 10^{-2}$ $1.2 \times 10^{-2}$ $1.2 \times 10^{-2}$ $2.7 \times 10^{-5}$ $1.5 \times 10^{-5}$ $1.2 \times 10^{-5}$ $1.1 \times 10^{-5}$ $9.2 \times 10^{-6}$ $1.1 \times 10^{-5}$ $1.1 \times 10^{-5}$ $9.1 \times 10^{-6}$ $1.3 \times 10^{-5}$ $1.0 \times 10^{-5}$ $1.2 \times 10^{-4}$ $5.3 \times 10^{-4}$	Digital Multimeter, Current Shunt /HCT-CS-053-40103



401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC Current	40103	0 mV (±) (0 ~ 10) mV (10 ~ 100) mV (0.1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V	0.24 μV  1.7×10 <sup>-5</sup> 7.9×10 <sup>-6</sup> 7.2×10 <sup>-6</sup> 8.0×10 <sup>-6</sup> 8.2×10 <sup>-6</sup>	
Temperature Calibrator via Electrical Standards	40104			Digital Multimeter /HCT-CS-205-40104
DC Current Source		1 mA (1 ~ 10) mA (10 ~ 20) mA (20 ~ 30) mA	63 nA  1.7×10 <sup>-5</sup> 1.4×10 <sup>-5</sup> 7.0×10 <sup>-5</sup>	
DC Voltage Source		(-10 ~ 0) mV 0 mV (0 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 30) V	2.8×10 <sup>-5</sup> 0.13 μV 2.8×10 <sup>-5</sup> 4.9×10 <sup>-6</sup> 7.3×10 <sup>-6</sup> 6.6×10 <sup>-6</sup>	
Resistance Source		10 Ω (10 ~ 100) Ω 100 Ω ~ 1 kΩ (1 ~ 100) kΩ	0.12 mΩ  1.1×10 <sup>-5</sup> 2.3×10 <sup>-5</sup> 1.1×10 <sup>-5</sup>	
DC Current Meter		1 mA (1 ~ 10) mA (10 ~ 30) mA (20 ~ 30) mA (30 ~ 50) mA (50 ~ 70) mA (70 ~ 100) mA	80 nA  4.8×10 <sup>-5</sup> 4.4×10 <sup>-5</sup> 8.3×10 <sup>-5</sup> 7.0×10 <sup>-5</sup> 6.5×10 <sup>-5</sup> 6.1×10 <sup>-5</sup>	
DC Voltage Meter		(-10 ~ 0) mV 0 mV (0 ~ 1) mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 200) V (200 ~ 300) V	6.0×10 <sup>-5</sup> 0.50 μV 5.2×10 <sup>-4</sup> 6.0×10 <sup>-5</sup> 1.4×10 <sup>-5</sup> 6.2×10 <sup>-5</sup> 2.4×10 <sup>-5</sup> 8.8×10 <sup>-6</sup> 1.2×10 <sup>-5</sup> 9.9×10 <sup>-6</sup>	
Resistance Meter		10 Ω (10 ~ 100) Ω (0.1 ~ 10) kΩ (10 ~ 100) kΩ	0.28 mΩ  1.6×10 <sup>-5</sup> 1.2×10 <sup>-5</sup> 1.4×10 <sup>-5</sup>	
DC current shunts Resistance	40105	25 μΩ (25 ~ 50) μΩ (50 ~ 100) μΩ (0.1 ~ 8) mΩ (8 ~ 10) mΩ (10 ~ 16) mΩ (16 ~ 80) mΩ (0.08 ~ 0.1) Ω (0.1 ~ 0.16) Ω (0.16 ~ 0.4) Ω (0.4 ~ 0.8) Ω	6.1 nΩ  1.8×10 <sup>-4</sup> 1.7×10 <sup>-4</sup> 1.6×10 <sup>-4</sup> 4.0×10 <sup>-4</sup> 1.5×10 <sup>-4</sup> 1.7×10 <sup>-4</sup> 1.2×10 <sup>-4</sup> 2.1×10 <sup>-4</sup> 1.1×10 <sup>-4</sup> 1.1×10 <sup>-4</sup>	Digital multimeters, Multimeter calibrators, Current calibrators /HCT-CS-054-40105

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistance	40105	(0.8 ~ 1) Ω (1 ~ 1.6) Ω (1.6 ~ 4) Ω (4 ~ 8) Ω (8 ~ 10) Ω (10 ~ 16) Ω (16 ~ 40) Ω (40 ~ 80) Ω (80 ~ 100) Ω (100 ~ 800) Ω (800 ~ 1 000) Ω	$6.2 \times 10^{-5}$ $1.3 \times 10^{-4}$ $1.8 \times 10^{-4}$ $6.2 \times 10^{-5}$ $5.0 \times 10^{-5}$ $7.6 \times 10^{-5}$ $9.6 \times 10^{-5}$ $4.9 \times 10^{-5}$ $5.3 \times 10^{-5}$ $5.2 \times 10^{-5}$ $1.2 \times 10^{-4}$	
Galvanometers/null detectors DC Voltage	40106	3 μV (3 ~ 10) μV (10 ~ 30) μV (30 ~ 100) μV (100 ~ 300) μV (0.3 ~ 1) mV (1 ~ 3) mV (3 ~ 10) mV (10 ~ 30) mV (30 ~ 100) mV (100 ~ 300) mV (0.3 ~ 1) V (1 ~ 3) V (3 ~ 10) V (10 ~ 30) V (30 ~ 100) V (100 ~ 300) V (300 ~ 1 000) V	58 nV $5.8 \times 10^{-3}$ $9.6 \times 10^{-3}$ $5.8 \times 10^{-3}$ $9.6 \times 10^{-3}$ $5.8 \times 10^{-3}$ $9.6 \times 10^{-3}$ $5.8 \times 10^{-3}$ $9.6 \times 10^{-3}$ $5.8 \times 10^{-3}$ $9.6 \times 10^{-3}$ $5.8 \times 10^{-3}$ $9.6 \times 10^{-3}$ $5.8 \times 10^{-3}$ $9.6 \times 10^{-3}$ $5.8 \times 10^{-3}$ $9.6 \times 10^{-3}$ $5.8 \times 10^{-3}$	Multimeter calibrators, Current shunts /HCT-CS-247-40106
DC Power Supply DC Voltage  DC Current	40108	0 V (±) (0 ~ 40) mV (40 ~ 80) mV (80 ~ 100) mV (100 ~ 400) mV (400 ~ 800) mV (0.8 ~ 1) V (1 ~ 4) V (4 ~ 8) V (8 ~ 10) V (10 ~ 40) V (40 ~ 80) V (80 ~ 100) V (100 ~ 400) V (400 ~ 800) V (800 ~ 1 000) V (1 000 ~ 1 500) V (1 500 ~ 2 000) V  0 A (±) (0 ~ 40) μA (40 ~ 80) μA (80 ~ 100) μA (100 ~ 400) μA (0.4 ~ 1) mA (1 ~ 4) mA (4 ~ 10) mA	0.63 μV $1.8 \times 10^{-5}$ $1.0 \times 10^{-5}$ $8.0 \times 10^{-6}$ $1.7 \times 10^{-5}$ $8.8 \times 10^{-6}$ $7.3 \times 10^{-6}$ $1.6 \times 10^{-5}$ $8.6 \times 10^{-6}$ $7.2 \times 10^{-6}$ $1.7 \times 10^{-5}$ $9.4 \times 10^{-6}$ $8.1 \times 10^{-6}$ $1.8 \times 10^{-5}$ $1.0 \times 10^{-5}$ $9.0 \times 10^{-6}$ $6.7 \times 10^{-4}$ $5.0 \times 10^{-4}$  0.65 nA $3.2 \times 10^{-5}$ $2.6 \times 10^{-5}$ $2.5 \times 10^{-5}$ $3.0 \times 10^{-5}$ $2.5 \times 10^{-5}$ $3.0 \times 10^{-5}$ $2.5 \times 10^{-5}$	Digital Multimeter, Current Shunt /HCT-CS-057-40108

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC Current	40108	(10 ~ 40) mA (40 ~ 100) mA (100 ~ 400) mA (400 ~ 800) mA (0.8 ~ 1) A (1 ~ 4) A (4 ~ 10) A (10 ~ 40) A (40 ~ 100) A (100 ~ 300) A (300 ~ 1 000) A	$3.3 \times 10^{-5}$ $2.6 \times 10^{-5}$ $3.2 \times 10^{-5}$ $2.8 \times 10^{-5}$ $2.7 \times 10^{-5}$ $3.7 \times 10^{-5}$ $3.3 \times 10^{-5}$ $4.8 \times 10^{-5}$ $4.5 \times 10^{-5}$ $2.0 \times 10^{-4}$ $2.2 \times 10^{-4}$	
Ripple		1 mV (1 ~ 5) mV (5 ~ 10) mV (10 ~ 20) mV	98 $\mu$ V $2.2 \times 10^{-2}$ $1.9 \times 10^{-2}$ $4.3 \times 10^{-2}$	
Load&Line Regulation		1 mV (1 ~ 5) mV (5 ~ 500) mV	0.7 $\mu$ V $1.4 \times 10^{-4}$ $1.2 \times 10^{-4}$	
DC voltage dividers	40110	Ratio 1 000 : 1 1 kV (1 ~ 5) kV (5 ~ 100) kV 10 000 : 1 1 kV (1 ~ 5) kV (5 ~ 100) kV	$4.2 \times 10^{-4}$ $3.9 \times 10^{-4}$ $3.8 \times 10^{-4}$ $4.2 \times 10^{-4}$ $3.9 \times 10^{-4}$ $3.8 \times 10^{-4}$	High voltage deviders /HCT-CS-348-40110
DC voltage standards	40111	DC Voltage 1.018 V 10 V	$4.8 \times 10^{-7}$ $3.2 \times 10^{-7}$	Standard cells, Digital multimeters /HCT-CS-275-40111
DC voltmeters	40112	DC Voltage 0 V (+) (0 ~ 4) mV (4 ~ 8) mV (8 ~ 10) mV (10 ~ 40) mV (40 ~ 80) mV (80 ~ 100) mV (100 ~ 400) mV (400 ~ 800) mV (0.8 ~ 1) V (1 ~ 4) V (4 ~ 8) V (8 ~ 10) V (10 ~ 40) V (40 ~ 80) V (80 ~ 100) V (100 ~ 400) V (400 ~ 1 000) V (-) (0 ~ -4) mV (-4 ~ -8) mV (-8 ~ -10) mV (-10 ~ -40) mV	0.78 $\mu$ V $2.0 \times 10^{-4}$ $1.0 \times 10^{-4}$ $8.5 \times 10^{-5}$ $2.6 \times 10^{-5}$ $1.7 \times 10^{-5}$ $1.5 \times 10^{-5}$ $1.7 \times 10^{-5}$ $1.0 \times 10^{-5}$ $9.0 \times 10^{-6}$ $1.6 \times 10^{-5}$ $8.9 \times 10^{-6}$ $7.6 \times 10^{-6}$ $1.7 \times 10^{-5}$ $1.0 \times 10^{-5}$ $8.8 \times 10^{-6}$ $2.0 \times 10^{-5}$ $1.1 \times 10^{-5}$ $2.0 \times 10^{-4}$ $1.0 \times 10^{-4}$ $8.5 \times 10^{-5}$ $2.6 \times 10^{-5}$	Current calibrators, Multimeter calibrators /HCT-CS-197-40112

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC Voltage	40112	(-40 ~ -80) mV (-80 ~ -100) mV (-100 ~ -400) mV (-400 ~ -800) mV (-0.8 ~ -1) V (-1 ~ -4) V (-4 ~ -8) V (-8 ~ -10) V (-10 ~ -40) V (-40 ~ -80) V (-80 ~ -100) V (-100 ~ -400) V (-400 ~ -1 000) V	$1.7 \times 10^{-5}$ $1.5 \times 10^{-5}$ $1.7 \times 10^{-5}$ $1.0 \times 10^{-5}$ $9.0 \times 10^{-6}$ $1.6 \times 10^{-5}$ $8.9 \times 10^{-6}$ $7.6 \times 10^{-6}$ $1.7 \times 10^{-5}$ $1.0 \times 10^{-5}$ $8.8 \times 10^{-6}$ $2.0 \times 10^{-5}$ $1.1 \times 10^{-5}$	
Static/Ionic voltmeters Static Voltage (Positive)  Static Voltage (Nagative)	40113	(+) 10 V (10 ~ 100) V (100 ~ 500) V (0.5 ~ 1) kV (1 ~ 10) kV (10 ~ 15) kV (15 ~ 20) kV (20 ~ 25) kV (25 ~ 48) kV  (-) -10 V (-10 ~ -100) V (-100 ~ -500) V (-0.5 ~ -1) kV (-1 ~ -10) kV (-10 ~ -15) kV (-15 ~ -20) kV (-20 ~ -25) kV (-25 ~ -48) kV	62 mV $6.2 \times 10^{-4}$ $1.2 \times 10^{-4}$ $6.2 \times 10^{-4}$ $1.4 \times 10^{-3}$ $1.6 \times 10^{-3}$ $1.2 \times 10^{-3}$ $1.4 \times 10^{-3}$ $1.3 \times 10^{-3}$  62 mV $6.2 \times 10^{-4}$ $1.2 \times 10^{-4}$ $6.2 \times 10^{-4}$ $1.3 \times 10^{-3}$ $1.5 \times 10^{-3}$ $1.2 \times 10^{-3}$ $1.4 \times 10^{-3}$ $1.3 \times 10^{-3}$	Multimeter calibrators, High voltage generators / HCT-CS-058-40113

402. Resistance, capacitance and inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Capacitance bridges/indicators	40201	10 Hz	5.8 mHz	Counters, Standard capacitors, Digital Multimeters /HCT-CS-059-40201
Frequency		(10 ~ 100) Hz	$5.8 \times 10^{-5}$	
		(0.1 ~ 1) kHz	$5.8 \times 10^{-6}$	
		(1 ~ 10) kHz	$5.8 \times 10^{-7}$	
		(10 ~ 100)kHz	$5.9 \times 10^{-8}$	
		(0.1 ~ 1) MHz	$8.2 \times 10^{-9}$	
		(1 ~ 10) MHz	$5.9 \times 10^{-8}$	
		(10 ~ 30) MHz	$2.0 \times 10^{-8}$	
AC Voltage		100 mV		
		20 Hz	21 $\mu$ V	
		(0.02 ~ 1) kHz	$1.9 \times 10^{-4}$	
		(1 ~ 10) kHz	$2.3 \times 10^{-4}$	
		(10 ~ 100) kHz	$1.1 \times 10^{-3}$	
		100 kHz ~ 1 MHz	$2.5 \times 10^{-1}$	
		(0.1 ~ 1) V		
		20 Hz	$6.4 \times 10^{-4}$	
		(0.02 ~ 10) kHz	$6.0 \times 10^{-4}$	
		(10 ~ 100) kHz	$1.0 \times 10^{-3}$	
		(0.1 ~ 1) MHz	$3.5 \times 10^{-2}$	
		(1 ~ 10) V		
		20 Hz	$6.4 \times 10^{-4}$	
		(0.02 ~ 10) kHz	$5.9 \times 10^{-4}$	
		(10 ~ 100) kHz	$1.0 \times 10^{-3}$	
		(0.1 ~ 1) MHz	$3.5 \times 10^{-2}$	
		(10 ~ 20) V		
		20 Hz	$9.9 \times 10^{-4}$	
(0.02 ~ 1) kHz		$3.4 \times 10^{-4}$		
(1 ~ 10) kHz		$4.4 \times 10^{-4}$		
(10 ~ 100) kHz		$1.8 \times 10^{-3}$		
Capacitance		1 pF		
		60 Hz	0.76 fF	
		(60 ~ 400) Hz	0.75 fF	
		(0.4 ~ 1) kHz	0.76 fF	
		(0.001 ~ 1) MHz	0.76 fF	
		(1 ~ 2) MHz	0.78 fF	
		(2 ~ 3) MHz	0.86 fF	
	(3 ~ 4) MHz	0.98 fF		
	(4 ~ 5) MHz	1.2 fF		
	(5 ~ 10) MHz	2.7 fF		
	(10 ~ 13) MHz	3.8 fF		
	(1 ~ 10) pF			
	60 Hz ~ 5 MHz	3.6 fF		
	(5 ~ 10) MHz	3.8 fF		
	(10 ~ 13) MHz	3.9 fF		
	(10 ~ 100) pF			
	(60 ~ 400) Hz	35 fF		
	400 Hz ~ 4 MHz	36 fF		
(4 ~ 5) MHz	38 fF			
(5 ~ 10) MHz	48 fF			
(10 ~ 13) MHz	61 fF			

402. Resistance, capacitance and inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Capacitance	40201	(100 ~ 1 000) pF (60 ~ 400) Hz 400 Hz ~ 1 MHz (1 ~ 2) MHz (2 ~ 3) MHz (3 ~ 4) MHz (4 ~ 5) MHz (5 ~ 10) MHz (10 ~ 13) MHz  (1 ~ 10) nF 60 Hz 120 Hz ~ 100 kHz  (10 ~ 100) nF 60 Hz 120 Hz ~ 100 kHz  (0.1 ~ 1) μF 60 Hz 120 Hz ~ 10 kHz (10 ~ 100) kHz  (1 ~ 10) μF 100 Hz (0.1 ~ 1) kHz  (10 ~ 100) μF 100 Hz (0.1 ~ 1) kHz  (0.1 ~ 1) mF 100 Hz (0.1 ~ 1) kHz	0.35 pF 0.36 pF 0.38 pF 0.45 pF 0.57 pF 0.72 pF 2.0 pF 2.9 pF  1.4 pF 0.82 pF  36 pF 8.2 pF  0.66 nF 0.11 nF 0.13 nF  4.7 nF 3.2 nF  77 nF 71 nF  1.2 μF 2.3 μF	
Decade capacitors	40202	1 kHz 1 pF (1 ~ 1 000) pF (1 ~ 10) nF (10 ~ 100) nF (100 ~ 1 000) nF (1 ~ 10) μF  120 Hz 10 μF (10 ~ 100) μF (0.1 ~ 1) mF	0.062 fF $6.0 \times 10^{-5}$ $1.1 \times 10^{-4}$ $2.6 \times 10^{-4}$ $5.1 \times 10^{-4}$ $1.7 \times 10^{-3}$  14 nF $2.0 \times 10^{-3}$ $2.2 \times 10^{-3}$	Capacitance bridges LCR meters / HCT-CS-060-40202
Standard capacitors	40204	1 kHz 1 pF 10 pF 100 pF 1 nF 10 nF 100 nF 1 μF 10 μF  120 Hz 100 μF 1 mF	12 aF 0.12 fF 1.2 fF 12 fF 0.21 pF 5.1 pF 0.12 nF 12 nF  0.12 μF 1.3 μF	Capacitance bridges LCR meters / HCT-CS-061-40204

402. Resistance, capacitance and inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Earth testers Earth resistance	40205	0.01 Ω (0.01 ~ 0.1) Ω (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) kΩ (1 ~ 10) kΩ (10 ~ 100) kΩ	59 μΩ $3.0 \times 10^{-3}$ $3.2 \times 10^{-4}$ $3.0 \times 10^{-4}$ $2.9 \times 10^{-4}$ $2.9 \times 10^{-4}$ $2.9 \times 10^{-4}$ $2.9 \times 10^{-4}$	Decade resistor, Standard resistances /HCT-CS-062-40205
Voltage		50 Hz ~ 1 kHz 1 V (1 ~ 100) V (100 ~ 600) V (600 ~ 1 000) V	5.8 mV $5.9 \times 10^{-4}$ $4.5 \times 10^{-4}$ $3.6 \times 10^{-4}$	
Current		50 Hz 1 A (1 ~ 10) A (10 ~ 50) A (50 ~ 100) A  50 Hz ~ 1 kHz 1 A (1 ~ 10) A (10 ~ 50) A (50 ~ 100) A	2.4 mA $2.4 \times 10^{-3}$ $2.3 \times 10^{-3}$ $2.3 \times 10^{-3}$  2.5 mA $2.4 \times 10^{-3}$ $1.8 \times 10^{-3}$ $1.8 \times 10^{-3}$	
Inductors	40208	1 kHz 0.1 mH (0.1 ~ 1) mH 1 mH ~ 1 H	0.12 μH $1.2 \times 10^{-3}$ $6.3 \times 10^{-4}$	LCR meters / HCT-CS-063-40208
Insulation testers Insulation Resistance	40210	1 kΩ 1 kΩ ~ 1 MΩ 1 MΩ ~ 10 MΩ 10 MΩ ~ 100 MΩ 100 MΩ ~ 1 GΩ 1 GΩ ~ 10 GΩ 10 GΩ ~ 100 GΩ 100 GΩ ~ 1 TΩ	0.62 Ω $6.2 \times 10^{-4}$ $3.3 \times 10^{-4}$ $6.0 \times 10^{-4}$ $1.1 \times 10^{-3}$ $2.2 \times 10^{-3}$ $3.9 \times 10^{-3}$ $7.0 \times 10^{-3}$	High resistance meters, Multimeter calibrators /HCT-CS-064-40210
Insulation voltage		25 V (25 ~ 800) V (0.8 ~ 1) kV (1 ~ 10) kV	5.8 mV $5.8 \times 10^{-4}$ $1.2 \times 10^{-2}$ $7.3 \times 10^{-3}$	
AC Voltage		50 Hz ~ 1 kHz 10 V 10 V ~ 600 V	5.9 mV $5.8 \times 10^{-4}$	
Resistance		1 Ω 1 Ω ~ 100 Ω 100 Ω ~ 100 kΩ	5.8 mΩ $5.9 \times 10^{-4}$ $5.8 \times 10^{-4}$	
DC Voltage		1 V 1 V ~ 1 000 V	0.58 mV $5.8 \times 10^{-5}$	

402. Resistance, capacitance and inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Q-meters  AC Voltage    Frequency	40211	1 kHz 10 mV 100 mV 1 V 10 V  1 kHz (1 ~ 10) kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (1 ~ 10) MHz (10 ~ 100) MHz	 58 μV 0.58 mV 0.59 mV 5.9 mV  5.8 mHz $5.8 \times 10^{-7}$ $5.9 \times 10^{-8}$ $1.1 \times 10^{-8}$ $5.9 \times 10^{-8}$ $1.1 \times 10^{-8}$	Frequency counters, Digital multimeters /HCT-CS-065-40211
Resistance bridges & Similar instrument  MEASURING ARM    RATIO ARM	40213	0.01 Ω (0.01 ~ 0.1) Ω (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) kΩ (1 ~ 10) kΩ (10 ~ 100) kΩ (0.1 ~ 1) MΩ (1 ~ 10) MΩ (10 ~ 100) MΩ  × 0.001 × 0.01 × 0.1 × 1 × 10 × 100 × 1 000	9.6 μΩ 11 μΩ 19 μΩ 0.11 mΩ 0.92 mΩ 9.1 mΩ 90 mΩ 1.0 Ω 11 Ω 0.24 kΩ 16 kΩ  $5.9 \times 10^{-8}$ $5.9 \times 10^{-7}$ $5.9 \times 10^{-6}$ $5.9 \times 10^{-5}$ $5.9 \times 10^{-4}$ $5.9 \times 10^{-3}$ $6.0 \times 10^{-2}$	Standard resistance, Digital multimeters /HCT-CS-066-40213
Resistance meters  DC Resistance          Frequency	40214	25 μΩ 50 μΩ 100 μΩ 1 mΩ (1 ~ 10) mΩ (10 ~ 100) mΩ (100 ~ 1 000) mΩ (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 10) kΩ (10 ~ 100) kΩ (0.1 ~ 1) MΩ (1 ~ 10) MΩ (10 ~ 100) MΩ (0.1 ~ 1) GΩ (1 ~ 10) GΩ (10 ~ 100) GΩ (0.1 ~ 1) TΩ (1 ~ 10) TΩ  10 Hz (10 ~ 100) Hz (0.1 ~ 1) kHz (1 ~ 10) kHz (10 ~ 100) kHz (0.1 ~ 1) MHz	0.2 μΩ 0.4 μΩ 0.7 μΩ 0.25 μΩ $1.2 \times 10^{-4}$ $6.8 \times 10^{-6}$ $3.3 \times 10^{-6}$ $6.7 \times 10^{-5}$ $3.7 \times 10^{-5}$ $3.1 \times 10^{-5}$ $3.7 \times 10^{-5}$ $4.7 \times 10^{-5}$ $8.7 \times 10^{-5}$ $3.7 \times 10^{-4}$ $7.3 \times 10^{-4}$ $1.8 \times 10^{-3}$ $3.1 \times 10^{-3}$ $7.0 \times 10^{-3}$ $1.9 \times 10^{-1}$  5.8 mHz $5.8 \times 10^{-5}$ $5.8 \times 10^{-6}$ $5.8 \times 10^{-7}$ $5.9 \times 10^{-8}$ $8.2 \times 10^{-9}$	Standard resistances, High resistance meters, Digital multimeters, Counters HCT-CS-067-40214



402. Resistance, capacitance and inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC Voltage	40214	1 kHz	7.2 μV	
AC Resistance		10 mV	$1.8 \times 10^{-4}$	
		(10 ~ 100) mV	$1.1 \times 10^{-4}$	
		(0.1 ~ 1) V	$1.1 \times 10^{-4}$	
		(1 ~ 10) V	$1.2 \times 10^{-4}$	
		(10 ~ 100) V	$1.4 \times 10^{-4}$	
		(100 ~ 1 000) V		
DC Voltage		1 kHz	1.4 μΩ	
		1 mΩ	12 μΩ	
		10 mΩ	0.12 mΩ	
		100 mΩ	9.1 mΩ	
		1 Ω	3.6 mΩ	
		10 Ω	35 mΩ	
		100 Ω	0.36 Ω	
		1 kΩ	3.6 Ω	
10 kΩ	36 Ω			
100 kΩ				
100 mV	0.79 μV			
(0.1 ~ 1) V	$7.3 \times 10^{-6}$			
(1 ~ 10) V	$7.2 \times 10^{-6}$			
(10 ~ 100) V	$8.1 \times 10^{-6}$			
(100 ~ 1 000) V	$9.0 \times 10^{-6}$			
Resistors	40215	1 mΩ	18 nΩ	Digital multimeters, LCR meters / HCT-CS-068-40215
Standard Resistance(DC)		(1 ~ 10) mΩ	$1.8 \times 10^{-5}$	
		(10 ~ 100) mΩ	$1.7 \times 10^{-5}$	
		(0.1 ~ 1) Ω	$4.9 \times 10^{-6}$	
		(1 ~ 10) Ω	$1.8 \times 10^{-6}$	
		(10 ~ 25) Ω	$5.2 \times 10^{-6}$	
		(25 ~ 100) Ω	$5.9 \times 10^{-6}$	
		(0.1 ~ 1) kΩ	$5.3 \times 10^{-6}$	
		(1 ~ 10) kΩ	$2.7 \times 10^{-6}$	
		(10 ~ 100) kΩ	$7.8 \times 10^{-6}$	
		(0.1 ~ 1) MΩ	$1.2 \times 10^{-5}$	
		(1 ~ 10) MΩ	$2.2 \times 10^{-5}$	
		(10 ~ 100) MΩ	$2.7 \times 10^{-5}$	
		(0.1 ~ 1) GΩ	$4.1 \times 10^{-5}$	
		(1 ~ 10) GΩ	$3.5 \times 10^{-4}$	
Standard Resistance(AC)	1 mΩ	1.5 μΩ		
	50 Hz	$1.5 \times 10^{-3}$		
	(0.05 ~ 1) kHz			
	(1 ~ 10) mΩ	$1.5 \times 10^{-3}$		
	50 Hz	$1.5 \times 10^{-3}$		
	(0.05 ~ 1) kHz			
	(10 ~ 100) mΩ	$8.3 \times 10^{-3}$		
	50 Hz	$8.3 \times 10^{-3}$		
	(0.05 ~ 1) kHz			
	(0.1 ~ 1) Ω	$1.2 \times 10^{-3}$		
400 Hz	$1.2 \times 10^{-3}$			
(0.4 ~ 1) kHz				

402. Resistance, capacitance and inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.		
Standard Resistance(AC)	40215	(1 ~ 10) Ω				
		400 Hz	$3.8 \times 10^{-4}$			
		(0.4 ~ 1) kHz	$3.8 \times 10^{-4}$			
		(0.001 ~ 1) MHz	$1.2 \times 10^{-2}$			
		(1 ~ 2) MHz	$1.2 \times 10^{-2}$			
		(2 ~ 3) MHz	$1.2 \times 10^{-2}$			
		(3 ~ 4) MHz	$1.2 \times 10^{-2}$			
		(4 ~ 5) MHz	$1.2 \times 10^{-2}$			
		(5 ~ 10) MHz	$1.3 \times 10^{-2}$			
		(10 ~ 13) MHz	$1.4 \times 10^{-2}$			
		(10 ~100) Ω				
		400 Hz	$3.6 \times 10^{-4}$			
		(0.4 ~ 1) kHz	$3.6 \times 10^{-4}$			
		(0.001 ~ 1) MHz	$1.2 \times 10^{-2}$			
		(1 ~ 2) MHz	$1.2 \times 10^{-2}$			
		(2 ~ 3) MHz	$1.2 \times 10^{-2}$			
		(3 ~ 4) MHz	$1.2 \times 10^{-2}$			
		(4 ~ 5) MHz	$1.2 \times 10^{-2}$			
		(5 ~ 10) MHz	$1.2 \times 10^{-2}$			
		(10 ~ 13) MHz	$1.2 \times 10^{-2}$			
		(0.1 ~ 1) kΩ				
		400 Hz	$3.6 \times 10^{-4}$			
		(0.4 ~ 1) kHz	$3.6 \times 10^{-4}$			
		(1 ~ 100) kHz	$1.2 \times 10^{-2}$			
		(0.1 ~ 1) MHz	$1.2 \times 10^{-2}$			
		(1 ~ 2) MHz	$1.2 \times 10^{-2}$			
		(2 ~ 3) MHz	$1.2 \times 10^{-2}$			
		(3 ~ 4) MHz	$1.2 \times 10^{-2}$			
		(4 ~ 5) MHz	$1.2 \times 10^{-2}$			
		(5 ~ 10) MHz	$1.2 \times 10^{-2}$			
		(10 ~ 13) MHz	$1.2 \times 10^{-2}$			
		(1 ~ 10) kΩ				
		400 Hz	$3.6 \times 10^{-4}$			
		(0.4 ~ 1) kHz	$3.6 \times 10^{-4}$			
		(1 ~ 100) kHz	$1.2 \times 10^{-2}$			
		(0.1 ~ 1) MHz	$1.2 \times 10^{-2}$			
		(10 ~ 100) kΩ				
		400 Hz	$3.7 \times 10^{-4}$			
		(0.4 ~ 1) kHz	$3.7 \times 10^{-4}$			
		(1 ~ 100) kHz	$1.2 \times 10^{-2}$			
		(0.1 ~ 1) MHz	$1.2 \times 10^{-2}$			
		Decade Resistance(DC)		1 mΩ	7.5 uΩ	
				1 mΩ ~ 10 mΩ	$3.8 \times 10^{-3}$	
				10 mΩ ~ 100 mΩ	$3.8 \times 10^{-4}$	
				100 mΩ ~ 1 Ω	$3.8 \times 10^{-4}$	
				1 Ω ~ 10 Ω	$6.5 \times 10^{-5}$	
				10 Ω ~ 100 Ω	$2.1 \times 10^{-5}$	
				100 Ω ~ 1 kΩ	$2.3 \times 10^{-5}$	
1 kΩ ~ 10 kΩ	$2.2 \times 10^{-5}$					
10 kΩ ~ 100 kΩ	$1.9 \times 10^{-5}$					
100 kΩ ~ 1 MΩ	$1.6 \times 10^{-5}$					
1 MΩ ~ 10 MΩ	$4.0 \times 10^{-5}$					
10 MΩ ~100 MΩ	$1.3 \times 10^{-4}$					
100 MΩ ~ 1 GΩ	$7.0 \times 10^{-4}$					
1 GΩ ~ 10 GΩ	$1.2 \times 10^{-3}$					
10 GΩ ~ 100 GΩ	$3.1 \times 10^{-3}$					
100 GΩ ~ 1 TΩ	$4.7 \times 10^{-3}$					

402. Resistance, capacitance and inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Decade Resistance(AC)	40215	1 kHz 100 mΩ 100 mΩ ~ 10 Ω 10 Ω ~ 100 kΩ	0.32 mΩ $1.2 \times 10^{-3}$ $6.5 \times 10^{-4}$	
Electrical conductivity meters	40216	14.36 MS/m 22.90 MS/m 34.26 MS/m 58.38 MS/m	0.081 MS/m 0.16 MS/m 0.21 MS/m 0.32 MS/m	Standard specimens /HCT-CS-227-40216
Impedance bridges/LCR meters	40217			Counters, Standard Resistance, Capacitance, Inductance, Digital multimeters /HCT-CS-093-40217
Frequency		10 Hz 10 Hz ~ 100 Hz 100 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz (0.1 ~ 1) MHz (1 ~ 10) MHz (10 ~ 30) MHz	5.8 mHz $5.8 \times 10^{-5}$ $5.8 \times 10^{-6}$ $5.8 \times 10^{-7}$ $5.9 \times 10^{-8}$ $8.7 \times 10^{-9}$ $5.8 \times 10^{-9}$ $2.0 \times 10^{-8}$	
AC Voltage		1 mV 20 Hz (0.02 ~ 1) kHz (1 ~ 10) kHz (10 ~ 100) kHz  (1 ~ 10) mV 20 Hz (0.02 ~ 1) kHz (1 ~ 10) kHz (10 ~ 100) kHz  (10 ~ 100) mV 20 Hz (0.02 ~ 1) kHz (1 ~ 10) kHz (10 ~ 100) kHz  (0.1 ~ 1) V 20 Hz (0.02 ~ 10) kHz (10 ~ 100) kHz (0.1 ~ 1) MHz  (1 ~ 10) V 20 Hz (0.02 ~ 10) kHz (10 ~ 100) kHz (0.1 ~ 1) MHz  (10 ~ 20) V 20 Hz (0.02 ~ 1) kHz (1 ~ 10) kHz (10 ~ 100) kHz	7 uV $6.0 \times 10^{-3}$ $1.1 \times 10^{-2}$ $3.0 \times 10^{-2}$  $8.0 \times 10^{-4}$ $7.0 \times 10^{-4}$ $1.2 \times 10^{-3}$ $4.0 \times 10^{-3}$  $2.0 \times 10^{-4}$ $1.8 \times 10^{-4}$ $2.2 \times 10^{-4}$ $1.1 \times 10^{-3}$  $6.4 \times 10^{-4}$ $5.9 \times 10^{-4}$ $1.0 \times 10^{-3}$ $3.5 \times 10^{-2}$  $6.4 \times 10^{-4}$ $5.9 \times 10^{-4}$ $1.0 \times 10^{-3}$ $3.5 \times 10^{-2}$  $8.8 \times 10^{-4}$ $3.4 \times 10^{-4}$ $3.7 \times 10^{-4}$ $1.8 \times 10^{-3}$	
DC Voltage		100 mV 100 mV ~ 10 V (10 ~ 40) V	0.8 uV $1.6 \times 10^{-3}$ $5.8 \times 10^{-5}$	

402. Resistance, capacitance and inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC Current	40217	1 A	0.63 mA	
		1 A ~ 10 A	$5.0 \times 10^{-4}$	
		10 A ~ 20 A	$1.0 \times 10^{-3}$	
		20 A ~ 40 A	$1.0 \times 10^{-3}$	
Resistance	40217	1 mΩ		
		50 Hz	1.4 uΩ	
		50 Hz ~ 1 kHz	$1.4 \times 10^{-3}$	
		(1 ~ 10) mΩ		
		50 Hz	$1.2 \times 10^{-3}$	
		50 Hz ~ 1 kHz	$1.2 \times 10^{-3}$	
		(10 ~ 100) mΩ		
		50 Hz	$1.2 \times 10^{-3}$	
		50 Hz ~ 1 kHz	$1.2 \times 10^{-3}$	
		(0.1 ~ 1) Ω		
		400 Hz	$1.2 \times 10^{-3}$	
		400 Hz ~ 1 kHz	$1.2 \times 10^{-3}$	
Capacitance	40217	(1 ~ 10) Ω		
		400 Hz	$6.0 \times 10^{-4}$	
		400 Hz ~ 1 kHz	$6.0 \times 10^{-4}$	
		1 kHz ~ 5 MHz	$1.2 \times 10^{-2}$	
		(5 ~ 10) MHz	$1.3 \times 10^{-2}$	
		(10 ~ 13) MHz	$1.4 \times 10^{-2}$	
		(10 ~ 100) Ω		
		400 Hz	$3.5 \times 10^{-4}$	
		400 Hz ~ 1 kHz	$3.5 \times 10^{-4}$	
		1 kHz ~ 13 MHz	$1.2 \times 10^{-2}$	
		100 Ω ~ 1 kΩ		
		400 Hz	$3.6 \times 10^{-4}$	
400 Hz ~ 1 kHz	$3.6 \times 10^{-4}$			
1 kHz ~ 13 MHz	$1.2 \times 10^{-2}$			
Capacitance	40217	(1 ~ 10) kΩ		
		400 Hz	$3.6 \times 10^{-4}$	
		400 Hz ~ 1 kHz	$3.6 \times 10^{-4}$	
		1 kHz ~ 1 MHz	$1.2 \times 10^{-2}$	
		(10 ~ 100) kΩ		
		400 Hz	$3.6 \times 10^{-4}$	
		400 Hz ~ 1 kHz	$3.6 \times 10^{-4}$	
		1 kHz ~ 1 MHz	$1.2 \times 10^{-2}$	
		1 pF		
		60 Hz	0.76 fF	
		(60 ~ 400) Hz	$7.5 \times 10^{-4}$	
		(0.4 ~ 1) kHz	$7.6 \times 10^{-4}$	
(0.001 ~ 1) MHz	$7.6 \times 10^{-4}$			
(1 ~ 2) MHz	$7.8 \times 10^{-4}$			
(2 ~ 3) MHz	$8.6 \times 10^{-4}$			
(3 ~ 4) MHz	$9.8 \times 10^{-4}$			
(4 ~ 5) MHz	$1.2 \times 10^{-3}$			
(5 ~ 10) MHz	$2.6 \times 10^{-3}$			
(10 ~ 13) MHz	$3.8 \times 10^{-3}$			

402. Resistance, capacitance and inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Capacitance	40217	(1 ~ 10) pF		
		60 Hz	$3.6 \times 10^{-4}$	
		(60 ~ 400) Hz	$3.6 \times 10^{-4}$	
		(0.4 ~ 1) kHz	$3.6 \times 10^{-4}$	
		(0.001 ~ 1) MHz	$3.6 \times 10^{-4}$	
		(1 ~ 2) MHz	$3.6 \times 10^{-4}$	
		(2 ~ 3) MHz	$3.6 \times 10^{-4}$	
		(3 ~ 4) MHz	$3.6 \times 10^{-4}$	
		(4 ~ 5) MHz	$3.6 \times 10^{-4}$	
		(5 ~ 10) MHz	$3.8 \times 10^{-4}$	
		(10 ~ 13) MHz	$3.9 \times 10^{-4}$	
		(10 ~ 100) pF		
		60 Hz	$3.5 \times 10^{-4}$	
		(60 ~ 400) Hz	$3.5 \times 10^{-4}$	
		(0.4 ~ 1) kHz	$3.6 \times 10^{-4}$	
		(0.001 ~ 1) MHz	$3.6 \times 10^{-4}$	
		(1 ~ 2) MHz	$3.6 \times 10^{-4}$	
		(2 ~ 3) MHz	$3.6 \times 10^{-4}$	
		(3 ~ 4) MHz	$3.6 \times 10^{-4}$	
		(4 ~ 5) MHz	$3.8 \times 10^{-4}$	
		(5 ~ 10) MHz	$4.9 \times 10^{-4}$	
		(10 ~ 13) MHz	$6.1 \times 10^{-4}$	
		(100 ~ 1 000) pF		
		60 Hz	$3.5 \times 10^{-4}$	
		(60 ~ 400) Hz	$3.5 \times 10^{-4}$	
		(0.4 ~ 1) kHz	$3.6 \times 10^{-4}$	
		(0.001 ~ 1) MHz	$3.6 \times 10^{-4}$	
		(1 ~ 2) MHz	$3.8 \times 10^{-4}$	
		(2 ~ 3) MHz	$4.5 \times 10^{-4}$	
		(3 ~ 4) MHz	$5.7 \times 10^{-4}$	
		(4 ~ 5) MHz	$7.2 \times 10^{-4}$	
		(5 ~ 10) MHz	$2.0 \times 10^{-3}$	
		(10 ~ 13) MHz	$3.0 \times 10^{-3}$	
		(1 ~ 10) nF		
		60 Hz	$1.4 \times 10^{-4}$	
		(60 ~ 120) Hz	$8.2 \times 10^{-5}$	
		(120 ~ 400) Hz	$7.7 \times 10^{-5}$	
		(0.4 ~ 1) kHz	$8.2 \times 10^{-5}$	
		(1 ~ 10) kHz	$8.2 \times 10^{-5}$	
		(10 ~ 100) kHz	$8.2 \times 10^{-5}$	
		(10 ~ 100) nF		
		60 Hz	$3.6 \times 10^{-4}$	
		(60 ~ 120) Hz	$8.2 \times 10^{-5}$	
		(120 ~ 400) Hz	$9.3 \times 10^{-5}$	
		(0.4 ~ 1) kHz	$8.2 \times 10^{-5}$	
		(1 ~ 10) kHz	$8.2 \times 10^{-5}$	
		(10 ~ 100) kHz	$8.2 \times 10^{-5}$	
		(0.1 ~ 1) μF		
60 Hz	$6.6 \times 10^{-4}$			
(60 ~ 120) Hz	$1.1 \times 10^{-4}$			
(120 ~ 400) Hz	$1.8 \times 10^{-4}$			
(0.4 ~ 1) kHz	$1.1 \times 10^{-4}$			
(1 ~ 10) kHz	$1.1 \times 10^{-4}$			
(10 ~ 100) kHz	$1.3 \times 10^{-4}$			
(1 ~ 10) μF				
100 Hz	$4.7 \times 10^{-4}$			
(0.1 ~ 1) kHz	$3.2 \times 10^{-4}$			

402. Resistance, capacitance and inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Capacitance	40217	(10 ~ 100) $\mu$ F		
		100 Hz	$7.7 \times 10^{-4}$	
		(0.1 ~ 1) kHz	$7.1 \times 10^{-4}$	
		(0.1 ~ 1) mF		
		100 Hz	$1.2 \times 10^{-3}$	
		(0.1 ~ 1) kHz	$2.3 \times 10^{-3}$	
Inductance		1 kHz		
		100 $\mu$ H	21 nH	
		(0.1 ~ 1) mH	$6.4 \times 10^{-4}$	
		(1 ~ 10) mH	$5.9 \times 10^{-4}$	
		(10 ~ 100) mH	$1.0 \times 10^{-3}$	
		(0.1 ~ 1) H	$3.5 \times 10^{-2}$	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC ammeters  AC Current	40301	10 $\mu$ A 50 Hz ~ 1 kHz  (10 $\mu$ A ~ 100 $\mu$ A) 40 Hz 40 Hz ~ 1 kHz (1 ~ 10) kHz  (100 $\mu$ A ~ 10 mA) 40 Hz 40 Hz ~ 1 kHz (1 ~ 10) kHz  (10 ~ 100) mA 40 Hz 40 Hz ~ 1 kHz (1 ~ 10) kHz  (100 mA ~ 1 A) 40 Hz 40 Hz ~ 1 kHz (1 ~ 10) kHz  (1 ~ 10) A (50 ~ 60) Hz 60 Hz ~ 1 kHz  (10 ~ 20) A (50 ~ 60) Hz 60 Hz ~ 1 kHz  (20 ~ 50) A (50 ~ 60) Hz  (50 ~ 100) A (50 ~ 60) Hz	12 nA  $3.2 \times 10^{-4}$ $2.3 \times 10^{-4}$ $2.1 \times 10^{-3}$  $2.4 \times 10^{-4}$ $1.8 \times 10^{-4}$ $2.1 \times 10^{-3}$  $2.4 \times 10^{-4}$ $1.8 \times 10^{-4}$ $2.0 \times 10^{-3}$  $3.5 \times 10^{-4}$ $3.4 \times 10^{-4}$ $8.4 \times 10^{-3}$  $1.3 \times 10^{-3}$ $2.0 \times 10^{-3}$  $7.1 \times 10^{-4}$ $1.5 \times 10^{-3}$  $6.2 \times 10^{-4}$  $4.1 \times 10^{-4}$	Multimeter calibrators / HCT-CS-070-40301
AC Voltage		1 V 40 Hz 40 Hz ~ 10 kHz  (1 ~ 10) V 40 Hz 40 Hz ~ 10 kHz	$0.38 \mu$ V $3.6 \times 10^{-4}$  $1.4 \times 10^{-4}$ $6.8 \times 10^{-5}$	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC Voltage	40301	(10 ~ 30) V 40 Hz 40 Hz ~ 1 kHz	$2.1 \times 10^{-4}$ $1.2 \times 10^{-4}$	
		(30 ~ 75) V 40 Hz 40 Hz ~ 1 kHz	$1.5 \times 10^{-4}$ $9.1 \times 10^{-5}$	
		(75 ~ 150) V 40 Hz 40 Hz ~ 1 kHz	$1.3 \times 10^{-4}$ $7.3 \times 10^{-5}$	
		(150 ~ 300) V 50 Hz 50 Hz ~ 1 kHz	$4.3 \times 10^{-4}$ $1.5 \times 10^{-4}$	
		(300 ~ 750) V 50 Hz 50 Hz ~ 1 kHz	$3.9 \times 10^{-4}$ $1.0 \times 10^{-4}$	
Clamp ammeters/voltmeters	40302			Multimeter calibrators, Coil / HCT-CS-071-40302
DC Voltage		100 mV (100 mV ~ 1 V) (1 V ~ 10 V) (10 V ~ 100 V) (100 V ~ 1 000 V)	$6.3 \mu V$ $62 \mu V$ $0.62 mV$ $6.0 mV$ $62 mV$	
AC Voltage		100 mV 40 Hz 40 Hz ~ 20 kHz (20 ~ 50) kHz (50 ~ 100) kHz	$13 \mu V$ $10 \mu V$ $17 \mu V$ $38 \mu V$	
		(100 mV ~ 1 V) 40 Hz 40 Hz ~ 20 kHz (20 ~ 100) kHz	$0.14 mV$ $85 \mu V$ $0.15 mV$	
		(1 ~ 10) V 40 Hz ~ 10 kHz (10 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz	$1.4 mV$ $0.85 mV$ $1.2 mV$ $1.4 mV$	
		(10 ~ 100) V 40 Hz ~ 10 kHz (10 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz	$15 mV$ $9.4 mV$ $13 mV$ $22 mV$	
		(100 ~ 1 000) V 50 Hz 50 Hz ~ 1 kHz	$0.38 V$ $0.11 V$	
DC Current		10 $\mu A$ (10 ~ 100) $\mu A$ 100 $\mu A$ ~ 1 mA (1 ~ 10) mA (10 ~ 100) mA 100 mA ~ 1 A (1 ~ 10) A (10 ~ 50) A	12 nA $0.12 \mu A$ $1.4 \mu A$ $14 \mu A$ $0.14 mA$ $1.4 mA$ $14 mA$ $85 mA$	



403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.			
DC Current	40302	(50 ~ 100) A	0.14 A				
		(100 ~ 200) A	0.25 A				
		(200 ~ 300) A	0.36 A				
		(300 ~ 400) A	0.48 A				
		(400 ~ 500) A	0.59 A				
		(500 ~ 750) A	1.1 A				
		(750 ~ 900) A	1.3 A				
		(900 ~ 1 000) A	1.4 A				
		(1 000 ~ 1 500) A	1.9 A				
		(1 500 ~ 2 000) A	2.5 A				
		(2 000 ~ 2 500) A	3.0 A				
		AC Current	40302		10 $\mu$ A		
					50 Hz ~ 10 kHz	12 nA	
					(10 ~ 100) $\mu$ A		
					50 Hz ~ 1 kHz	0.12 $\mu$ A	
(1 kHz ~ 10 kHz)	0.24 $\mu$ A						
100 $\mu$ A ~ 1 mA							
40 Hz ~ 1 kHz	1.4 $\mu$ A						
(1 kHz ~ 10 kHz)	2.5 $\mu$ A						
(1 ~ 10) mA							
40 Hz ~ 1 kHz	14 $\mu$ A						
(1 kHz ~ 10 kHz)	24 $\mu$ A						
(10 ~ 100) mA							
40 Hz ~ 1 kHz	0.14 mA						
(1 kHz ~ 10 kHz)	0.20 mA						
100 mA ~ 1 A							
40 Hz ~ 1 kHz	1.4 mA						
(1 kHz ~ 10 kHz)	8.5 mA						
(1 ~ 10) A							
(40 ~ 60) Hz	19 mA						
60 Hz ~ 1 kHz	24 mA						
(10 ~ 100) A							
(50 ~ 60) Hz	0.32 A						
(100 ~ 200) A							
(50 ~ 60) Hz	0.39 A						
(200 ~ 300) A							
(50 ~ 60) Hz	0.48 A						
(300 ~ 400) A							
(50 ~ 60) Hz	0.58 A						
(400 ~ 500) A							
(50 ~ 60) Hz	0.68 A						
(500 ~ 750) A							
(50 ~ 60) Hz	1.3 A						
(750 ~ 900) A							
(50 ~ 60) Hz	1.5 A						
(900 ~ 1 000) A							
(50 ~ 60) Hz	1.6 A						

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC Current	40302	(1 000 ~ 1 500) A (50 ~ 60) Hz	2.1 A	
		(1 500 ~ 2 000) A (50 ~ 60) Hz	2.6 A	
		(2 000 ~ 2 500) A (50 ~ 60) Hz	3.2 A	
		(2 500 ~ 3 000) A (50 ~ 60) Hz	3.7 A	
Resistance	40302	1 Ω	0.88 mΩ	
		(1 Ω ~ 10 Ω)	1.2 mΩ	
		(10 Ω ~ 100 Ω)	8.5 mΩ	
		(100 Ω ~ 1 kΩ)	83 mΩ	
		(1 kΩ ~ 10 kΩ)	1.1 Ω	
		(10 kΩ ~ 100 kΩ)	6.9 Ω	
		(100 kΩ ~ 1 MΩ)	0.19 kΩ	
		(1 MΩ ~ 10 MΩ) (10 MΩ ~ 100 MΩ)	1.4 kΩ 17 kΩ	
Frequency	40302	40 Hz	7.8 mHz	
		(40 ~ 50) Hz	8.6 mHz	
		(50 ~ 60) Hz	10 mHz	
		(60 ~ 300) Hz	71 mHz	
		(300 ~ 400) Hz	80 mHz	
		400 Hz ~ 1 kHz	0.14 Hz	
Current Probe	40302	DC		
		10 mA	15 μA	
		(10 ~ 50) mA	59 μA	
		(50 ~ 100) mA	0.12 mA	
		(100 ~ 500) mA	0.59 mA	
		(500 mA ~ 1 A)	1.2 mA	
		(1 ~ 5) A	6.0 mA	
		(5 ~ 10) A	12 mA	
		(10 ~ 20) A	24 mA	
		(20 ~ 40) A	48 mA	
		(40 ~ 60) A	71 mA	
		(60 ~ 80) A	94 mA	
		(80 ~ 100) A	0.12 A	
		(100 ~ 500) A	0.59 A	
		(500 ~ 1 000) A	1.2 A	
		(50 ~ 60) Hz		
		10 mA	79 μA	
		(10 ~ 50) mA	0.10 mA	
		(50 ~ 100) mA	0.15 mA	
		(100 ~ 500) mA	0.63 mA	
		(500 mA ~ 1 A)	1.3 mA	
		(1 ~ 5) A	6.1 mA	
		(5 ~ 10) A	12 mA	
		(10 ~ 20) A	25 mA	
(20 ~ 40) A	47 mA			
(40 ~ 60) A	71 mA			
(60 ~ 80) A	95 mA			
(80 ~ 100) A	0.12 A			
(100 ~ 500) A	0.69 A			
(500 ~ 1 000) A	1.4 A			
(1 000 ~ 3 000) A	3.7 A			

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Current Coil (AC Ratio)	40302	(50 ~ 60) Hz		
		2	0.15 %	
		10	0.06 %	
		25	0.13 %	
		50	0.08 %	
Current Coil (DC Ratio)		2	0.04 %	
		10	0.04 %	
		25	0.13 %	
	50	0.08 %		
AC voltage/current calibrators AC Voltage	40303	0.1 V		Multimeters / HCT-CS-072-40303
		50 Hz	18 μV	
		50 Hz ~ 1 kHz	$1.6 \times 10^{-4}$	
		(1 ~ 10) kHz	$1.8 \times 10^{-4}$	
		(10 ~ 30) kHz	$4.6 \times 10^{-4}$	
		(30 ~ 100) kHz	$1.1 \times 10^{-3}$	
		(0.1 ~ 0.4) V		
		50 Hz	0.077 mV	
		50 Hz ~ 1 kHz	$1.8 \times 10^{-4}$	
		(1 ~ 10) kHz	$1.9 \times 10^{-4}$	
		(10 ~ 30) kHz	$4.0 \times 10^{-4}$	
		(30 ~ 100) kHz	$1.2 \times 10^{-3}$	
		(0.4 ~ 0.8) V		
		50 Hz	0.11 mV	
		50 Hz ~ 1 kHz	$1.3 \times 10^{-4}$	
		(1 ~ 10) kHz	$1.4 \times 10^{-4}$	
		(10 ~ 30) kHz	$3.1 \times 10^{-4}$	
		(30 ~ 100) kHz	$8.8 \times 10^{-4}$	
		(0.8 ~ 1) V		
		50 Hz	0.13 mV	
		50 Hz ~ 1 kHz	$1.1 \times 10^{-4}$	
		(1 ~ 10) kHz	$1.3 \times 10^{-4}$	
		(10 ~ 30) kHz	$2.9 \times 10^{-4}$	
		(30 ~ 100) kHz	$8.2 \times 10^{-4}$	
		(1 ~ 4) V		
		50 Hz	0.77 mV	
		(50 ~ 300) Hz	$1.8 \times 10^{-4}$	
		300 Hz ~ 10 kHz	$2.0 \times 10^{-4}$	
		(10 ~ 30) kHz	$4.0 \times 10^{-4}$	
		(30 ~ 100) kHz	$1.2 \times 10^{-3}$	
		(4 ~ 8) V		
		50 Hz	1.1 mV	
		50 Hz ~ 1 kHz	$1.3 \times 10^{-4}$	
		(1 ~ 10) kHz	$1.4 \times 10^{-4}$	
		(10 ~ 30) kHz	$3.1 \times 10^{-4}$	
		(30 ~ 100) kHz	$8.8 \times 10^{-4}$	
		(8 ~ 10) V		
		50 Hz	1.3 mV	
		50 Hz ~ 1 kHz	$1.2 \times 10^{-4}$	
		(1 ~ 10) kHz	$1.3 \times 10^{-4}$	
		(10 ~ 30) kHz	$2.9 \times 10^{-4}$	
		(30 ~ 100) kHz	$8.2 \times 10^{-4}$	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC Voltage	40303	(10 ~ 40) V		
		50 Hz	8.1 mV	
		50 Hz ~ 1 kHz	$1.9 \times 10^{-4}$	
		(1 ~ 10) kHz	$2.0 \times 10^{-4}$	
		(10 ~ 30) kHz	$4.3 \times 10^{-4}$	
		(30 ~ 100) kHz	$1.2 \times 10^{-3}$	
		(40 ~ 80) V		
		50 Hz	12 mV	
		50 Hz ~ 1 kHz	$1.3 \times 10^{-4}$	
		(1 ~ 10) kHz	$1.5 \times 10^{-4}$	
		(10 ~ 30) kHz	$3.1 \times 10^{-4}$	
		(30 ~ 100) kHz	$8.8 \times 10^{-4}$	
		(80 ~ 100) V		
		50 Hz	13 mV	
		50 Hz ~ 1 kHz	$1.1 \times 10^{-4}$	
		(1 ~ 10) kHz	$1.3 \times 10^{-4}$	
		(10 ~ 30) kHz	$3.0 \times 10^{-4}$	
		(30 ~ 100) kHz	$8.2 \times 10^{-4}$	
		(100 ~ 400) V		
		50 Hz	0.082 V	
		(50 ~ 300) Hz	$2.1 \times 10^{-4}$	
		300 Hz ~ 1 kHz	$2.1 \times 10^{-4}$	
		(1 ~ 10) kHz	$2.2 \times 10^{-4}$	
		(400 ~ 800) V		
50 Hz	0.15 V			
(50 ~ 300) Hz	$1.9 \times 10^{-4}$			
300 Hz ~ 1 kHz	$1.5 \times 10^{-4}$			
(1 ~ 10) kHz	$1.9 \times 10^{-4}$			
(800 ~ 1 000) V				
50 Hz	0.16 V			
(50 ~ 300) Hz	$1.6 \times 10^{-4}$			
300 Hz ~ 1 kHz	$1.4 \times 10^{-4}$			
(1 ~ 10) kHz	$1.6 \times 10^{-4}$			
AC Current	40303	100 $\mu$ A		
		50 Hz	71 nA	
		50 Hz ~ 1 kHz	$7.0 \times 10^{-4}$	
		(1 ~ 10) kHz	$7.1 \times 10^{-4}$	
		(0.1 ~ 0.4) mA		
		50 Hz	0.10 $\mu$ A	
		50 Hz ~ 1 kHz	$2.3 \times 10^{-4}$	
		(1 ~ 10) kHz	$2.5 \times 10^{-4}$	
		(0.4 ~ 0.8) mA		
		50 Hz	0.14 $\mu$ A	
		50 Hz ~ 1 kHz	$1.5 \times 10^{-4}$	
		(1 ~ 10) kHz	$1.8 \times 10^{-4}$	
		(0.8 ~ 1) mA		
		50 Hz	0.16 $\mu$ A	
		50 Hz ~ 1 kHz	$1.4 \times 10^{-4}$	
		(1 ~ 10) kHz	$1.7 \times 10^{-4}$	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
AC Current	40303	(1 ~ 4) mA			
		50 Hz	0.90 μA		
		50 Hz ~ 1 kHz	$2.1 \times 10^{-4}$		
		(1 ~ 10) kHz	$2.3 \times 10^{-4}$		
		(4 ~ 8) mA			
		50 Hz	1.3 μA		
		50 Hz ~ 1 kHz	$1.4 \times 10^{-4}$		
		(1 ~ 10) kHz	$1.6 \times 10^{-4}$		
		(8 ~ 10) mA			
		50 Hz	1.4 μA		
		50 Hz ~ 1 kHz	$1.2 \times 10^{-4}$		
		(1 ~ 10) kHz	$1.4 \times 10^{-4}$		
		(10 ~ 40) mA			
		50 Hz	8.9 μA		
		50 Hz ~ 1 kHz	$2.1 \times 10^{-4}$		
		(1 ~ 10) kHz	$2.2 \times 10^{-4}$		
		(40 ~ 80) mA			
		50 Hz	12 μA		
		50 Hz ~ 1 kHz	$1.3 \times 10^{-4}$		
		(1 ~ 10) kHz	$1.5 \times 10^{-4}$		
		(80 ~ 100) mA			
		50 Hz	14 μA		
		50 Hz ~ 1 kHz	$1.2 \times 10^{-4}$		
		(1 ~ 10) kHz	$1.4 \times 10^{-4}$		
		(0.1 ~ 0.4) A			
		50 Hz	90 μA		
		50 Hz ~ 1 kHz	$2.1 \times 10^{-4}$		
		(1 ~ 10) kHz	$2.3 \times 10^{-4}$		
		(0.4 ~ 0.8) A			
		50 Hz	0.12 mA		
		50 Hz ~ 1 kHz	$1.4 \times 10^{-4}$		
		(1 ~ 10) kHz	$1.5 \times 10^{-4}$		
		(0.8 ~ 1) A			
		50 Hz	0.14 mA		
		50 Hz ~ 1 kHz	$1.2 \times 10^{-4}$		
		(1 ~ 10) kHz	$1.4 \times 10^{-4}$		
		(1 ~ 4) A			
		50 Hz	0.91 mA		
		50 Hz ~ 1 kHz	$2.1 \times 10^{-4}$		
		(1 ~ 10) kHz	$2.3 \times 10^{-4}$		
		(4 ~ 8) A			
		50 Hz	1.3 mA		
		50 Hz ~ 1 kHz	$1.4 \times 10^{-4}$		
		(1 ~ 10) kHz	$1.8 \times 10^{-4}$		
		(8 ~ 10) A			
		50 Hz	1.5 mA		
		50 Hz ~ 1 kHz	$1.3 \times 10^{-4}$		
		(1 ~ 10) kHz	$1.6 \times 10^{-4}$		

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC Current	40303	(10 ~ 30) A 50 Hz 50 Hz ~ 1 kHz	5.7 mA $1.7 \times 10^{-4}$	
		(30 ~ 50) A 50 Hz 50 Hz ~ 1 kHz	7.8 mA $1.4 \times 10^{-4}$	
		(50 ~ 80) A 50 Hz 50 Hz ~ 1 kHz	14 mA $1.6 \times 10^{-4}$	
		(80 ~ 100) A 50 Hz 50 Hz ~ 1 kHz	16 mA $1.5 \times 10^{-4}$	
Clamp meter		(50 ~ 60) Hz 1 A (1 ~ 3) A (3 ~ 8) A (8 ~ 10) A (10 ~ 20) A (20 ~ 30) A (30 ~ 50) A (50 ~ 80) A (80 ~ 100) A (100 ~ 200) A (200 ~ 300) A (300 ~ 500) A (500 ~ 800) A (800 ~ 1 000) A	0.21 A 0.25 A 0.33 A 0.37 A 0.53 A 0.71 A 1.6 A 2.0 A 2.4 A 4.1 A 5.8 A 16 A 21 A 24 A	
Wattmeter calibrators	40304	(50 ~ 60) Hz 0.12 W (0.12 ~ 0.48) W (0.48 ~ 0.6) W (0.6 ~ 2.4) W (2.4 ~ 24) W (24 ~ 240) W (240 ~ 600) W (600 ~ 1 200) W (1 200 ~ 2 400) W (2 400 ~ 4 800) W	0.35 mW $1.6 \times 10^{-3}$ $7.5 \times 10^{-4}$ $7.3 \times 10^{-4}$ $1.1 \times 10^{-3}$ $9.2 \times 10^{-4}$ $7.7 \times 10^{-4}$ $7.6 \times 10^{-4}$ $8.0 \times 10^{-4}$ $7.8 \times 10^{-4}$	Power calibrators / HCT-CS-275-40304
Power factor		(50 ~ 60) Hz -1 ~ 1	0.000 68	
Harmonic TVD-V TVD-I		(50 ~ 60) Hz (0.5 ~ 20) % (0.5 ~ 20) %	0.046 % 0.046 %	
Frequency		20 Hz (20 ~ 60) Hz (60 ~ 100) Hz (100 ~ 400) Hz 400 Hz ~ 1 kHz	$3.5 \times 10^{-4}$ $1.2 \times 10^{-4}$ $1.4 \times 10^{-4}$ $1.2 \times 10^{-4}$ $1.4 \times 10^{-4}$	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC current shunts AC Resistance	40305	40 Hz 0.001 Ω (0.001 ~ 0.01) Ω (0.01 ~ 0.1) Ω (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (100 ~ 1 000) Ω  (40 ~ 100) Hz 0.001 Ω (0.001 ~ 0.01) Ω (0.01 ~ 0.1) Ω (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (100 ~ 1 000) Ω  100 Hz ~ 1 kHz 0.001 Ω (0.001 ~ 0.01) Ω (0.01 ~ 0.1) Ω (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (100 ~ 1 000) Ω	0.44 μ Ω 13 μ Ω 40 μ Ω 0.27 mΩ 3.0 mΩ 30 mΩ 0.36 Ω  0.63 μ Ω 20 μ Ω 38 μ Ω 0.22 mΩ 2.4 mΩ 24 mΩ 0.28 Ω  2.0 μ Ω 20 μ Ω 37 μ Ω 0.20 mΩ 2.3 mΩ 23 mΩ 0.27 Ω	Current Sources / HCT-CS-073-40305
역률계, 무효율계 등	40310	(50 ~ 60) Hz -1 ~ 1	0.000 24	Power calibrators / HCT-CS-074-40310
AC power meters AC Power	40311	(50 ~ 60) Hz 0.06 W (0.06 ~ 0.12) W (0.12 ~ 0.48) W (0.48 ~ 0.6) W (0.6 ~ 1.2) W (1.2 ~ 2.4) W (2.4 ~ 6) W (6 ~ 12) W (12 ~ 24) W (24 ~ 48) W (48 ~ 60) W (60 ~ 120) W (120 ~ 240) W (240 ~ 480) W (480 ~ 600) W (600 ~ 1 200) W (1.2 ~ 2.4) kW (2.4 ~ 4.8) kW (4.8 ~ 9.6) kW (9.6 ~ 19.2) kW	0.12 mW 0.12 mW 0.13 mW 0.15 mW 0.17 mW 0.31 mW 1.1 mW 1.7 mW 3.1 mW 6.0 mW 7.4 mW 17 mW 31 mW 60 mW 75 mW 0.17 W 0.55 W 1.1 W 2.2 W 4.5 W	Power calibrators, Multimeter calibrators / HCT-CS-075-40311

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC Power	40311	0.10 W	0.16 mW	
		(0.1 ~ 1.2) W	0.19 mW	
		(1.2 ~ 2.4) W	0.36 mW	
		(2.4 ~ 3) W	0.55 mW	
		(3 ~ 4.8) W	0.71 mW	
		(4.8 ~ 6) W	1.7 mW	
		(6 ~ 12) W	2.9 mW	
		(12 ~ 24) W	4.4 mW	
		(24 ~ 48) W	9.8 mW	
		(48 ~ 60) W	35 mW	
		(60 ~ 120) W	40 mW	
		(120 ~ 240) W	79 mW	
		(240 ~ 480) W	0.14 mW	
		(480 ~ 500) W	0.24 mW	
		500 W ~ 1 kW	0.42 W	
		(1 ~ 1.2) kW	0.84 W	
		(1.2 ~ 3) kW	2.9 W	
(3 ~ 6) kW	5.8 W			
(6 ~ 12) kW	6.4 W			
(12 ~ 50) kW	24 W			
Power factor		(50 ~ 60) Hz	0.000 16	
		-1 ~ 1		
AC Voltage		(50 ~ 60) Hz		
		1 V		85 μV
		(1 ~ 2) V		0.13 mV
		(2 ~ 5) V		0.37 mV
		(5 ~ 10) V		0.85 mV
		(10 ~ 20) V		1.3 mV
		(20 ~ 50) V		4.5 mV
		(50 ~ 60) V		5.0 mV
		(60 ~ 100) V		9.4 mV
		(100 ~ 150) V		12 mV
		(150 ~ 200) V		15 mV
		(200 ~ 300) V		46 mV
		(300 ~ 500) V		64 mV
		(500 ~ 600) V		73 mV
		(600 ~ 750) V		95 mV
(750 ~ 1 000) V	0.14 V			
AC Current		(50 ~ 60) Hz		
		1 mA		0.19 μA
		(1 ~ 10) mA		1.9 μA
		(10 ~ 20) mA		3.0 μA
		(20 ~ 50) mA		11 μA
		(50 ~ 100) mA		18 μA
		(100 ~ 200) mA		29 μA
		(200 ~ 500) mA		0.21 mA
		(0.5 ~ 1) A		0.35 mA
		(1 ~ 2) A		0.62 mA
		(2 ~ 5) A		2.9 mA
		(5 ~ 10) A		5.6 mA
		(10 ~ 20) A		15 mA
		(20 ~ 30) A		29 mA
(30 ~ 50) A	31 mA			
DC Voltage		1 V	62 μV	
		(1 ~ 2) V	63 μV	
		(2 ~ 5) V	67 μV	
		(5 ~ 20) V	0.64 mV	
		(20 ~ 50) V	0.70 mV	



403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC Voltage	40311	(50 ~ 60) V	0.73 mV	
		(60 ~ 200) V	6.5 mV	
		(200 ~ 300) V	6.8 mV	
		(300 ~ 500) V	7.5 mV	
		(500 ~ 600) V	8.0 mV	
		(600 ~ 750) V	8.7 mV	
		(750 ~ 1 000) V	62 mV	
		DC Current	40311	
(1 ~ 10) mA	0.78 $\mu$ A			
(10 ~ 20) mA	1.1 $\mu$ A			
(20 ~ 50) mA	3.6 $\mu$ A			
(50 ~ 100) mA	8.6 $\mu$ A			
(100 ~ 200) mA	13 $\mu$ A			
(200 ~ 500) mA	62 $\mu$ A			
500 mA ~ 1 A	0.13 mA			
(1 ~ 2) A	0.21 mA			
(2 ~ 5) A	2.2 mA			
(5 ~ 10) A	2.6 mA			
(10 ~ 20) A	4.3 mA			
(20 ~ 30) A	8.7 mA			
(30 ~ 50) A	11 mA			
Harmonic Voltage	40311	(50 ~ 60) Hz	0.036 %	
		(0.5 % ~ 20 %)		
Harmonic Current	40311	(50 ~ 60) Hz	0.034 %	
		(0.5 % ~ 20 %)		
Frequency	40311	20 Hz	2.8 mHz	
		(20 Hz ~ 1 kHz)	$1.3 \times 10^{-4}$	
Flicker $P_{st}$	40311	(1 ~ 4 000) cpm	0.39 %	
		1		
$P_{inst,max}$ Sinusoidal	40311	(0.5 ~ 33.333) Hz	0.38 %	
		1		
Square	40311	(0.5 ~ 28) Hz	0.40 %	
		1		
		(28 ~ 30.5) Hz	1.1 %	
		1		
(30.5 ~ 33.333) Hz	0.40 %			
$P_{st}$ Range	40311	1 620 cpm	0.39 %	
		0.25		
		(0.25 ~ 5)		

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power supplies AC Voltage	40312	100 mV 20 Hz (20 ~ 50) Hz 50 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz  (0.1 ~ 0.4) V 20 Hz 20 Hz ~ 10 kHz (10 ~ 100) kHz  (0.4 ~ 0.8) V 20 Hz (20 ~ 50) Hz 50 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz  (0.8 ~ 1) V 20 Hz (20 ~ 50) Hz 50 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz  (1 ~ 4) V 20 Hz 20 Hz ~ 10 kHz (10 ~ 100) kHz  (4 ~ 8) V 20 Hz (20 ~ 50) Hz 50 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz  (8 ~ 10) V 20 Hz (20 ~ 50) Hz 50 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz  (10 ~ 50) V 20 Hz (20 ~ 50) Hz 50 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz  (50 ~ 80) V 20 Hz (20 ~ 50) Hz 50 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz	21 μV $1.9 \times 10^{-4}$ $1.7 \times 10^{-4}$ $1.9 \times 10^{-4}$ $1.1 \times 10^{-3}$  0.11 mV $2.5 \times 10^{-4}$ $1.2 \times 10^{-3}$  0.14 mV $1.7 \times 10^{-4}$ $1.4 \times 10^{-4}$ $1.7 \times 10^{-4}$ $8.8 \times 10^{-4}$  0.16 mV $1.4 \times 10^{-4}$ $1.2 \times 10^{-4}$ $1.4 \times 10^{-4}$ $8.2 \times 10^{-4}$  1.1 mV $2.5 \times 10^{-4}$ $1.2 \times 10^{-3}$  1.4 mV $1.7 \times 10^{-4}$ $1.5 \times 10^{-4}$ $1.7 \times 10^{-4}$ $8.8 \times 10^{-4}$  1.6 mV $1.4 \times 10^{-4}$ $1.3 \times 10^{-4}$ $1.4 \times 10^{-4}$ $8.2 \times 10^{-4}$  15 mV $2.2 \times 10^{-4}$ $2.0 \times 10^{-4}$ $2.2 \times 10^{-4}$ $1.1 \times 10^{-3}$  17 mV $1.7 \times 10^{-4}$ $1.5 \times 10^{-4}$ $1.7 \times 10^{-4}$ $8.9 \times 10^{-4}$	Multimeters / HCT-CS-076-40312

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC Voltage	40312	(80 ~ 100) V		
		20 Hz	19 mV	
		(20 ~ 50) Hz	$1.5 \times 10^{-4}$	
		50 Hz ~ 1 kHz	$1.3 \times 10^{-4}$	
		(1 ~ 10) kHz	$1.5 \times 10^{-4}$	
		(10 ~ 100) kHz	$8.2 \times 10^{-4}$	
		(100 ~ 150) V		
		50 Hz	0.12 V	
		(50 ~ 100) Hz	$8.0 \times 10^{-4}$	
		100 Hz ~ 1 kHz	$5.4 \times 10^{-4}$	
		(1 ~ 10) kHz	$8.0 \times 10^{-4}$	
		(150 ~ 200) V		
		50 Hz	0.14 V	
		(50 ~ 100) Hz	$7.0 \times 10^{-4}$	
		100 Hz ~ 1 kHz	$5.0 \times 10^{-4}$	
		(1 ~ 10) kHz	$7.0 \times 10^{-4}$	
		(200 ~ 300) V		
		50 Hz	0.14 V	
		(50 ~ 100) Hz	$4.7 \times 10^{-4}$	
		100 Hz ~ 1 kHz	$3.7 \times 10^{-4}$	
(1 ~ 10) kHz	$4.7 \times 10^{-4}$			
(300 ~ 500) V				
50 Hz	0.63 V			
50 Hz ~ 10 kHz	$1.3 \times 10^{-3}$			
(500 ~ 800) V				
50 Hz	0.63 V			
50 Hz ~ 10 kHz	$7.9 \times 10^{-4}$			
(800 ~ 1 000) V				
50 Hz	0.63 V			
50 Hz ~ 10 kHz	$6.4 \times 10^{-4}$			
(1 ~ 1.5) kV				
60 Hz	0.012 kV			
DC Voltage	40312	100 mV	6.2 $\mu$ V	
		(0.1 ~ 1) V	59 $\mu$ V	
		(1 ~ 10) V	0.58 mV	
		(10 ~ 100) V	2.2 mV	
		(100 ~ 400) V	0.062 V	
		(400 ~ 1 000) V	0.62 V	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC Current	40312	100 $\mu$ A		
		50 Hz ~ 10 kHz	71 nA	
		(0.1 ~ 0.4) mA		
		50 Hz ~ 10 kHz	0.12 $\mu$ A	
		(0.4 ~ 0.8) mA		
		50 Hz ~ 10 kHz	0.15 $\mu$ A	
		(0.8 ~ 1) mA		
		50 Hz ~ 10 kHz	0.18 $\mu$ A	
		(1 ~ 4) mA		
		50 Hz ~ 10 kHz	1.1 $\mu$ A	
		(4 ~ 8) mA		
		50 Hz ~ 10 kHz	1.4 $\mu$ A	
		(8 ~ 10) mA		
		50 Hz ~ 10 kHz	1.6 $\mu$ A	
		(10 ~ 40) mA		
		50 Hz ~ 10 kHz	11 $\mu$ A	
		(40 ~ 80) mA		
		50 Hz ~ 10 kHz	14 $\mu$ A	
		(80 ~ 100) mA		
		50 Hz ~ 10 kHz	15 $\mu$ A	
(0.1 ~ 0.4) A				
50 Hz ~ 10 kHz	0.11 mA			
(0.4 ~ 0.8) A				
50 Hz ~ 10 kHz	0.14 mA			
(0.8 ~ 1) A				
50 Hz ~ 10 kHz	0.16 mA			
(1 ~ 4) A				
50 Hz ~ 10 kHz	1.1 mA			
(4 ~ 8) A				
50 Hz ~ 10 kHz	1.5 mA			
(8 ~ 10) A				
50 Hz ~ 10 kHz	1.7 mA			
(10 ~ 20) A				
50 Hz ~ 10 kHz	7.5 mA			
(20 ~ 30) A				
50 Hz ~ 10 kHz	8.1 mA			
(30 ~ 45) A				
50 Hz ~ 10 kHz	9.3 mA			
DC Current		100 $\mu$ A	10 nA	
		(0.1 ~ 1) mA	64 nA	
		(1 ~ 10) mA	0.64 $\mu$ A	
		(10 ~ 100) mA	6.4 $\mu$ A	
		(0.1 ~ 1) A	64 $\mu$ A	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC Current	40312	(1 ~ 10) A (10 ~ 40) A (40 ~ 80) A (80 ~ 100) A	0.68 mA 25 mA 48 mA 53 mA	
Frequency	40312	20 Hz (20 ~ 100) Hz 100 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 40) kHz (40 ~ 80) kHz (80 ~ 100) kHz	9 μHz $7.8 \times 10^{-7}$ $8.4 \times 10^{-7}$ $9.8 \times 10^{-7}$ $1.7 \times 10^{-6}$ $8.9 \times 10^{-7}$ $7.5 \times 10^{-7}$	
Puncture/safety testers	40313			High voltage voltmeters Digital multimeter / HCT-CS-077-40313
AC Voltage		(50 ~ 60) Hz 0.1 kV (0.1 ~ 1) kV (1 ~ 2) kV (2 ~ 4) kV (4 ~ 6) kV (6 ~ 8) kV (8 ~ 10) kV (10 ~ 20) kV (20 ~ 30) kV (30 ~ 40) kV (40 ~ 50) kV (50 ~ 60) kV (60 ~ 70) kV (70 ~ 75) kV	0.62 V 0.64 V 5.1 V 9.1 V 14 V 18 V 24 V 47 V 62 V 85 V 0.10 kV 0.13 kV 0.14 kV 0.15 kV	
DC Voltage		(±) 0.1 kV (0.1 ~ 1) kV (1 ~ 2) kV (2 ~ 4) kV (4 ~ 6) kV (6 ~ 8) kV (8 ~ 10) kV (10 ~ 20) kV (20 ~ 30) kV (30 ~ 40) kV (40 ~ 50) kV (50 ~ 60) kV (60 ~ 70) kV (70 ~ 80) kV (80 ~ 90) kV (90 ~ 100) kV	0.64 V 0.64 V 1.4 V 2.6 V 3.8 V 5.0 V 11 V 17 V 28 V 32 V 43 V 47 V 58 V 62 V 73 V 77 V	
AC Cut-off Current		(50 ~ 60) Hz 0.1 mA (0.1 ~ 0.5) mA (0.5 ~ 1) mA (1 ~ 2) mA (2 ~ 5) mA (5 ~ 10) mA (10 ~ 50) mA (50 ~ 100) mA	0.11 μA 0.42 μA 0.86 μA 3.2 μA 4.2 μA 8.6 μA 40 μA 83 μA	
DC Cut-off Current		0.1 mA (0.1 ~ 0.5) mA (0.5 ~ 1) mA (1 ~ 5) mA	68 nA 83 nA 0.64 μA 0.83 μA	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC Cut-off Current	40313	(5 ~ 10) mA	6.4 $\mu$ A	
		(10 ~ 50) mA	8.3 $\mu$ A	
		(50 ~ 100) mA	64 $\mu$ A	
Insulation Voltage		25 V	0.90 mV	
		(25 ~ 50) V	0.94 mV	
		(50 ~ 100) V	1.1 mV	
		(100 ~ 500) V	9.0 mV	
		(500 ~ 800) V	10 mV	
		(0.8 ~ 4) kV	2.6 V	
		(4 ~ 6) kV	3.8 V	
		(6 ~ 8) kV	5.0 V	
		(8 ~ 10) kV	11 V	
Insulation Resistance		(1 ~ 7) k $\Omega$	0.65 $\Omega$	
		(7 ~ 10) k $\Omega$	0.99 $\Omega$	
		(10 ~ 100) k $\Omega$	6.6 $\Omega$	
	(100 ~ 200) k $\Omega$	95 $\Omega$		
	(200 ~ 500) k $\Omega$	0.19 k $\Omega$		
	(500 ~ 700) k $\Omega$	0.26 k $\Omega$		
	(700 ~ 1 000) k $\Omega$	0.36 k $\Omega$		
	(1 ~ 100) M $\Omega$	$1.3 \times 10^{-3}$		
	(100 ~ 1 000) M $\Omega$	$2.6 \times 10^{-3}$		
	(1 ~ 10) G $\Omega$	$6.5 \times 10^{-3}$		
	(10 ~ 100) G $\Omega$	$1.2 \times 10^{-2}$		
Ground Bond AC Current	(50 ~ 60) Hz	1.0 mA 1.6 mA 7.5 mA 8.2 mA 9.0 mA 9.9 mA 14 mA		
	1 A			
	(1 ~ 10) A			
	(10 ~ 20) A			
	(20 ~ 30) A			
	(30 ~ 40) A			
	(40 ~ 50) A			
	(50 ~ 60) A			
Ground Bond Resistance	(50 ~ 60) Hz	1.3 m $\Omega$ 2.4 m $\Omega$ $1.2 \times 10^{-2}$		
	100 m $\Omega$			
	(100 ~ 200) m $\Omega$			
	(200 ~ 500) m $\Omega$			
Time	(1 ~ 5) s	2 ms		
	(5 s ~ 30) s	0.04 s		
	(30 s ~ 60) s	0.07 s		
Power recorders AC Power	40314	(50 ~ 60) Hz	2.5 mW 2.5 mW 2.9 mW 3.6 mW 4.6 mW 9.1 mW 27 mW 46 mW 91 mW 0.17 W 0.24 W 0.46 W 0.91 W 1.7 W 2.4 W	Power calibrators Multimeter calibrators / HCT-CS-078-40314
	1.5 W			
	(1.5 ~ 3) W			
	(3 ~ 12) W			
	(12 ~ 15) W			
	(15 ~ 30) W			
	(30 ~ 60) W			
	(60 ~ 120) W			
	(120 ~ 300) W			
	(300 ~ 600) W			
	(600 ~ 1 200) W			
	(1.2 ~ 1.5) kW			
	(1.5 ~ 3) kW			
	(3 ~ 6) kW			
	(6 ~ 12) kW			
	(12 ~ 15) kW			

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC Power	40314	(15 ~ 30) kW	4.6 W	
		(30 ~ 60) kW	9.1 W	
		(60 ~ 120) kW	20 W	
		(120 ~ 240) kW	31 W	
DC Power	40314	0.5 W	3.7 mW	
		(0.5 ~ 2.5) W	4.4 mW	
		(2.5 ~ 7.5) W	5.3 mW	
		(7.5 ~ 15) W	5.9 mW	
		(15 ~ 24) W	6.6 mW	
		(24 ~ 30) W	7.2 mW	
		(30 ~ 50) W	9.5 mW	
		(50 ~ 60) W	12 mW	
		(60 ~ 75) W	15 mW	
		(75 ~ 125) W	42 mW	
		(125 ~ 300) W	71 mW	
		(300 ~ 600) W	0.12 W	
		(600 ~ 750) W	0.25 W	
		(0.75 ~ 1.25) kW	0.87 W	
		(1.25 ~ 3) kW	1.6 W	
		(3 ~ 12.5) kW	6.0 W	
		(12.5 ~ 15) kW	11 W	
		(15 ~ 30) kW	36 W	
		(30 ~ 60) kW	72 W	
		(60 ~ 120) kW	0.15 kW	
		(120 ~ 250) kW	0.29 kW	
		(250 ~ 500) kW	0.60 kW	
		(12.5 ~ 15) kW	2.3 W	
		(15 ~ 25) kW	2.9 W	
(25 ~ 30) kW	4.6 W			
(30 ~ 60) kW	7.8 W			
(60 ~ 125) kW	17 W			
(125 ~ 250) kW	32 W			
(250 ~ 500) kW	60 W			
Power Factor	40314	(50 ~ 60) Hz	0.000 16	
		-1 ~ 1		
AC Voltage	40314	(50 ~ 60) Hz	85 μV	
		1 V		
		(1 ~ 2) V		
		(2 ~ 5) V		
		(5 ~ 10) V		
		(10 ~ 20) V		
		(20 ~ 50) V		
		(50 ~ 60) V		
		(60 ~ 100) V		
		(100 ~ 150) V		
		(150 ~ 200) V		
		(200 ~ 300) V		
		(300 ~ 500) V		
		(500 ~ 600) V		
		(600 ~ 750) V		
		(750 ~ 1 000) V		
AC Current	40314	(50 ~ 60) Hz	0.24 mA	
		100 mA		
		100 mA ~ 1 A		
		(1 ~ 10) A		
		(10 ~ 100) A		
		(100 ~ 200) A		
		(200 ~ 300) A		
		(200 ~ 300) A		

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC Current	40314	(300 ~ 400) A (400 ~ 500) A (500 ~ 750) A (750 ~ 900) A (900 ~ 1 000) A (1 000 ~ 1 500) A (1 500 ~ 2 000) A (2 000 ~ 2 500) A (2 500 ~ 3 000) A	0.99 A 1.3 A 2.0 A 2.3 A 2.6 A 3.7 A 4.8 A 5.9 A 7.1 A	
AC voltmeters	40318			Multimeter calibrators, Digitor Multimeters AC voltage standard /HCT-CS-079-40318
DC Voltage		2 mV (2 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V -2 mV (-2 ~ -10) mV (-10 ~ -100) mV (-0.1 ~ -1) V (-1 ~ -10) V (-10 ~ -100) V (-100 ~ -1 000) V	53 μV $5.9 \times 10^{-5}$ $1.4 \times 10^{-5}$ $6.7 \times 10^{-6}$ $4.6 \times 10^{-6}$ $6.3 \times 10^{-6}$ $8.1 \times 10^{-6}$ 53 μV $5.9 \times 10^{-5}$ $1.4 \times 10^{-5}$ $6.7 \times 10^{-6}$ $4.6 \times 10^{-6}$ $6.3 \times 10^{-6}$ $8.1 \times 10^{-6}$	
AC Voltage		2 mV 40 Hz 40 Hz ~ 50 kHz (50 ~ 100) kHz (100 ~ 500) kHz 500 kHz ~ 1 MHz  (2 ~ 10) mV 40 Hz 40 Hz ~ 20 kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 500) kHz 500 kHz ~ 1 MHz  (10 ~ 100) mV 40 Hz 40 Hz ~ 20 kHz (20 ~ 50) kHz (50 ~ 100) kHz 100 kHz ~ 1 MHz  100 mV ~ 1 V 10 Hz (10 ~ 20) Hz (20 ~ 40) Hz 40 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 20) kHz (20 ~ 50) kHz	3.9 μV $2.0 \times 10^{-3}$ $2.3 \times 10^{-3}$ $5.5 \times 10^{-3}$ $6.0 \times 10^{-3}$  3.5 μV $3.5 \times 10^{-4}$ $3.7 \times 10^{-4}$ $6.1 \times 10^{-4}$ $1.5 \times 10^{-3}$ $1.8 \times 10^{-3}$  6.0 μV $6.0 \times 10^{-5}$ $1.2 \times 10^{-4}$ $1.8 \times 10^{-4}$ $7.0 \times 10^{-4}$  0.23 mV $7.2 \times 10^{-5}$ $3.8 \times 10^{-5}$ $2.0 \times 10^{-5}$ $2.6 \times 10^{-5}$ $2.0 \times 10^{-5}$ $4.8 \times 10^{-5}$	



403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC Voltage	40318	(50 ~ 100) kHz	$6.0 \times 10^{-5}$	
		(100 ~ 500) kHz	$5.0 \times 10^{-4}$	
		500 kHz ~ 1 MHz	$5.3 \times 10^{-4}$	
		(1 ~ 10) V		
		10 Hz	2.3 mV	
		(10~ 20) Hz	$7.2 \times 10^{-5}$	
		(20 ~ 40) Hz	$3.7 \times 10^{-5}$	
		40 Hz ~ 20 kHz	$2.5 \times 10^{-5}$	
		(20 ~ 50) kHz	$4.7 \times 10^{-5}$	
		(50 ~ 100) kHz	$5.9 \times 10^{-5}$	
		(100 ~ 500) kHz	$5.0 \times 10^{-4}$	
		500 kHz ~ 1 MHz	$5.6 \times 10^{-4}$	
		(10 ~ 100) V		
		40 Hz	3.7 mV	
		40 Hz ~ 20 kHz	$3.7 \times 10^{-5}$	
(20 ~ 50) kHz	$7.6 \times 10^{-5}$			
(50 ~ 100) kHz	$8.3 \times 10^{-5}$			
(100 ~ 1 000) V				
40 Hz	35 mV			
40 Hz ~ 20 kHz	$3.5 \times 10^{-5}$			
Frequency		10 Hz	5.8 mHz	
		10 Hz ~ 10 MHz	$5.8 \times 10^{-5}$	
		(10 ~ 50) MHz	$1.2 \times 10^{-4}$	
Frequency Response		0 dB (0.774 6 V)		
		20 Hz ~ 100 kHz	0.002 dB	
		(100 ~ 200) kHz	0.005 dB	
Output Voltage		1 V		
		100 Hz	1.1 mV	
		100 Hz ~ 20 kHz	$1.0 \times 10^{-3}$	
		(20 ~ 50) kHz	$2.0 \times 10^{-3}$	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF amplifiers	40401	10 Hz		Multimeter calibrators Digital multimeters /HCT-CS-181-40401
Gain		100 mV	33 μV	
		(0.1 ~ 1) V	$3.3 \times 10^{-4}$	
		(1 ~ 10) V	$3.6 \times 10^{-4}$	
		(10 ~ 100) V	$1.9 \times 10^{-4}$	
		(10 ~ 100) Hz		
		100 mV	25 μV	
		(0.1 ~ 1) V	$2.5 \times 10^{-4}$	
		(1 ~ 10) V	$3.1 \times 10^{-4}$	
		(10 ~ 100) V	$1.3 \times 10^{-4}$	
		(0.1 ~ 1) kHz		
		100 mV	25 μV	
		(0.1 ~ 1) V	$2.5 \times 10^{-4}$	
		(1 ~ 10) V	$1.7 \times 10^{-4}$	
		(10 ~ 100) V	$1.3 \times 10^{-4}$	
		(1 ~ 10) kHz		
		100 mV	26 μV	
		(0.1 ~ 1) V	$2.6 \times 10^{-4}$	
		(1 ~ 10) V	$3.2 \times 10^{-4}$	
		(10 ~ 100) V	$1.4 \times 10^{-4}$	
	(10 ~ 100) kHz			
	100 mV	0.12 mV		
	(0.1 ~ 1) V	$1.2 \times 10^{-3}$		
	(1 ~ 10) V	$1.8 \times 10^{-3}$		
	(10 ~ 100) V	$8.2 \times 10^{-4}$		
	10 Hz ~ 1 kHz			
	(0 ~ 60) dB	0.006 dB		
	(1 ~ 20) kHz			
	(0 ~ 60) dB	0.007 dB		
	(20 ~ 100) kHz			
	(0 ~ 40) dB	0.010 dB		
Charge type amplifier		10 Hz		
Gain		10 mV	60 μV	
		(10 ~ 100) mV	$6.0 \times 10^{-3}$	
		(0.1 ~ 1) V	$6.8 \times 10^{-4}$	
		(1 ~ 9) V	$4.5 \times 10^{-4}$	
		(10 ~ 100) Hz		
		10 mV	60 μV	
		(10 ~ 100) mV	$6.0 \times 10^{-3}$	
		(0.1 ~ 1) V	$6.1 \times 10^{-4}$	
		(1 ~ 9) V	$4.5 \times 10^{-4}$	
		(0.1 ~ 1) kHz		
		10 mV	60 μV	
		(10 ~ 100) mV	$6.0 \times 10^{-3}$	
		(0.1 ~ 1) V	$6.1 \times 10^{-4}$	
		(1 ~ 9) V	$4.5 \times 10^{-4}$	
		(1 ~ 10) kHz		
		10 mV	60 μV	
		(10 ~ 100) mV	$6.0 \times 10^{-3}$	
		(0.1 ~ 1) V	$6.1 \times 10^{-4}$	
		(1 ~ 9) V	$4.5 \times 10^{-4}$	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Charge type amplifier Gain	40401	(10 ~ 20) kHz 10 mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 9) V	60 μV $6.1 \times 10^{-3}$ $7.6 \times 10^{-4}$ $1.1 \times 10^{-3}$	
DC/LF attenuators Attenuator	40402	20 Hz ~ 20 kHz (0 ~ 50) dB (50 ~ 60) dB  (20 ~ 50) kHz (0 ~ 50) dB (50 ~ 60) dB  (50 ~ 100) kHz (0 ~ 50) dB (50 ~ 60) dB	0.017 dB 0.044 dB  0.044 dB 0.056 dB  0.044 dB 0.056 dB	Function Generator, Digital Multimeters /HCT-CS-081-40402
Multimeter calibrators DC Voltage          AC Voltage	40403	0 mV (0 ~ 100) mV (-0 ~ -100) mV (0.1 ~ 1) V (-0.1 ~ -1) V (1 ~ 10) V (-1 ~ -10) V (10 ~ 100) V (-10 ~ -100) V (100 ~ 1 000) V (-100 ~ -1 000) V  (10 Hz) 100 μV (0.1 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V  (10 ~ 40) Hz 100 μV (0.1 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V  (40 ~ 100) Hz 100 μV (0.1 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V  (100 ~ 500) Hz 100 μV (0.1 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V	0.13 μV $2.3 \times 10^{-6}$ $2.3 \times 10^{-6}$ $2.3 \times 10^{-6}$ $2.3 \times 10^{-6}$ $2.3 \times 10^{-6}$ $2.3 \times 10^{-6}$ $2.3 \times 10^{-6}$ $2.3 \times 10^{-6}$ $2.3 \times 10^{-6}$  0.27 μV $4.0 \times 10^{-5}$ $3.3 \times 10^{-5}$ $4.7 \times 10^{-5}$ $5.2 \times 10^{-5}$  0.13 μV $3.3 \times 10^{-5}$ $2.7 \times 10^{-5}$ $1.7 \times 10^{-5}$ $3.2 \times 10^{-5}$ $1.5 \times 10^{-5}$  0.13 μV $2.0 \times 10^{-5}$ $2.6 \times 10^{-5}$ $2.0 \times 10^{-5}$ $3.5 \times 10^{-5}$ $2.0 \times 10^{-5}$  0.13 μV $3.1 \times 10^{-5}$ $1.7 \times 10^{-5}$ $2.8 \times 10^{-5}$ $2.3 \times 10^{-5}$ $2.2 \times 10^{-5}$	Standard cell, Standard resistance, Standard divider, Digital multimeters, AC calibrator /HCT-CS-082-40403

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC Voltage	40403	500 Hz ~ 1 kHz		
		100 μV	0.13 μV	
		(0.1 ~ 100) mV	$3.1 \times 10^{-5}$	
		(0.1 ~ 1) V	$2.1 \times 10^{-5}$	
		(1 ~ 10) V	$2.0 \times 10^{-5}$	
		(10 ~ 100) V	$2.6 \times 10^{-5}$	
		(100 ~ 1 000) V	$2.7 \times 10^{-5}$	
		(1 ~ 10) kHz		
		100 μV	0.13 μV	
		(0.1 ~ 100) mV	$2.7 \times 10^{-5}$	
		(0.1 ~ 1) V	$1.7 \times 10^{-5}$	
		(1 ~ 10) V	$1.5 \times 10^{-5}$	
		(10 ~ 100) V	$2.8 \times 10^{-5}$	
		(100 ~ 1 000) V	$3.1 \times 10^{-5}$	
		(10 ~ 20) kHz		
		100 μV	0.13 μV	
		(0.1 ~ 100) mV	$2.9 \times 10^{-5}$	
		(0.1 ~ 1) V	$2.6 \times 10^{-5}$	
		(1 ~ 10) V	$1.5 \times 10^{-5}$	
		(10 ~ 100) V	$4.9 \times 10^{-5}$	
		(100 ~ 1 000) V	$2.7 \times 10^{-5}$	
		(20 ~ 30) kHz		
		100 μV	0.21 μV	
		(0.1 ~ 100) mV	$3.0 \times 10^{-5}$	
		(0.1 ~ 1) V	$2.1 \times 10^{-5}$	
		(1 ~ 10) V	$2.5 \times 10^{-5}$	
		(10 ~ 100) V	$4.8 \times 10^{-5}$	
		(100 ~ 1 000) V	$4.4 \times 10^{-5}$	
		(30 ~ 50) kHz		
		100 μV	3.5 μV	
		(0.1 ~ 100) mV	$4.3 \times 10^{-5}$	
		(0.1 ~ 1) V	$2.7 \times 10^{-5}$	
		(1 ~ 10) V	$3.0 \times 10^{-5}$	
		(10 ~ 100) V	$4.1 \times 10^{-5}$	
		(100 ~ 600) V	$1.1 \times 10^{-4}$	
		(50 ~ 100) kHz		
		100 μV	0.28 μV	
		(0.1 ~ 100) mV	$6.7 \times 10^{-5}$	
		(0.1 ~ 1) V	$3.6 \times 10^{-5}$	
		(1 ~ 10) V	$5.8 \times 10^{-5}$	
		(10 ~ 100) V	$6.5 \times 10^{-5}$	
		(100 ~ 600) V	$1.2 \times 10^{-4}$	
		(100 ~ 200) kHz		
		100 μV	0.50 μV	
		(0.1 ~ 100) mV	$1.1 \times 10^{-4}$	
		(0.1 ~ 1) V	$5.9 \times 10^{-5}$	
		(1 ~ 10) V	$5.9 \times 10^{-5}$	
		(10 ~ 60) V	$1.4 \times 10^{-4}$	
(200 ~ 300) kHz				
100 μV	0.50 μV			
(0.1 ~ 100) mV	$1.3 \times 10^{-4}$			
(0.1 ~ 1) V	$5.9 \times 10^{-5}$			
(1 ~ 10) V	$6.3 \times 10^{-5}$			
(10 ~ 60) V	$1.8 \times 10^{-4}$			

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC Voltage	40403	(300 ~ 500) kHz		
		100 μV	0.74 μV	
		(0.1 ~ 100) mV	$1.5 \times 10^{-4}$	
		(0.1 ~ 1) V	$1.3 \times 10^{-4}$	
		(1 ~ 20) V	$5.0 \times 10^{-5}$	
		(0.5 ~ 1) MHz		
		100 μV	0.88 μV	
		(0.1 ~ 100) mV	$4.8 \times 10^{-4}$	
		(0.1 ~ 1) V	$2.7 \times 10^{-4}$	
		(1 ~ 20) V	$2.5 \times 10^{-4}$	
		(1 ~ 2) MHz		
		100 μV	0.20 μV	
		(0.1 ~ 100) mV	$4.4 \times 10^{-4}$	
		(0.1 ~ 1) V	$5.5 \times 10^{-4}$	
		(1 ~ 3) V	$2.4 \times 10^{-4}$	
		(2 ~ 5) MHz		
		100 μV	0.31 μV	
		(0.1 ~ 100) mV	$9.1 \times 10^{-4}$	
		(0.1 ~ 1) V	$9.0 \times 10^{-4}$	
		(1 ~ 3) V	$7.3 \times 10^{-4}$	
		(5 ~ 10) MHz		
		100 μV	0.31 μV	
		(0.1 ~ 100) mV	$1.1 \times 10^{-3}$	
		(0.1 ~ 1) V	$8.4 \times 10^{-4}$	
(1 ~ 3) V	$9.1 \times 10^{-4}$			
(10 ~ 20) MHz				
100 μV	0.47 μV			
(0.1 ~ 100) mV	$1.1 \times 10^{-3}$			
(0.1 ~ 1) V	$7.8 \times 10^{-4}$			
(1 ~ 3) V	$8.3 \times 10^{-4}$			
(20 ~ 30) MHz				
100 μV	1.4 μV			
(0.1 ~ 100) mV	$2.2 \times 10^{-3}$			
(0.1 ~ 1) V	$1.3 \times 10^{-3}$			
(1 ~ 3) V	$1.3 \times 10^{-3}$			
DC Current		0 μA	0.81 nA	
		(0 ~ 100) μA	$1.0 \times 10^{-5}$	
		(-0 ~ -100) μA	$1.0 \times 10^{-5}$	
		(0.1 ~ 1) mA	$1.1 \times 10^{-5}$	
		(-0.1 ~ -1) mA	$1.1 \times 10^{-5}$	
		(1 ~ 10) mA	$1.2 \times 10^{-5}$	
		(-1 ~ -10) mA	$1.2 \times 10^{-5}$	
		(10 ~ 100) mA	$8.7 \times 10^{-6}$	
		(-10 ~ -100) mA	$8.7 \times 10^{-6}$	
		(0.1 ~ 1) A	$8.8 \times 10^{-6}$	
		(-0.1 ~ -1) A	$8.8 \times 10^{-6}$	
		(1 ~ 10) A	$1.6 \times 10^{-5}$	
		(-1 ~ -10) A	$1.6 \times 10^{-5}$	
		(10 ~ 20) A	$5.4 \times 10^{-5}$	
		(-10 ~ -20) A	$5.4 \times 10^{-5}$	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
AC Current	40403	(10 Hz)			
		1 $\mu$ A	0.81 nA		
		(1 ~ 100) $\mu$ A	$3.2 \times 10^{-4}$		
		(0.1 ~ 1) mA	$2.9 \times 10^{-4}$		
		(1 ~ 10) mA	$2.9 \times 10^{-4}$		
		(10 ~ 100) mA	$2.9 \times 10^{-4}$		
		(0.1 ~ 1) A	$2.9 \times 10^{-4}$		
		(1 ~ 3) A	$9.9 \times 10^{-4}$		
		(10 ~ 40) Hz			
		1 $\mu$ A	32 nA		
		(1 ~ 100) $\mu$ A	$3.2 \times 10^{-4}$		
		(0.1 ~ 1) mA	$2.9 \times 10^{-4}$		
		(1 ~ 10) mA	$2.9 \times 10^{-4}$		
		(10 ~ 100) mA	$2.9 \times 10^{-4}$		
		(0.1 ~ 1) A	$2.9 \times 10^{-4}$		
		(1 ~ 3) A	$9.9 \times 10^{-4}$		
		(40 ~ 45) Hz			
		1 $\mu$ A	6.6 nA		
		(1 ~ 100) $\mu$ A	$7.3 \times 10^{-5}$		
		(0.1 ~ 1) mA	$3.7 \times 10^{-5}$		
		(1 ~ 10) mA	$4.3 \times 10^{-5}$		
		(10 ~ 100) mA	$4.4 \times 10^{-5}$		
		(0.1 ~ 1) A	$4.5 \times 10^{-5}$		
		(1 ~ 10) A	$2.9 \times 10^{-4}$		
		(10 ~ 20) A	$2.9 \times 10^{-4}$		
		(45 ~ 100) Hz			
		1 $\mu$ A	6.6 nA		
		(1 ~ 100) $\mu$ A	$7.3 \times 10^{-5}$		
		(0.1 ~ 1) mA	$3.7 \times 10^{-5}$		
		(1 ~ 10) mA	$4.3 \times 10^{-5}$		
		(10 ~ 100) mA	$4.4 \times 10^{-5}$		
		(0.1 ~ 1) A	$4.7 \times 10^{-5}$		
		(1 ~ 10) A	$4.9 \times 10^{-5}$		
		(10 ~ 20) A	$4.7 \times 10^{-5}$		
		(100 ~ 200) Hz			
		1 $\mu$ A	6.6 nA		
		(1 ~ 100) $\mu$ A	$7.2 \times 10^{-5}$		
		(0.1 ~ 1) mA	$3.8 \times 10^{-5}$		
		(1 ~ 10) mA	$4.3 \times 10^{-5}$		
		(10 ~ 100) mA	$4.5 \times 10^{-5}$		
		(0.1 ~ 1) A	$4.7 \times 10^{-5}$		
		(1 ~ 10) A	$4.8 \times 10^{-5}$		
		(10 ~ 20) A	$4.7 \times 10^{-5}$		
		(200 ~ 500) Hz			
		1 $\mu$ A	6.6 nA		
		(1 ~ 100) $\mu$ A	$7.2 \times 10^{-5}$		
		(0.1 ~ 1) mA	$3.7 \times 10^{-5}$		
		(1 ~ 10) mA	$4.3 \times 10^{-5}$		
(10 ~ 100) mA	$4.4 \times 10^{-5}$				
(0.1 ~ 1) A	$4.5 \times 10^{-5}$				
(1 ~ 10) A	$6.3 \times 10^{-5}$				
(10 ~ 20) A	$4.8 \times 10^{-5}$				
500 Hz ~ 1 kHz					
1 $\mu$ A	6.6 nA				
(1 ~ 100) $\mu$ A	$7.2 \times 10^{-5}$				
(0.1 ~ 1) mA	$3.7 \times 10^{-5}$				

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC Current	40403	(1 ~ 10) mA	$4.3 \times 10^{-5}$	
		(10 ~ 100) mA	$4.5 \times 10^{-5}$	
		(0.1 ~ 1) A	$4.3 \times 10^{-5}$	
		(1 ~ 10) A	$4.9 \times 10^{-5}$	
		(10 ~ 20) A	$4.8 \times 10^{-5}$	
		(1 ~ 2) kHz		
		1 $\mu$ A	6.7 nA	
		(1 ~ 100) $\mu$ A	$7.3 \times 10^{-5}$	
		(0.1 ~ 1) mA	$3.7 \times 10^{-5}$	
		(1 ~ 10) mA	$4.3 \times 10^{-5}$	
		(10 ~ 100) mA	$4.3 \times 10^{-5}$	
		(0.1 ~ 1) A	$4.4 \times 10^{-5}$	
		(1 ~ 10) A	$4.9 \times 10^{-5}$	
		(10 ~ 20) A	$4.8 \times 10^{-5}$	
		(2 ~ 5) kHz		
		1 $\mu$ A	6.7 nA	
		(1 ~ 100) $\mu$ A	$7.3 \times 10^{-5}$	
		(0.1 ~ 1) mA	$3.5 \times 10^{-5}$	
		(1 ~ 10) mA	$4.2 \times 10^{-5}$	
		(10 ~ 100) mA	$4.6 \times 10^{-5}$	
		(0.1 ~ 1) A	$4.5 \times 10^{-5}$	
		(1 ~ 10) A	$4.9 \times 10^{-5}$	
		(10 ~ 20) A	$4.9 \times 10^{-5}$	
		(5 ~ 10) kHz		
		1 $\mu$ A	6.7 nA	
		(1 ~ 100) $\mu$ A	$7.3 \times 10^{-5}$	
		(0.1 ~ 1) mA	$3.7 \times 10^{-5}$	
		(1 ~ 10) mA	$4.3 \times 10^{-5}$	
		(10 ~ 100) mA	$4.5 \times 10^{-5}$	
		(0.1 ~ 1) A	$5.0 \times 10^{-5}$	
		(1 ~ 3) A	$2.6 \times 10^{-4}$	
		(10 ~ 30) kHz		
		1 $\mu$ A	12 nA	
		(1 ~ 100) $\mu$ A	$1.2 \times 10^{-4}$	
		(0.1 ~ 1) mA	$6.8 \times 10^{-5}$	
		(1 ~ 10) mA	$7.2 \times 10^{-5}$	
(10 ~ 100) mA	$7.2 \times 10^{-5}$			
Resistance		0 $\Omega$	4.6 $\mu\Omega$	
		(0 ~ 1) $\Omega$	$9.9 \times 10^{-6}$	
		(1 ~ 10) $\Omega$	$2.5 \times 10^{-5}$	
		(10 ~ 100) $\Omega$	$7.7 \times 10^{-6}$	
		(0.1 ~ 1) k $\Omega$	$7.3 \times 10^{-6}$	
		(1 ~ 10) k $\Omega$	$4.9 \times 10^{-6}$	
		(10 ~ 100) k $\Omega$	$7.3 \times 10^{-6}$	
		(0.1 ~ 1) M $\Omega$	$9.6 \times 10^{-6}$	
		(1 ~ 10) M $\Omega$	$1.2 \times 10^{-5}$	
		(10 ~ 100) M $\Omega$	$2.5 \times 10^{-5}$	
		(100 ~ 1 000) M $\Omega$	$3.2 \times 10^{-5}$	
		(1 ~ 10) G $\Omega$	$5.8 \times 10^{-4}$	
(10 ~ 100) G $\Omega$	$1.2 \times 10^{-3}$			
Frequency		1 Hz	0.58 $\mu$ Hz	
		(1 ~ 10) Hz	$5.8 \times 10^{-7}$	
		(10 ~ 100) Hz	$5.8 \times 10^{-7}$	
		(0.1 ~ 1) kHz	$5.8 \times 10^{-7}$	
		(1 ~ 10) kHz	$5.8 \times 10^{-7}$	
		(10 ~ 100) kHz	$5.8 \times 10^{-7}$	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Frequency	40403	(0.1 ~ 1) MHz (1 ~ 10) MHz (10 ~ 30) MHz	$5.8 \times 10^{-7}$ $5.8 \times 10^{-7}$ $1.9 \times 10^{-7}$	Meter calibrators, Digital multimeters, Frequency counters, LCR meters /HCT-CS-276-40403
Multi function calibrator DC Voltage (Meter)		1 mV -1 mV (1 ~ 100) mV (-1 ~ -100) mV (0.1 ~ 1) V (-0.1 ~ -1) V (1 ~ 10) V (-1 ~ -10) V (10 ~ 100) V (-10 ~ -100) V (100 ~ 1 000) V (-100 ~ -1 000) V	0.59 $\mu$ V 0.59 $\mu$ V $1.6 \times 10^{-5}$ $1.6 \times 10^{-5}$ $8.0 \times 10^{-6}$ $8.0 \times 10^{-6}$ $5.5 \times 10^{-6}$ $5.5 \times 10^{-6}$ $8.5 \times 10^{-6}$ $8.5 \times 10^{-6}$ $1.0 \times 10^{-5}$ $1.0 \times 10^{-5}$	
DC Current (Meter)		1 $\mu$ A -1 $\mu$ A (1 ~ 100) $\mu$ A (-1 ~ -100) $\mu$ A (0.1 ~ 1) mA (-0.1 ~ -1) mA (1 ~ 10) mA (-1 ~ -10) mA (10 ~ 100) mA (-10 ~ -100) mA (0.1 ~ 1) A (-0.1 ~ -1) A (1 ~ 10) A (-1 ~ -10) A (10 ~ 20) A (-10 ~ -20) A	2.4 nA 2.4 nA $1.4 \times 10^{-4}$ $1.4 \times 10^{-4}$ $5.6 \times 10^{-5}$ $5.6 \times 10^{-5}$ $5.2 \times 10^{-5}$ $5.2 \times 10^{-5}$ $5.7 \times 10^{-5}$ $5.7 \times 10^{-5}$ $1.2 \times 10^{-4}$ $1.2 \times 10^{-4}$ $1.0 \times 10^{-4}$ $9.9 \times 10^{-5}$ $1.0 \times 10^{-4}$ $9.9 \times 10^{-5}$	
Resistance (Meter)		1 $\Omega$ (1 ~ 100) $\Omega$ (0.1 ~ 1) k $\Omega$ (1 ~ 10) k $\Omega$ (10 ~ 100) k $\Omega$ (0.1 ~ 1) M $\Omega$ (1 ~ 10) M $\Omega$ (10 ~ 100) M $\Omega$ (0.1 ~ 1) G $\Omega$	11 $\mu$ $\Omega$ $6.9 \times 10^{-6}$ $6.9 \times 10^{-6}$ $4.6 \times 10^{-6}$ $7.0 \times 10^{-6}$ $9.3 \times 10^{-6}$ $1.2 \times 10^{-5}$ $2.5 \times 10^{-5}$ $6.2 \times 10^{-4}$	



404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC Voltage (Meter)	40403	(10 ~ 40) Hz		
		1 mV	4.8 μV	
		(1 ~ 100) mV	$1.8 \times 10^{-4}$	
		(0.1 ~ 1) V	$1.2 \times 10^{-4}$	
		(1 ~ 10) V	$1.3 \times 10^{-4}$	
		(10 ~ 100) V	$1.3 \times 10^{-4}$	
		(100 ~ 1 000) V	$3.7 \times 10^{-4}$	
		(40 ~ 500) Hz		
		1 mV	4.7 μV	
		(1 ~ 100) mV	$1.5 \times 10^{-4}$	
		(0.1 ~ 1) V	$5.9 \times 10^{-5}$	
		(1 ~ 10) V	$5.8 \times 10^{-5}$	
		(10 ~ 100) V	$7.1 \times 10^{-5}$	
		(100 ~ 1 000) V	$8.7 \times 10^{-5}$	
		500 Hz ~ 1 kHz		
		1 mV	4.7 μV	
		(1 ~ 100) mV	$1.5 \times 10^{-4}$	
		(0.1 ~ 1) V	$5.9 \times 10^{-5}$	
		(1 ~ 10) V	$5.8 \times 10^{-5}$	
		(10 ~ 100) V	$7.1 \times 10^{-5}$	
		(100 ~ 1 000) V	$8.7 \times 10^{-5}$	
		(1 ~ 10) kHz		
		1 mV	4.7 μV	
		(1 ~ 100) mV	$1.5 \times 10^{-4}$	
(0.1 ~ 1) V	$6.1 \times 10^{-5}$			
(1 ~ 10) V	$5.8 \times 10^{-5}$			
(10 ~ 100) V	$7.1 \times 10^{-5}$			
(100 ~ 1 000) V	$2.0 \times 10^{-4}$			
(10 ~ 20) kHz				
1 mV	4.7 μV			
(1 ~ 100) mV	$1.5 \times 10^{-4}$			
(0.1 ~ 1) V	$7.7 \times 10^{-5}$			
(1 ~ 10) V	$5.8 \times 10^{-5}$			
(10 ~ 100) V	$7.1 \times 10^{-5}$			
(100 ~ 1 000) V	$2.0 \times 10^{-4}$			

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC Voltage (Meter)	40403	(20 ~ 50) kHz 1 mV (1 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V	4.9 μV $2.2 \times 10^{-4}$ $9.4 \times 10^{-5}$ $9.4 \times 10^{-5}$ $1.1 \times 10^{-4}$	
		(50 ~ 100) kHz 1 mV (1 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V	6.4 μV $5.5 \times 10^{-4}$ $1.4 \times 10^{-4}$ $1.3 \times 10^{-4}$ $2.1 \times 10^{-4}$	
AC Current (Meter)	40403	(10 ~ 40) Hz 1 μA (1 ~ 100) μA (0.1 ~ 1) mA (1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 10) A (10 ~ 20) A	12 nA $3.1 \times 10^{-4}$ $2.3 \times 10^{-4}$ $2.3 \times 10^{-4}$ $2.3 \times 10^{-4}$ $3.4 \times 10^{-4}$ $1.5 \times 10^{-4}$ $1.5 \times 10^{-4}$	
		(40 ~ 500) Hz 1 μA (1 ~ 100) μA (0.1 ~ 1) mA (1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 10) A (10 ~ 20) A	9.4 nA $2.2 \times 10^{-4}$ $1.7 \times 10^{-4}$ $1.7 \times 10^{-4}$ $1.6 \times 10^{-4}$ $3.4 \times 10^{-4}$ $5.5 \times 10^{-4}$ $4.7 \times 10^{-4}$	
		500 Hz ~ 1 kHz 1 μA (1 ~ 100) μA (0.1 ~ 1) mA (1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 10) A (10 ~ 20) A	9.4 nA $2.2 \times 10^{-4}$ $1.7 \times 10^{-4}$ $1.7 \times 10^{-4}$ $1.6 \times 10^{-4}$ $3.4 \times 10^{-4}$ $5.5 \times 10^{-4}$ $4.7 \times 10^{-4}$	
		(1 ~ 5) kHz 1 μA (1 ~ 100) μA (0.1 ~ 1) mA (1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A	14 nA $4.9 \times 10^{-4}$ $3.8 \times 10^{-4}$ $3.4 \times 10^{-4}$ $3.2 \times 10^{-4}$ $6.6 \times 10^{-4}$	
		(5 ~ 10) kHz 1 μA (1 ~ 100) μA (0.1 ~ 1) mA (1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A	76 nA $2.1 \times 10^{-3}$ $2.1 \times 10^{-3}$ $1.9 \times 10^{-3}$ $1.5 \times 10^{-3}$ $8.3 \times 10^{-3}$	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC Voltage (Source)	40403	1 mV	0.12 $\mu$ V	
		-1 mV	0.12 $\mu$ V	
		(1 ~ 100) mV	$4.9 \times 10^{-6}$	
		(-1 ~ -100) mV	$4.9 \times 10^{-6}$	
		(0.1 ~ 1) V	$3.9 \times 10^{-6}$	
		(-0.1 ~ -1) V	$3.9 \times 10^{-6}$	
		(1 ~ 10) V	$3.7 \times 10^{-6}$	
		(-1 ~ -10) V	$3.7 \times 10^{-6}$	
		(10 ~ 100) V	$5.2 \times 10^{-6}$	
		(-10 ~ -100) V	$5.2 \times 10^{-6}$	
(100 ~ 1 000) V	$5.5 \times 10^{-6}$			
(-100 ~ -1 000) V	$5.5 \times 10^{-6}$			
DC Current (Source)		1 $\mu$ A	0.48 nA	
		-1 $\mu$ A	0.48 nA	
		(1 ~ 100) $\mu$ A	$6.1 \times 10^{-5}$	
		(-1 ~ -100) $\mu$ A	$6.1 \times 10^{-5}$	
		(0.1 ~ 1) mA	$1.1 \times 10^{-5}$	
		(-0.1 ~ -1) mA	$1.1 \times 10^{-5}$	
		(1 ~ 10) mA	$1.1 \times 10^{-5}$	
		(-1 ~ -10) mA	$1.1 \times 10^{-5}$	
		(10 ~ 100) mA	$2.3 \times 10^{-5}$	
		(-10 ~ -100) mA	$2.3 \times 10^{-5}$	
		(0.1 ~ 1) A	$3.8 \times 10^{-5}$	
		(-0.1 ~ -1) A	$3.8 \times 10^{-5}$	
		(1 ~ 10) A	$1.6 \times 10^{-4}$	
(-1 ~ -10) A	$1.6 \times 10^{-4}$			
Resistance (Source)		1 $\Omega$	18 $\mu$ $\Omega$	
		(1 ~ 100) $\Omega$	$9.1 \times 10^{-6}$	
		(0.1 ~ 1) k $\Omega$	$9.0 \times 10^{-6}$	
		(1 ~ 10) k $\Omega$	$9.0 \times 10^{-6}$	
		(10 ~ 100) k $\Omega$	$9.1 \times 10^{-6}$	
		(0.1 ~ 1) M $\Omega$	$1.0 \times 10^{-5}$	
		(1 ~ 10) M $\Omega$	$1.1 \times 10^{-5}$	
		(10 ~ 100) M $\Omega$	$3.2 \times 10^{-5}$	
AC Current (Source)		(10 ~ 40) Hz		
		1 $\mu$ A	0.42 nA	
		(1 ~ 100) $\mu$ A	$4.2 \times 10^{-4}$	
		(0.1 ~ 1) mA	$4.1 \times 10^{-4}$	
		(1 ~ 10) mA	$4.1 \times 10^{-4}$	
		(10 ~ 100) mA	$4.1 \times 10^{-4}$	
		(0.1 ~ 1) A	$8.1 \times 10^{-4}$	
		(1 ~ 10) A	$1.0 \times 10^{-3}$	
		(40 ~ 500) Hz		
		1 $\mu$ A	0.42 nA	
		(1 ~ 100) $\mu$ A	$4.2 \times 10^{-4}$	
		(0.1 ~ 1) mA	$4.1 \times 10^{-4}$	
		(1 ~ 10) mA	$4.1 \times 10^{-4}$	
		(10 ~ 100) mA	$4.1 \times 10^{-4}$	
		(0.1 ~ 1) A	$8.1 \times 10^{-4}$	
		(1 ~ 10) A	$1.0 \times 10^{-3}$	
		500 Hz ~ 1 kHz		
		1 $\mu$ A	0.42 nA	
		(1 ~ 100) $\mu$ A	$4.2 \times 10^{-4}$	
		(0.1 ~ 1) mA	$4.1 \times 10^{-4}$	
(1 ~ 10) mA	$4.1 \times 10^{-4}$			
(10 ~ 100) mA	$4.1 \times 10^{-4}$			
(0.1 ~ 1) A	$8.1 \times 10^{-4}$			
(1 ~ 10) A	$1.0 \times 10^{-3}$			

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC Current (Source)	40403	(1 ~ 5) kHz 1 μA (1 ~ 100) μA (0.1 ~ 1) mA (1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 10) A	0.46 nA $4.6 \times 10^{-4}$ $4.1 \times 10^{-4}$ $4.2 \times 10^{-4}$ $4.1 \times 10^{-4}$ $9.4 \times 10^{-4}$ $3.1 \times 10^{-3}$	
DC Voltage (Electrical temperature)		(5 ~ 10) kHz 1 μA (1 ~ 100) μA (0.1 ~ 1) mA (1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 10) A	0.45 nA $4.6 \times 10^{-4}$ $4.1 \times 10^{-4}$ $4.2 \times 10^{-4}$ $4.1 \times 10^{-4}$ $1.0 \times 10^{-3}$ $3.0 \times 10^{-3}$	
Time Mark		0 mV (-10 ~ 0) mV (0 ~ 1) mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V	0.49 μV $4.9 \times 10^{-5}$ $4.9 \times 10^{-4}$ $4.9 \times 10^{-5}$ $4.9 \times 10^{-6}$ $5.8 \times 10^{-5}$	
Frequency		1 ns (1 ~ 10) ns (10 ~ 100) ns (0.1 ~ 1) μs (1 ~ 10) μs (10 ~ 100) μs (0.1 ~ 1) ms (1 ~ 10) ms (10 ~ 100) ms (0.1 ~ 1) s	2.7 ps $2.7 \times 10^{-4}$ $2.7 \times 10^{-5}$ $2.8 \times 10^{-6}$ $6.4 \times 10^{-7}$ $5.8 \times 10^{-7}$ $5.8 \times 10^{-7}$ $5.8 \times 10^{-7}$ $5.8 \times 10^{-7}$ $5.8 \times 10^{-7}$	
Capacitance		1 Hz (1 ~ 10) Hz (10 ~ 100) Hz (0.1 ~ 1) kHz (1 ~ 10) kHz (10 ~ 100) kHz (0.1 ~ 1) MHz (1 ~ 10) MHz (10 ~ 100) MHz (0.1 ~ 1) GHz	0.58 μHz $5.8 \times 10^{-7}$ $5.8 \times 10^{-7}$ $5.8 \times 10^{-7}$ $5.8 \times 10^{-7}$ $5.8 \times 10^{-7}$ $5.8 \times 10^{-7}$ $5.8 \times 10^{-7}$ $5.8 \times 10^{-7}$	
		DC 1 mF (1 ~ 10) mF (10 ~ 110) mF	0.60 μF $1.1 \times 10^{-4}$ $4.8 \times 10^{-5}$	
		50 Hz 100 μF 100 Hz 1 μF (1 ~ 10) μF (10 ~ 100) μF	32 nF  0.39 nF $2.5 \times 10^{-4}$ $3.2 \times 10^{-4}$	
		120 Hz 100 μF	32 nF	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Capacitance	40403	1 kHz		
		1 pF	0.39 fF	
		(1 ~ 10) pF	$3.8 \times 10^{-4}$	
		(10 ~ 100) pF	$2.6 \times 10^{-4}$	
		(0.1 ~ 1) nF	$2.6 \times 10^{-4}$	
		(1 ~ 10) nF	$2.5 \times 10^{-4}$	
		(10 ~ 100) nF	$2.5 \times 10^{-4}$	
		(1 ~ 10) $\mu$ F	$2.5 \times 10^{-4}$	
		(0.1 ~ 1) $\mu$ F	$2.5 \times 10^{-4}$	
		(1 ~ 10) kHz		
		1 pF	0.39 fF	
		(1 ~ 10) pF	$3.8 \times 10^{-4}$	
		(10 ~ 100) pF	$3.8 \times 10^{-4}$	
		(0.1 ~ 1) nF	$3.8 \times 10^{-4}$	
		(1 ~ 10) nF	$3.8 \times 10^{-4}$	
		(10 ~ 100) nF	$3.8 \times 10^{-4}$	
		(0.1 ~ 1) $\mu$ F	$3.8 \times 10^{-4}$	
		(10 ~ 100) kHz		
		1 pF	0.39 fF	
		(1 ~ 10) pF	$3.8 \times 10^{-4}$	
		(10 ~ 100) pF	$3.8 \times 10^{-4}$	
		(0.1 ~ 1) nF	$3.8 \times 10^{-4}$	
		(1 ~ 10) nF	$3.8 \times 10^{-4}$	
		(100 ~ 500) kHz		
		1 pF	0.39 fF	
		(1 ~ 10) pF	$3.8 \times 10^{-4}$	
		(10 ~ 100) pF	$3.8 \times 10^{-4}$	
		(0.1 ~ 1) nF	$3.8 \times 10^{-4}$	
		(0.5 ~ 1) MHz		
		1 pF	0.63 fF	
		(1 ~ 10) pF	$6.3 \times 10^{-4}$	
		(10 ~ 100) pF	$6.3 \times 10^{-4}$	
		(0.1 ~ 1) nF	$6.3 \times 10^{-4}$	
Inductance(Source)		1 kHz		
		100 $\mu$ H	28 nH	
		(0.1 ~ 1) mH	$2.7 \times 10^{-4}$	
		(1 ~ 10) mH	$2.7 \times 10^{-4}$	
		(10 ~ 100) mH	$2.7 \times 10^{-4}$	
		(0.1 ~ 1) H	$2.7 \times 10^{-4}$	
		(1 ~ 10) H	$2.8 \times 10^{-4}$	
		(1 ~ 10) kHz		
		100 $\mu$ H	28 nH	
		(0.1 ~ 1) mH	$2.7 \times 10^{-4}$	
		(1 ~ 10) mH	$2.7 \times 10^{-4}$	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Oscilloscope calibrators DC Voltage	40404	(±) 0 mV (0 ~ 2.5) mV (2.5 ~ 5) mV (5 ~ 10) mV (10 ~ 25) mV (25 ~ 50) mV (50 ~ 100) mV (100 ~ 250) mV (250 ~ 500) mV (0.5 ~ 1) V (1 ~ 2.5) V (2.5 ~ 5) V (5 ~ 10) V (10 ~ 25) V (25 ~ 50) V (50 ~ 100) V (100 ~ 150) V (150 ~ 200) V	0.060 μV $1.1 \times 10^{-4}$ $5.6 \times 10^{-5}$ $6.4 \times 10^{-5}$ $2.6 \times 10^{-5}$ $1.3 \times 10^{-5}$ $5.8 \times 10^{-5}$ $2.4 \times 10^{-5}$ $1.3 \times 10^{-5}$ $5.8 \times 10^{-5}$ $2.4 \times 10^{-5}$ $1.2 \times 10^{-5}$ $5.8 \times 10^{-5}$ $2.4 \times 10^{-5}$ $1.3 \times 10^{-5}$ $5.8 \times 10^{-5}$ $4.0 \times 10^{-5}$ $3.1 \times 10^{-5}$	Digital multimeters, Counters, Power meters, Oscilloscopes, Spectrum analyzers AC Calibrators / HCT-CS-083-40404
Square/Edge Wave Voltage		1 kHz 5 mV (5 ~ 10) mV (10 ~ 25) mV (25 ~ 50) mV (50 ~ 100) mV (100 ~ 250) mV (250 ~ 500) mV (0.5 ~ 1) V (1 ~ 2.5) V (2.5 ~ 5) V (5 ~ 10) V (10 ~ 25) V (25 ~ 50) V (50 ~ 100) V (100 ~ 130) V (130 ~ 200) V	3.2 μV $7.0 \times 10^{-4}$ $3.4 \times 10^{-4}$ $2.2 \times 10^{-4}$ $1.7 \times 10^{-4}$ $1.4 \times 10^{-4}$ $7.4 \times 10^{-5}$ $1.3 \times 10^{-2}$ $1.2 \times 10^{-4}$ $1.7 \times 10^{-4}$ $1.6 \times 10^{-4}$ $1.3 \times 10^{-4}$ $1.0 \times 10^{-4}$ $1.4 \times 10^{-4}$ $1.1 \times 10^{-4}$ $8.0 \times 10^{-5}$	
Square/Edge Wave Frequency		100 kHz 10 mV (10 ~ 25) mV (25 ~ 50) mV (50 ~ 100) mV (100 ~ 250) mV (250 ~ 500) mV (0.5 ~ 1) V (1 ~ 2.5) V	28 μV $2.7 \times 10^{-3}$ $1.7 \times 10^{-3}$ $1.3 \times 10^{-3}$ $1.1 \times 10^{-3}$ $1.5 \times 10^{-3}$ $1.2 \times 10^{-3}$ $8.2 \times 10^{-4}$	
Edge TD Pulse Drive		10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz 10 kHz ~ 10 MHz	5.8 μHz $5.8 \times 10^{-8}$ $2.8 \times 10^{-8}$ $5.8 \times 10^{-8}$	
Edge TD Pulse Drive		(10 ~ 100) Hz 11 V (11 ~ 100) V (0.1 ~ 1) kHz	5.4 mV $5.5 \times 10^{-5}$	
Edge TD Pulse Drive		11 V (11 ~ 100) V	5.4 mV $5.5 \times 10^{-5}$	
Edge Duty Cycle		50 %	0.058 %	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Edge Rise Time	40404	300 ps (300 ~ 500) ps	0.64 ps $1.5 \times 10^{-3}$	
Leveled Sine Wave (Harmonic)		50 kHz ~ 6 GHz -10 dBc (-10 ~ -80) dBc	0.64 dB 0.64 dB	
RF output levels (V : pp)		50 kHz ~ 600 MHz 60 mV (60 ~ 300) mV (300 ~ 600) mV 600 mV ~ 5.5 V	1.0 mV $1.6 \times 10^{-2}$ $1.5 \times 10^{-2}$ $1.6 \times 10^{-2}$	
		600 MHz ~ 1 GHz 60 mV (60 ~ 300) mV (300 ~ 600) mV 600 mV ~ 3.5 V	1.0 mV $1.6 \times 10^{-2}$ $1.5 \times 10^{-2}$ $1.6 \times 10^{-2}$	
		(1 ~ 2) GHz 60 mV (60 ~ 300) mV (300 ~ 600) mV 600 mV ~ 3 V	1.0 mV $1.6 \times 10^{-2}$ $1.5 \times 10^{-2}$ $1.6 \times 10^{-2}$	
		(2 ~ 6) GHz 60 mV (60 ~ 300) mV (300 ~ 600) mV 600 mV ~ 1.2 V	1.0 mV $1.6 \times 10^{-2}$ $1.5 \times 10^{-2}$ $1.6 \times 10^{-2}$	
Leveled Sine Wave (Frequency)		500 MHz (0.5 ~ 6) GHz	1.4 Hz $3.2 \times 10^{-8}$	
Leveled Sine Wave (Amplitude)		10 Hz 5 mV (5 ~ 100) mV (0.1 ~ 1) V (1 ~ 5.5) V	4.3 $\mu$ V $5.9 \times 10^{-5}$ $6.6 \times 10^{-4}$ $1.2 \times 10^{-4}$	
		(0.01 ~ 50) kHz 5 mV (5 ~ 100) mV (0.1 ~ 1) V (1 ~ 5.5) V	5.1 $\mu$ V $1.1 \times 10^{-4}$ $6.6 \times 10^{-4}$ $1.4 \times 10^{-4}$	
Wave Generator (Square)		10 Hz 10 mV (10 ~ 900) mV (0.9 ~ 2.5) V (2.5 ~ 3.75) V (3.75 ~ 55) V	3.3 $\mu$ V $6.7 \times 10^{-5}$ $3.2 \times 10^{-4}$ $2.1 \times 10^{-4}$ $4.4 \times 10^{-5}$	
		(0.01 ~ 1) kHz 10 mV (10 ~ 900) mV (0.9 ~ 2.5) V (2.5 ~ 3.75) V (3.75 ~ 55) V	3.0 $\mu$ V $2.6 \times 10^{-5}$ $2.6 \times 10^{-4}$ $1.9 \times 10^{-4}$ $2.1 \times 10^{-5}$	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Wave Generator (Square)	40404	(1 ~ 10) kHz		
		2.5 V	0.73 mV	
		(2.5 ~ 3.75) V	$1.3 \times 10^{-5}$	
		(3.75 ~ 55) V	$4.3 \times 10^{-5}$	
Wave Generator (Sine)		10 Hz		
		10 mV	3.1 $\mu$ V	
		(0.01 ~ 55) V	$3.0 \times 10^{-5}$	
		(0.01 ~ 1) kHz		
		10 mV	3.0 $\mu$ V	
		(0.01 ~ 55) V	$2.1 \times 10^{-5}$	
Wave Generator (Triangle)		10 Hz		
		10 mV	3.0 $\mu$ V	
		(0.01 ~ 55) V	$2.6 \times 10^{-5}$	
		(0.01 ~ 1) kHz		
		10 mV	3.0 $\mu$ V	
		(0.01 ~ 55) V	$1.8 \times 10^{-5}$	
Pulse Generator (Priod)		10 ns	0.58 ps	
		(0.01 ~ 20) $\mu$ s	$2.9 \times 10^{-5}$	
		(20 ~ 100) $\mu$ s	$5.8 \times 10^{-6}$	
Pulse Generator (Width)		4 ns	1.2 ps	
		(4 ~ 100) ns	$1.0 \times 10^{-3}$	
Time mark	1 ns	2.7 ps		
	(1 ~ 2) ns	$1.4 \times 10^{-3}$		
	(2 ~ 5) ns	$5.4 \times 10^{-4}$		
	(5 ~ 10) ns	$2.7 \times 10^{-4}$		
	(10 ~ 20) ns	$1.4 \times 10^{-4}$		
	(20 ~ 50) ns	$5.4 \times 10^{-5}$		
	(50 ~ 100) ns	$2.7 \times 10^{-5}$		
	(100 ~ 200) ns	$1.4 \times 10^{-5}$		
	(200 ~ 500) ns	$5.4 \times 10^{-6}$		
	(0.5 ~ 1) $\mu$ s	$2.8 \times 10^{-6}$		
	(1 ~ 2) $\mu$ s	$1.4 \times 10^{-6}$		
	(2 ~ 5) $\mu$ s	$5.5 \times 10^{-7}$		
	(5 ~ 10) $\mu$ s	$6.4 \times 10^{-7}$		
	(10 ~ 20) $\mu$ s	$3.2 \times 10^{-7}$		
	(20 ~ 50) $\mu$ s	$1.3 \times 10^{-7}$		
	(50 ~ 100) $\mu$ s	$5.8 \times 10^{-7}$		
	(100 ~ 200) $\mu$ s	$2.9 \times 10^{-7}$		
	(200 ~ 500) $\mu$ s	$1.2 \times 10^{-7}$		
	(0.5 ~ 1) ms	$5.8 \times 10^{-7}$		
	(1 ~ 2) ms	$2.9 \times 10^{-7}$		
	(2 ~ 5) ms	$1.2 \times 10^{-7}$		
	(5 ~ 10) ms	$5.8 \times 10^{-7}$		
	(10 ~ 20) ms	$2.9 \times 10^{-7}$		
	(20 ~ 50) ms	$1.2 \times 10^{-7}$		
	(50 ~ 100) ms	$5.8 \times 10^{-7}$		
	(100 ~ 200) ms	$2.9 \times 10^{-7}$		
	(200 ~ 500) ms	$1.2 \times 10^{-7}$		
	(0.5 ~ 1) s	$5.8 \times 10^{-7}$		
	(1 ~ 2) s	$2.9 \times 10^{-7}$		
	(2 ~ 5) s	$1.2 \times 10^{-7}$		
	(5 ~ 10) s	$5.8 \times 10^{-7}$		
	(10 ~ 20) s	$2.9 \times 10^{-7}$		



404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Frequency	40404	50 mHz (50 ~ 100) mHz (100 ~ 200) mHz (200 ~ 500) mHz (0.5 ~ 1) Hz (1 ~ 2) Hz (2 ~ 5) Hz (5 ~ 10) Hz (10 ~ 20) Hz (20 ~ 50) Hz (50 ~ 100) Hz (100 ~ 200) Hz (200 ~ 500) Hz (0.5 ~ 1) kHz (1 ~ 2) kHz (2 ~ 5) kHz (5 ~ 10) kHz (10 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 200) kHz (200 ~ 500) kHz (0.5 ~ 1) MHz (1 ~ 2) MHz (2 ~ 5) MHz (5 ~ 10) MHz (10 ~ 20) MHz (20 ~ 50) MHz (50 ~ 100) MHz (100 ~ 200) MHz (200 ~ 500) MHz (0.5 ~ 1.1) GHz	5.8 nHz $5.8 \times 10^{-7}$ $2.9 \times 10^{-7}$ $1.2 \times 10^{-7}$ $5.8 \times 10^{-7}$ $2.9 \times 10^{-7}$ $1.2 \times 10^{-7}$ $5.8 \times 10^{-7}$ $2.9 \times 10^{-7}$ $1.2 \times 10^{-7}$ $5.8 \times 10^{-7}$ $2.9 \times 10^{-7}$ $1.2 \times 10^{-7}$ $5.8 \times 10^{-7}$ $2.9 \times 10^{-7}$ $1.2 \times 10^{-7}$ $5.8 \times 10^{-7}$ $2.9 \times 10^{-7}$ $1.2 \times 10^{-7}$ $5.8 \times 10^{-7}$ $2.9 \times 10^{-7}$ $1.2 \times 10^{-7}$ $5.8 \times 10^{-7}$ $2.9 \times 10^{-7}$ $1.2 \times 10^{-7}$ $5.8 \times 10^{-7}$ $2.9 \times 10^{-7}$ $1.2 \times 10^{-7}$ $5.8 \times 10^{-7}$ $2.9 \times 10^{-7}$ $1.2 \times 10^{-7}$ $5.2 \times 10^{-7}$	
MeasZ (Resistance)		40 Ω 40 Ω ~ 1.5 MΩ	12 mΩ $2.8 \times 10^{-4}$	
MeasZ (Capacitance)		50 pF (50 ~ 100) pF	0.10 pF $1.8 \times 10^{-3}$	
Video signal generators	40406			Frequency counters, Video signal analyzers, Oscilloscopes /HCT-CS-084-40406
DOT Frequency		10 kHz ~ 1 000 MHz	$5.8 \times 10^{-7}$	
SYNC Frequency		50 Hz ~ 1 MHz	$5.8 \times 10^{-7}$	
SYNC WIDTH(Time)		1 μs (1 ~ 100) μs	1.2 ns $1.2 \times 10^{-3}$	
Analog Video Level		100 mV (100 ~ 1 000) mV	1.2 mV $1.2 \times 10^{-2}$	
Analog Sync Level		1 V (1 ~ 5) V	20 mV $1.4 \times 10^{-2}$	
Audio Level		100 mV (100 ~ 1 000) mV	1.2 mV $1.2 \times 10^{-2}$	
S-Video Level		100 mV (100 ~ 1 000) mV	1.2 mV $1.2 \times 10^{-2}$	
Component Level		100 mV (100 ~ 1 000) mV	1.2 mV $1.2 \times 10^{-2}$	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Scart Video Level	40406	100 mV (100 ~ 1 000) mV	1.2 mV $1.2 \times 10^{-2}$	
Scart Audio Level		100 mV (100 ~ 1 000) mV	1.2 mV $1.2 \times 10^{-2}$	
NTSC,PAL,SECAM H-Timing Test (Time)		100 ns (100 ~ 300) ns	0.60 ns $3.9 \times 10^{-2}$	
		300 ns ~ 9 $\mu$ s	$1.4 \times 10^{-3}$	
(Level)		50 mV (50 ~ 900) mV	0.62 mV $6.2 \times 10^{-3}$	
NTSC,PAL,SECAM COLOR BAR (LUMINANCE Level)		50 mV (50 ~ 900) mV	0.32 mV $3.2 \times 10^{-3}$	
NTSC,PAL,SECAM COLOR BAR (CHROMINANCE Level)		50 mV (50 ~ 900) mV	0.32 mV $3.2 \times 10^{-3}$	
NTSC,PAL,SECAM COLOR BAR (CHROMINANCE Phase)		0 ° ~ 360 °	0.55 °	
		RF Frequency	10 kHz ~ 1 000 MHz $5.8 \times 10^{-7}$	
Sound Frequency	10 Hz ~ 1 MHz $5.8 \times 10^{-7}$			
SUB CARRIER Frequency				
	NTSC PAL	3.579 545 MHz 4.433 619 MHz	0.58 Hz 0.58 Hz	
Audio distortion analyzers/meters Distortion meter Input frequency	40407	1 Hz ~ 1 MHz	$6.2 \times 10^{-5}$	Multimeter calibrators, Distortion meter calibrators /HCT-CS-085-40407
AC input levels		2 mV		
		10 Hz	5.3 $\mu$ V	
		10 Hz ~ 20 kHz	$2.5 \times 10^{-3}$	
		(20 ~ 50) kHz	$2.6 \times 10^{-3}$	
(50 ~ 100) kHz		$3.6 \times 10^{-3}$		
(2 ~ 10) mV				
10 Hz		9.8 $\mu$ V		
10 Hz ~ 20 kHz		$1.1 \times 10^{-3}$		
(20 ~ 50) kHz		$1.2 \times 10^{-3}$		
(50 ~ 100) kHz		$1.8 \times 10^{-3}$		
(10 ~ 100) mV				
10 Hz		74 $\mu$ V		
10 Hz ~ 20 kHz		$6.3 \times 10^{-4}$		
(20 ~ 50) kHz		$6.5 \times 10^{-4}$		
(50 ~ 100) kHz		$9.4 \times 10^{-4}$		
(0.1 ~ 1) V				
10 Hz	0.69 mV			
10 Hz ~ 20 kHz	$6.1 \times 10^{-4}$			
(20 ~ 50) kHz	$6.2 \times 10^{-4}$			
(50 ~ 100) kHz	$6.3 \times 10^{-4}$			
(1 ~ 10) V				
10 Hz	6.9 mV			
10 Hz ~ 20 kHz	$6.1 \times 10^{-4}$			
(20 ~ 50) kHz	$6.2 \times 10^{-4}$			
(50 ~ 100) kHz	$6.2 \times 10^{-4}$			

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
AC input levels	40407	(10 ~ 100) V			
		10 Hz	70 mV		
		10 Hz ~ 20 kHz	$6.2 \times 10^{-4}$		
		(20 ~ 50) kHz	$6.2 \times 10^{-4}$		
		(50 ~ 100) kHz	$6.5 \times 10^{-4}$		
		(100 ~ 300) V			
		50 Hz	0.14 V		
		50 Hz ~ 1 kHz	$2.3 \times 10^{-4}$		
DC input levels			1 mV	6.2 $\mu$ V	
			1 mV ~ 100 V	$6.1 \times 10^{-4}$	
			(100 ~ 300) V	$2.1 \times 10^{-4}$	
Input distortion			(20 Hz ~ 20 kHz)		
		(-10 ~ -60) dB	0.31 dB		
		(-60 ~ -70) dB	0.38 dB		
		(-70 ~ -80) dB	0.55 dB		
		20 Hz ~ 20 kHz			
		0.01%	0.000 55 %		
		(0.01 ~ 30) %	$3.1 \times 10^{-2}$		
Distortion meter calibrators					
Output level		100 mV			
		20 Hz	65 $\mu$ V		
		20 Hz ~ 1 kHz	$6.3 \times 10^{-4}$		
		(1 ~ 20) kHz	$7.4 \times 10^{-4}$		
		(20 ~ 100) kHz	$1.1 \times 10^{-3}$		
		(0.1 ~ 1) V			
		20 Hz	0.63 mV		
		20 Hz ~ 1 kHz	$6.2 \times 10^{-4}$		
		(1 ~ 20) kHz	$6.7 \times 10^{-4}$		
		(20 ~ 100) kHz	$9.3 \times 10^{-4}$		
Output level		(1 ~ 10) V			
		20 Hz	6.3 mV		
		20 Hz ~ 1 kHz	$6.2 \times 10^{-4}$		
		(1 ~ 20) kHz	$6.7 \times 10^{-4}$		
		(20 ~ 100) kHz	$9.3 \times 10^{-4}$		
Output distortion		20 Hz ~ 100 kHz			
		(-10 ~ -20) dB	0.88 dB		
		20 Hz ~ 100 kHz			
		(-20 ~ -50) dB	1.1 dB		
		20 Hz ~ 100 kHz			
		(-50 ~ -80) dB	1.4 dB		
LF filters	40408			Audio analyzers, Function generators /HCT-CS-087-40408	
Frequency		30 Hz ~ 30 MHz	$5.8 \times 10^{-4}$		
Level		(0 ~ 90) dB			
		20 Hz ~ 100 kHz	0.010 dB		
LF/audio signal analyzers	40409			Multimeter calibrators, Digital multimeters /HCT-CS-088-40409	
Output Frequency		1 Hz ~1 MHz	$6.2 \times 10^{-5}$		

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Output level	40409	2 mV		
		20 Hz	7.9 μV	
		20 Hz ~ 1 kHz	$3.5 \times 10^{-3}$	
		(1 ~ 20) kHz	$6.0 \times 10^{-3}$	
		(20 ~ 100) kHz	$1.3 \times 10^{-2}$	
		(2 ~ 100) mV		
		20 Hz	20 μV	
		20 Hz ~ 1 kHz	$1.8 \times 10^{-4}$	
		(1 ~ 20) kHz	$4.5 \times 10^{-4}$	
		(20 ~ 100) kHz	$1.1 \times 10^{-3}$	
		(0.1 ~ 1) V		
		20 Hz	0.16 mV	
		20 Hz ~ 1 kHz	$1.2 \times 10^{-4}$	
		(1 ~ 20) kHz	$2.9 \times 10^{-4}$	
		(20 ~ 100) kHz	$8.2 \times 10^{-4}$	
		(1 ~ 10) V		
		20 Hz	1.6 mV	
		20 Hz ~ 1 kHz	$1.2 \times 10^{-4}$	
		(1 ~ 20) kHz	$2.9 \times 10^{-4}$	
		(20 ~ 100) kHz	$8.2 \times 10^{-4}$	
		(10 ~ 100) V		
		20 Hz	16 mV	
		20 Hz ~ 1 kHz	$1.7 \times 10^{-4}$	
		(1 ~ 20) kHz	$3.5 \times 10^{-4}$	
(20 ~ 100) kHz	$8.1 \times 10^{-4}$			
(10 ~ -10) dBm				
20 Hz	0.006 0 dB			
20 Hz ~ 20 kHz	0.006 3 dB			
(20 ~ 50) kHz	0.008 4 dB			
(50 ~ 100) kHz	0.008 5 dB			
(-10 ~ -30) dBm				
20 Hz	0.006 0 dB			
20 Hz ~ 20 kHz	0.008 0 dB			
(20 ~ 50) kHz	0.010 dB			
(50 ~ 100) kHz	0.010 dB			
(-30 ~ -40) dBm				
20 Hz	0.006 0 dB			
20 Hz ~ 20 kHz	0.006 8 dB			
(20 ~ 50) kHz	0.010 dB			
(50 ~ 100) kHz	0.024 dB			
Output DC Offset		-20 V ~ 0 mV	$6.2 \times 10^{-5}$	
		0 mV	6.2 μV	
		0 mV ~ 20 V	$6.2 \times 10^{-5}$	
Output flatness		20 Hz ~ 20 kHz	0.006 3 dB	
		(20 ~ 100) kHz	0.009 2 dB	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Output amplitude	40409	20 Hz ~ 1 kHz (-10 ~ -60) dB	0.061 dB	
		(1 ~ 20) kHz (-10 ~ -60) dB	0.11 dB	
		(20 ~ 50) kHz (-10 ~ -60) dB	0.11 dB	
		(50 ~ 100) kHz (-10 ~ -60) dB	0.11 dB	
Output impedance	40409	50 Ω	6.2 mΩ	
		600 Ω	62 mΩ	
Input frequency	40409	1 Hz ~ 1 MHz	$6.2 \times 10^{-5}$	
AC input levels	40409	2 mV		
		10 Hz	7.8 μV	
		10 Hz ~ 20 kHz	$3.8 \times 10^{-3}$	
		(20 ~ 50) kHz	$3.9 \times 10^{-3}$	
		(50 ~ 100) kHz	$4.6 \times 10^{-3}$	
		(2 ~ 100) mV		
		10 Hz	43 μV	
		10 Hz ~ 20 kHz	$1.6 \times 10^{-4}$	
		(20 ~ 50) kHz	$2.4 \times 10^{-4}$	
		(50 ~ 100) kHz	$5.7 \times 10^{-4}$	
		(0.1 ~ 1) V		
		10 Hz	0.14 mV	
		10 Hz ~ 20 kHz	$9.0 \times 10^{-5}$	
		(20 ~ 50) kHz	$1.5 \times 10^{-4}$	
		(50 ~ 100) kHz	$1.5 \times 10^{-4}$	
		(1 ~ 10) V		
10 Hz	3.4 mV			
10 Hz ~ 20 kHz	$9.0 \times 10^{-5}$			
(20 ~ 50) kHz	$1.2 \times 10^{-4}$			
(50 ~ 100) kHz	$1.4 \times 10^{-4}$			
(10 ~ 100) V				
10 Hz	10 mV			
10 Hz ~ 20 kHz	$1.0 \times 10^{-4}$			
(20 ~ 50) kHz	$1.3 \times 10^{-4}$			
(50 ~ 100) kHz	$2.2 \times 10^{-4}$			
(100 ~ 300) V				
10 Hz	0.13 V			
10 Hz ~ 10 kHz	$2.2 \times 10^{-4}$			
DC input levels	40409	1 mV	6.2 μV	
		(1 ~ 100) mV	$6.3 \times 10^{-5}$	
		(0.1 ~ 100) V	$6.2 \times 10^{-5}$	
		(100 ~ 300) V	$6.2 \times 10^{-5}$	
Filter(weight,low,high pass etc.)	40409	400 Hz ~ 80 kHz	$1.9 \times 10^{-4}$	
Distortion factor	40409	20 Hz ~ 20 kHz (-10 ~ -60) dB	0.31 dB	
		(-60 ~ -70) dB	0.38 dB	
		(-70 ~ -80) dB	0.56 dB	
		20 Hz ~ 20 kHz (0.001 ~ 0.01) %	$5.5 \times 10^{-2}$	
		(0.01 ~ 30) %	$3.1 \times 10^{-2}$	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Line frequency meters Frequency	40410	40 Hz ~ 1 kHz	$5.8 \times 10^{-4}$	Multimeter calibrators /HCT-CS-179-40410
Function generators Frequency Output level	40411	1 Hz ~ 3 GHz	$5.8 \times 10^{-9}$	Frequency counters, Digital multimeters, Spectrum analyzers, Oscilloscopes /HCT-CS-089-40411
		10 mV		
		20 Hz	$7.0 \mu V$	
		20 Hz ~ 20 kHz	$1.3 \times 10^{-3}$	
		(20 ~ 100) kHz	$3.2 \times 10^{-3}$	
		100 kHz ~ 1 MHz	$1.5 \times 10^{-2}$	
		(10 ~ 100) mV		
		20 Hz	$20 \mu V$	
		20 Hz ~ 20 kHz	$8.1 \times 10^{-4}$	
		(20 ~ 100) kHz	$2.0 \times 10^{-3}$	
		100 kHz ~ 1 MHz	$2.6 \times 10^{-2}$	
		(0.1 ~ 1) V		
		20 Hz	$0.16 mV$	
		20 Hz ~ 20 kHz	$2.9 \times 10^{-4}$	
		(20 ~ 100) kHz	$8.2 \times 10^{-4}$	
		100 kHz ~ 1 MHz	$2.4 \times 10^{-2}$	
		(1 ~ 10) V		
		20 Hz	$1.6 mV$	
		20 Hz ~ 20 kHz	$5.5 \times 10^{-4}$	
		(20 ~ 100) kHz	$1.8 \times 10^{-3}$	
		100 kHz ~ 1 MHz	$3.0 \times 10^{-2}$	
		(10 ~ 100) V		
		20 Hz	$16 mV$	
		20 Hz ~ 20 kHz	$5.5 \times 10^{-4}$	
		(20 ~ 100) kHz	$1.8 \times 10^{-3}$	
		(-60 ~ 20) dBm		
		20 Hz ~ 20 kHz	$0.007 dB$	
		(20 ~ 50) kHz	$0.008 dB$	
		(50 ~ 100) kHz	$0.013 dB$	
		100 kHz ~ 100 MHz	$0.16 dB$	
		(-20 V ~ 0 mV)	$5.8 \times 10^{-4}$	
DC Offset		0 mV	$5.8 \mu V$	
		(0 mV ~ 20 V)	$5.8 \times 10^{-4}$	
		(-10 ~ 10) dB		
Output flatness		20 Hz ~ 100 kHz	$0.016 dB$	
		100 kHz ~ 1 GHz	$0.018 dB$	
		(-80 ~ 0) dB		
Distortion factor		20 Hz ~ 100 MHz	$1.4 dB$	
		20 Hz ~ 1 kHz		
Output amplitude		(0 ~ -60) dB	$0.007 dB$	
		1 kHz ~ 20 kHz		
		(0 ~ -60) dB	$0.009 dB$	
		20 kHz ~ 100 kHz		
		(0 ~ -60) dB	$0.015 dB$	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Rise/Fall Time	40411	1 ns (1 ~ 10) ns (10 ~ 100) ns 100 ns ~ 1 s	5.9 ps $1.3 \times 10^{-3}$ $1.2 \times 10^{-3}$ $1.2 \times 10^{-3}$	
AM Modulation		5 % ~ 99 %	$1.2 \times 10^{-2}$	
FM Modulation		9 kHz ~ 400 kHz	$1.2 \times 10^{-2}$	
Duty Cycle		1 % ~ 99 %	$5.8 \times 10^{-3}$	
Genescopes Vertical gain	40412	1 kHz 100 mV 100 mV ~ 100 V	1.2 mV $1.2 \times 10^{-2}$	Oscilloscope calibrators / HCT-CS-110-40412
AC/DC high voltage voltmeters DC Voltage	40413	(+) 1 V (1 ~ 100) V (0.1 ~ 1) kV  (1 ~ 2) kV (2 ~ 5) kV (5 ~ 10) kV (10 ~ 20) kV (20 ~ 50) kV (50 ~ 100) kV  (-) -1 V -(1 ~ 100) V -(0.1 ~ 1) kV -(1 ~ 2) kV -(2 ~ 5) kV -(5 ~ 10) kV -(10 ~ 20) kV -(20 ~ 50) kV -(50 ~ 100) kV	1 mV $5.8 \times 10^{-4}$ $4.0 \times 10^{-5}$  $1.4 \times 10^{-3}$ $7.9 \times 10^{-4}$ $4.8 \times 10^{-4}$ $4.2 \times 10^{-4}$ $4.4 \times 10^{-4}$ $4.1 \times 10^{-4}$  1 mV $5.8 \times 10^{-4}$ $4.0 \times 10^{-5}$ $1.4 \times 10^{-3}$ $7.9 \times 10^{-4}$ $4.8 \times 10^{-4}$ $4.2 \times 10^{-4}$ $4.4 \times 10^{-4}$ $4.1 \times 10^{-4}$	High voltage generators / HCT-CS-092-40413
AC Voltage		(50 ~ 60) Hz 1 V (1 ~ 200) V (200 ~ 500) V (0.5 ~ 1) kV (1 ~ 20) kV (20 ~ 30) kV (30 ~ 40) kV (40 ~ 60) kV (60 ~ 70) kV	7 mV $6.2 \times 10^{-3}$ $8.6 \times 10^{-5}$ $1.1 \times 10^{-4}$ $2.8 \times 10^{-2}$ $2.3 \times 10^{-2}$ $2.1 \times 10^{-2}$ $2.0 \times 10^{-2}$ $1.9 \times 10^{-2}$	
Leakage current testers AC Current	40416	(10 Hz) 100 $\mu$ A (100 ~ 200) $\mu$ A (200 ~ 500) $\mu$ A (0.5 ~ 1) mA (1 ~ 2) mA (2 ~ 5) mA (5 ~ 10) mA (10 ~ 20) mA (20 ~ 50) mA (50 ~ 100) mA	78 nA $5.0 \times 10^{-4}$ $4.6 \times 10^{-4}$ $7.1 \times 10^{-4}$ $4.4 \times 10^{-4}$ $4.6 \times 10^{-4}$ $7.1 \times 10^{-4}$ $4.4 \times 10^{-4}$ $4.8 \times 10^{-4}$ $3.7 \times 10^{-4}$	Meter calibrators /HCT-CS-208-40416

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC Current	40416	(10 ~ 40) Hz		
		100 μA	69 nA	
		(100 ~ 200) μA	$4.0 \times 10^{-4}$	
		(200 ~ 500) μA	$3.2 \times 10^{-4}$	
		(0.5 ~ 1) mA	$6.6 \times 10^{-4}$	
		(1 ~ 2) mA	$3.7 \times 10^{-4}$	
		(2 ~ 5) mA	$3.2 \times 10^{-4}$	
		(5 ~ 10) mA	$6.6 \times 10^{-4}$	
		(10 ~ 20) mA	$3.7 \times 10^{-4}$	
		(20 ~ 50) mA	$3.2 \times 10^{-4}$	
		(50 ~ 100) mA	$2.5 \times 10^{-4}$	
		(0.04 ~ 1) kHz		
		20 μA	14 nA	
		(20 ~ 50) μA	$3.6 \times 10^{-4}$	
		(50 ~ 100) μA	$6.6 \times 10^{-4}$	
		(100 ~ 200) μA	$3.5 \times 10^{-4}$	
		(200 ~ 500) μA	$2.8 \times 10^{-4}$	
		(0.5 ~ 1) mA	$6.4 \times 10^{-4}$	
		(1 ~ 2) mA	$3.4 \times 10^{-4}$	
		(2 ~ 5) mA	$2.8 \times 10^{-4}$	
		(5 ~ 10) mA	$6.4 \times 10^{-4}$	
		(10 ~ 20) mA	$3.4 \times 10^{-4}$	
		(20 ~ 50) mA	$2.6 \times 10^{-4}$	
		(50 ~ 100) mA	$1.8 \times 10^{-4}$	
		(1 ~ 10) kHz		
		20 μA	0.11 μA	
		(20 ~ 50) μA	$3.0 \times 10^{-3}$	
		(50 ~ 100) μA	$2.2 \times 10^{-3}$	
		(100 ~ 200) μA	$1.7 \times 10^{-3}$	
		(200 ~ 500) μA	$3.0 \times 10^{-3}$	
		(0.5 ~ 1) mA	$2.2 \times 10^{-3}$	
		(1 ~ 2) mA	$1.7 \times 10^{-3}$	
		(2 ~ 5) mA	$2.8 \times 10^{-3}$	
(5 ~ 10) mA	$2.1 \times 10^{-3}$			
(10 ~ 20) mA	$1.7 \times 10^{-3}$			
(20 ~ 50) mA	$1.9 \times 10^{-3}$			
(50 ~ 100) mA	$1.6 \times 10^{-3}$			
DC Current	40416	1 μA	7.1 nA	
		(1 ~ 2) μA	$3.6 \times 10^{-3}$	
		(2 ~ 5) μA	$1.4 \times 10^{-3}$	
		(5 ~ 10) μA	$1.0 \times 10^{-3}$	
		(10 ~ 20) μA	$5.5 \times 10^{-4}$	
		(20 ~ 50) μA	$2.4 \times 10^{-4}$	
		(50 ~ 100) μA	$6.3 \times 10^{-4}$	
		(100 ~ 200) μA	$3.2 \times 10^{-4}$	
		(200 ~ 500) μA	$1.4 \times 10^{-4}$	
		(0.5 ~ 1) mA	$6.2 \times 10^{-4}$	
		(1 ~ 2) mA	$3.1 \times 10^{-4}$	
		(2 ~ 5) mA	$1.4 \times 10^{-4}$	
		(5 ~ 10) mA	$6.2 \times 10^{-4}$	
		(10 ~ 20) mA	$3.1 \times 10^{-4}$	
		(20 ~ 50) mA	$1.4 \times 10^{-4}$	
(50 ~ 100) mA	$8.6 \times 10^{-5}$			
AC Voltage	40416	(40 Hz)		
		1 mV	4.8 μV	
		(1 ~ 2) mV	$2.5 \times 10^{-3}$	
		(2 ~ 5) mV	$1.1 \times 10^{-3}$	



404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC Voltage	40416	(5 ~ 10) mV	$5.9 \times 10^{-4}$	
		(10 ~ 20) mV	$4.6 \times 10^{-4}$	
		(20 ~ 50) mV	$3.2 \times 10^{-4}$	
		(50 ~ 100) mV	$2.0 \times 10^{-4}$	
		(100 ~ 200) mV	$3.4 \times 10^{-4}$	
		(200 ~ 500) mV	$1.9 \times 10^{-4}$	
		(0.5 ~ 1) V	$6.3 \times 10^{-4}$	
		(1 ~ 2) V	$3.3 \times 10^{-4}$	
		(2 ~ 5) V	$2.0 \times 10^{-4}$	
		(5 ~ 10) V	$1.4 \times 10^{-4}$	
		(10 ~ 20) V	$3.3 \times 10^{-4}$	
		(20 ~ 50) V	$2.2 \times 10^{-4}$	
		(50 ~ 100) V	$1.5 \times 10^{-4}$	
		(0.04 ~ 1) kHz		
		1 mV	4.8 $\mu$ V	
		(1 ~ 2) mV	$2.5 \times 10^{-3}$	
		(2 ~ 5) mV	$1.1 \times 10^{-3}$	
		(5 ~ 10) mV	$5.8 \times 10^{-4}$	
		(10 ~ 20) mV	$4.5 \times 10^{-4}$	
		(20 ~ 50) mV	$2.8 \times 10^{-4}$	
		(50 ~ 100) mV	$1.7 \times 10^{-4}$	
		(100 ~ 200) mV	$3.3 \times 10^{-4}$	
		(200 ~ 500) mV	$1.4 \times 10^{-4}$	
		(0.5 ~ 1) V	$6.2 \times 10^{-4}$	
		(1 ~ 2) V	$3.2 \times 10^{-4}$	
		(2 ~ 5) V	$1.4 \times 10^{-4}$	
		(5 ~ 10) V	$8.5 \times 10^{-5}$	
		(10 ~ 20) V	$3.1 \times 10^{-4}$	
		(20 ~ 50) V	$1.5 \times 10^{-4}$	
		(50 ~ 100) V	$9.4 \times 10^{-5}$	
		(100 ~ 200) V	$7.5 \times 10^{-5}$	
		(200 ~ 500) V	$9.2 \times 10^{-5}$	
		(500 ~ 1 000) V	$8.7 \times 10^{-5}$	
		(1 ~ 10) kHz		
		1 mV	4.8 $\mu$ V	
		(1 ~ 2) mV	$2.5 \times 10^{-3}$	
		(2 ~ 5) mV	$1.1 \times 10^{-3}$	
		(5 ~ 10) mV	$5.8 \times 10^{-4}$	
		(10 ~ 20) mV	$4.5 \times 10^{-4}$	
		(20 ~ 50) mV	$2.8 \times 10^{-4}$	
		(50 ~ 100) mV	$1.7 \times 10^{-4}$	
		(100 ~ 200) mV	$3.3 \times 10^{-4}$	
		(200 ~ 500) mV	$1.4 \times 10^{-4}$	
		(0.5 ~ 1) V	$6.2 \times 10^{-4}$	
(1 ~ 2) V	$3.2 \times 10^{-4}$			
(2 ~ 5) V	$1.4 \times 10^{-4}$			
(5 ~ 10) V	$8.5 \times 10^{-5}$			
(10 ~ 20) V	$3.1 \times 10^{-4}$			
(20 ~ 50) V	$1.5 \times 10^{-4}$			
(50 ~ 100) V	$9.4 \times 10^{-5}$			
DC Voltage	40416	1 mV	0.80 $\mu$ V	
		(1 ~ 2) mV	$4.1 \times 10^{-4}$	
		(2 ~ 5) mV	$1.7 \times 10^{-4}$	
		(5 ~ 10) mV	$8.5 \times 10^{-5}$	
		(10 ~ 20) mV	$3.1 \times 10^{-4}$	
		(20 ~ 50) mV	$1.2 \times 10^{-4}$	
		(50 ~ 100) mV	$6.3 \times 10^{-5}$	
(100 ~ 200) mV	$3.1 \times 10^{-4}$			

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC Voltage	40416	(200 ~ 500) mV (0.5 ~ 1) V (1 ~ 2) V (2 ~ 5) V (5 ~ 10) V (10 ~ 20) V (20 ~ 50) V (50 ~ 100) V (100 ~ 200) V (200 ~ 500) V (500 ~ 1 000) V	$1.2 \times 10^{-4}$ $6.2 \times 10^{-4}$ $3.1 \times 10^{-4}$ $1.2 \times 10^{-4}$ $6.2 \times 10^{-5}$ $3.1 \times 10^{-4}$ $1.2 \times 10^{-4}$ $6.2 \times 10^{-5}$ $3.2 \times 10^{-5}$ $1.5 \times 10^{-5}$ $1.1 \times 10^{-5}$	
Input voltage to output current ratio		(20 Hz ~ 1 MHz) 1 ~ 1 384	$4.2 \times 10^{-3}$	
Input voltage to output voltage ratio		(20 Hz ~ 1 MHz) 0.5 ~ 689	$4.8 \times 10^{-3}$	
Resistance		10 Ω (10 ~ 20) Ω (20 ~ 50) Ω (50 ~ 100) Ω (100 ~ 200) Ω (200 ~ 500) Ω (0.5 ~ 1) kΩ (1 ~ 2) kΩ (2 ~ 5) kΩ (5 ~ 10) kΩ (10 ~ 20) kΩ (20 ~ 50) kΩ (50 ~ 100) kΩ (100 ~ 200) kΩ (200 ~ 500) kΩ (0.5 ~ 1) MΩ	0.12 mΩ $3.5 \times 10^{-5}$ $1.6 \times 10^{-5}$ $1.1 \times 10^{-5}$ $3.2 \times 10^{-5}$ $1.5 \times 10^{-5}$ $1.2 \times 10^{-5}$ $3.3 \times 10^{-5}$ $1.6 \times 10^{-5}$ $1.1 \times 10^{-5}$ $3.3 \times 10^{-5}$ $1.6 \times 10^{-5}$ $1.1 \times 10^{-5}$ $3.5 \times 10^{-5}$ $1.6 \times 10^{-5}$ $1.3 \times 10^{-5}$	
Capacitance		(1 kHz) 100 pF (100 ~ 200) pF (200 ~ 500) pF (0.5 ~ 1) nF (1 ~ 2) nF (2 ~ 5) nF (5 ~ 10) nF (10 ~ 20) nF (20 ~ 50) nF (50 ~ 100) nF (100 ~ 200) nF (200 ~ 500) nF (0.5 ~ 1) μF	1.2 fF $5.1 \times 10^{-5}$ $2.1 \times 10^{-5}$ $1.2 \times 10^{-5}$ $1.1 \times 10^{-4}$ $4.2 \times 10^{-5}$ $2.1 \times 10^{-5}$ $2.6 \times 10^{-4}$ $1.1 \times 10^{-4}$ $5.1 \times 10^{-5}$ $5.1 \times 10^{-4}$ $2.1 \times 10^{-4}$ $1.1 \times 10^{-4}$	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electronic AC/DC loads	40417			DC power supplies, Current shunts, Digital multimeters /HCT-CS-094-40417
DC loads CV Mode		100 mV (0.1 ~ 1 000) V	6.4 μV $1.2 \times 10^{-4}$	
CC Mode		100 mA (0.1 ~ 1) A (1 ~ 10) A (10 ~ 100) A (100 ~ 300) A (300 ~ 400) A	6.4 μA $6.4 \times 10^{-5}$ $1.2 \times 10^{-4}$ $1.2 \times 10^{-4}$ $2.3 \times 10^{-4}$ $2.1 \times 10^{-4}$	
CR Mode		0.1 Ω 100 A (0.1 ~ 1) Ω (100 ~ 10) A (1 ~ 100) Ω (10 ~ 0.1) A	7.3 mA $6.8 \times 10^{-5}$ $6.4 \times 10^{-5}$	
AC loads CV Mode		60 Hz 100 mV (0.1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V	16 μV $1.5 \times 10^{-4}$ $1.5 \times 10^{-4}$ $1.5 \times 10^{-4}$	
CC Mode		60 Hz 100 mA (0.1 ~ 1) A (1 ~ 10) A (10 ~ 20) A (20 ~ 40) A	92 μA $1.2 \times 10^{-3}$ $2.3 \times 10^{-3}$ $6.4 \times 10^{-4}$ $6.7 \times 10^{-4}$	
Charging/Discharging Tester		100 μA	5.8 nA	
Charging Current		(0.000 1 ~ 100) A (100 ~ 400) A (400 ~ 500) A (500 ~ 1 000) A	$1.2 \times 10^{-4}$ $2.1 \times 10^{-4}$ $2.0 \times 10^{-4}$ $3.1 \times 10^{-4}$	
Discharging Current		-100 μA (-0.000 1 ~ -100) A (-100 ~ -400) A (-400 ~ -500) A (-500 ~ -1 000) A	5.8 nA $1.2 \times 10^{-4}$ $2.1 \times 10^{-4}$ $2.0 \times 10^{-4}$ $3.1 \times 10^{-4}$	
Charging Voltage		100 mV (0.1 ~ 1 000) V (1 000 ~ 1 500) V	6.2 μV $1.2 \times 10^{-4}$ $1.1 \times 10^{-3}$	
Sense Voltage(Meter)		100 mV (0.1 ~ 1 000) V	6.4 μV $1.2 \times 10^{-4}$	
Analogue/digital multimeters	40419			Meter callibraters Current amplifiers, Standard resistance /HCT-CS-095-40419
DC Voltage		0 mV (0 ~ 100) mV (0 ~ -100) mV (0.1 ~ 1) V (-0.1 ~ -1) V (1 ~ 10) V	0.20 μV $4.1 \times 10^{-6}$ $4.1 \times 10^{-6}$ $1.5 \times 10^{-6}$ $1.5 \times 10^{-6}$ $8.2 \times 10^{-7}$	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC Voltage	40419	(-1 ~ -10) V	$8.2 \times 10^{-7}$	
		(10 ~ 100) V	$1.6 \times 10^{-6}$	
		(-10 ~ -100) V	$1.6 \times 10^{-6}$	
		(100 ~ 1 000) V	$2.1 \times 10^{-6}$	
		(-100 ~ -1 000) V	$2.1 \times 10^{-6}$	
DC Current	40419	100 nA	2.7 pA	
		-100 nA	2.7 pA	
		0 $\mu$ A	0.39 nA	
		(0 ~ 1) $\mu$ A	$3.8 \times 10^{-4}$	
		(0 ~ -1) $\mu$ A	$3.8 \times 10^{-4}$	
		(1 ~ 10) $\mu$ A	$3.8 \times 10^{-5}$	
		(-1 ~ -10) $\mu$ A	$3.8 \times 10^{-5}$	
		(10 ~ 100) $\mu$ A	$1.9 \times 10^{-5}$	
		(-10 ~ -100) $\mu$ A	$1.9 \times 10^{-5}$	
		(0.1 ~ 1) mA	$2.4 \times 10^{-5}$	
		(-0.1 ~ -1) mA	$2.4 \times 10^{-5}$	
		(1 ~ 10) mA	$2.5 \times 10^{-5}$	
		(-1 ~ -10) mA	$2.5 \times 10^{-5}$	
		(10 ~ 100) mA	$1.9 \times 10^{-5}$	
		(-10 ~ -100) mA	$1.9 \times 10^{-5}$	
		(0.1 ~ 1) A	$4.7 \times 10^{-5}$	
		(-0.1 ~ -1) A	$4.7 \times 10^{-5}$	
		(1 ~ 10) A	$8.7 \times 10^{-5}$	
(-1 ~ -10) A	$8.7 \times 10^{-5}$			
(10 ~ 20) A	$8.7 \times 10^{-5}$			
(-10 ~ -20) A	$8.7 \times 10^{-5}$			
Resistance	40419	0 $\Omega$	2.1 $\mu\Omega$	
		(0 ~ 100) $\Omega$	$2.2 \times 10^{-6}$	
		(0.1 ~ 1) k $\Omega$	$9.9 \times 10^{-7}$	
		(1 ~ 10) k $\Omega$	$1.2 \times 10^{-6}$	
		(10 ~ 100) k $\Omega$	$9.2 \times 10^{-7}$	
		(0.1 ~ 1) M $\Omega$	$1.5 \times 10^{-6}$	
		(1 ~ 10) M $\Omega$	$6.2 \times 10^{-6}$	
		(10 ~ 100) M $\Omega$	$3.1 \times 10^{-5}$	
		(0.1 ~ 1) G $\Omega$	$1.2 \times 10^{-5}$	
(1 ~ 10) G $\Omega$	$5.6 \times 10^{-4}$			
AC Voltage	40419	1 mV		
		10 Hz	0.64 $\mu$ V	
		(10 ~ 40) Hz	0.62 $\mu$ V	
		(40 ~ 500) Hz	0.63 $\mu$ V	
		(0.5 ~ 1) kHz	0.63 $\mu$ V	
		(1 ~ 10) kHz	0.63 $\mu$ V	
		(10 ~ 20) kHz	0.61 $\mu$ V	
		(20 ~ 50) kHz	0.91 $\mu$ V	
		(50 ~ 100) kHz	1.5 $\mu$ V	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC Voltage	40419	(100 ~ 200) kHz	1.3 μV	
		(200 ~ 500) kHz	1.8 μV	
		(0.5 ~ 1) MHz	6.4 μV	
		(1 ~ 100) mV		
		10 Hz	$8.2 \times 10^{-5}$	
		(10 ~ 40) Hz	$4.7 \times 10^{-5}$	
		(40 ~ 500) Hz	$4.5 \times 10^{-5}$	
		(0.5 ~ 1) kHz	$4.5 \times 10^{-5}$	
		(1 ~ 10) kHz	$4.5 \times 10^{-5}$	
		(10 ~ 20) kHz	$4.5 \times 10^{-5}$	
		(20 ~ 50) kHz	$6.7 \times 10^{-5}$	
		(50 ~ 100) kHz	$1.0 \times 10^{-4}$	
		(100 ~ 200) kHz	$2.0 \times 10^{-4}$	
		(200 ~ 500) kHz	$3.1 \times 10^{-4}$	
		(0.5 ~ 1) MHz	$6.5 \times 10^{-4}$	
		(0.1 ~ 1) V		
		10 Hz	$6.9 \times 10^{-5}$	
		(10 ~ 40) Hz	$3.8 \times 10^{-5}$	
		(40 ~ 500) Hz	$2.3 \times 10^{-5}$	
		(0.5 ~ 1) kHz	$2.3 \times 10^{-5}$	
		(1 ~ 10) kHz	$2.8 \times 10^{-5}$	
		(10 ~ 20) kHz	$2.8 \times 10^{-5}$	
		(20 ~ 50) kHz	$3.8 \times 10^{-5}$	
		(50 ~ 100) kHz	$4.8 \times 10^{-5}$	
		(100 ~ 200) kHz	$1.3 \times 10^{-4}$	
		(200 ~ 500) kHz	$3.0 \times 10^{-4}$	
		(0.5 ~ 1) MHz	$1.2 \times 10^{-3}$	
		(1 ~ 10) V		
		10 Hz	$6.8 \times 10^{-5}$	
		(10 ~ 40) Hz	$3.3 \times 10^{-5}$	
		(40 ~ 500) Hz	$2.2 \times 10^{-5}$	
		(0.5 ~ 1) kHz	$2.2 \times 10^{-5}$	
		(1 ~ 10) kHz	$2.2 \times 10^{-5}$	
		(10 ~ 20) kHz	$2.2 \times 10^{-5}$	
		(20 ~ 50) kHz	$3.1 \times 10^{-5}$	
		(50 ~ 100) kHz	$4.6 \times 10^{-5}$	
		(100 ~ 200) kHz	$1.7 \times 10^{-4}$	
		(200 ~ 500) kHz	$4.0 \times 10^{-4}$	
		(0.5 ~ 1) MHz	$1.5 \times 10^{-4}$	
		(10 ~ 100) V		
		10 Hz	$8.5 \times 10^{-5}$	
		(10 ~ 40) Hz	$4.1 \times 10^{-5}$	
		(40 ~ 500) Hz	$2.6 \times 10^{-5}$	
		(0.5 ~ 1) kHz	$2.6 \times 10^{-5}$	
		(1 ~ 10) kHz	$2.6 \times 10^{-5}$	
(10 ~ 20) kHz	$2.6 \times 10^{-5}$			
(20 ~ 50) kHz	$3.4 \times 10^{-5}$			
(50 ~ 100) kHz	$7.8 \times 10^{-5}$			
(100 ~ 1 000) V				
40 Hz	$2.6 \times 10^{-5}$			
(40 ~ 500) Hz	$2.0 \times 10^{-5}$			
(0.5 ~ 1) kHz	$2.0 \times 10^{-5}$			
(1 ~ 10) kHz	$8.9 \times 10^{-5}$			
(10 ~ 20) kHz	$1.3 \times 10^{-4}$			
(20 ~ 30) kHz	$1.9 \times 10^{-4}$			

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC Current	40419	10 $\mu$ A		
		10 Hz	10 nA	
		(10 ~ 40) Hz	7.3 nA	
		(40 ~ 500) Hz	1.4 nA	
		(0.5 ~ 1) kHz	3.8 nA	
		(1 ~ 5) kHz	6.7 nA	
		(5 ~ 10) kHz	9.7 nA	
		(10 ~ 100) $\mu$ A		
		10 Hz	$1.0 \times 10^{-4}$	
		(10 ~ 40) Hz	$7.3 \times 10^{-5}$	
		(40 ~ 500) Hz	$7.1 \times 10^{-5}$	
		(0.5 ~ 1) kHz	$7.1 \times 10^{-5}$	
		(1 ~ 5) kHz	$1.6 \times 10^{-4}$	
		(5 ~ 10) kHz	$4.1 \times 10^{-4}$	
		(0.1 ~ 1) mA		
		10 Hz	$1.1 \times 10^{-4}$	
		(10 ~ 40) Hz	$6.3 \times 10^{-5}$	
		(40 ~ 500) Hz	$6.2 \times 10^{-5}$	
		(0.5 ~ 1) kHz	$6.2 \times 10^{-5}$	
		(1 ~ 5) kHz	$1.2 \times 10^{-4}$	
		(5 ~ 10) kHz	$4.0 \times 10^{-4}$	
		(1 ~ 10) mA		
		10 Hz	$1.2 \times 10^{-4}$	
		(10 ~ 40) Hz	$6.6 \times 10^{-5}$	
		(40 ~ 500) Hz	$6.5 \times 10^{-5}$	
		(0.5 ~ 1) kHz	$6.5 \times 10^{-5}$	
		(1 ~ 5) kHz	$1.7 \times 10^{-4}$	
		(5 ~ 10) kHz	$6.1 \times 10^{-4}$	
		(10 ~ 100) mA		
		10 Hz	$1.3 \times 10^{-4}$	
		(10 ~ 40) Hz	$6.8 \times 10^{-5}$	
		(40 ~ 500) Hz	$6.6 \times 10^{-5}$	
		(0.5 ~ 1) kHz	$6.6 \times 10^{-5}$	
		(1 ~ 5) kHz	$1.7 \times 10^{-4}$	
		(5 ~ 10) kHz	$6.1 \times 10^{-4}$	
		(0.1 ~ 1) A		
		10 Hz	$1.3 \times 10^{-4}$	
		(10 ~ 40) Hz	$1.3 \times 10^{-4}$	
		(40 ~ 500) Hz	$1.1 \times 10^{-4}$	
		(0.5 ~ 1) kHz	$1.1 \times 10^{-4}$	
		(1 ~ 5) kHz	$3.4 \times 10^{-4}$	
		(5 ~ 10) kHz	$1.3 \times 10^{-3}$	
		(1 ~ 10) A		
		40 Hz	$7.2 \times 10^{-5}$	
		(40 ~ 500) Hz	$8.1 \times 10^{-5}$	
		(0.5 ~ 1) kHz	$6.5 \times 10^{-5}$	
		(1 ~ 5) kHz	$1.2 \times 10^{-4}$	
		(5 ~ 10) kHz	$2.2 \times 10^{-3}$	
(10 ~ 20) A				
40 Hz	$6.7 \times 10^{-5}$			
(40 ~ 500) Hz	$7.1 \times 10^{-5}$			
(0.5 ~ 1) kHz	$7.2 \times 10^{-5}$			
(1 ~ 5) kHz	$1.8 \times 10^{-4}$			
(5 ~ 10) kHz	$2.9 \times 10^{-4}$			

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Frequency	40419	10 Hz (10 ~ 100) Hz (0.1 ~ 1) kHz (1 ~ 10) kHz (10 ~ 100) kHz (0.1 ~ 1) MHz (1 ~ 10) MHz	0.24 mHz $2.5 \times 10^{-5}$ $2.3 \times 10^{-5}$ $1.0 \times 10^{-4}$ $2.5 \times 10^{-5}$ $2.3 \times 10^{-5}$ $2.3 \times 10^{-5}$	
Noise meters AC level(rms & Q-peak)	40420	100 mV 10 Hz 10 Hz ~ 50 kHz (50 ~ 100) kHz  (100 ~ 300) mV 10 Hz 10 Hz ~ 50 kHz (50 ~ 100) kHz  (0.3 ~ 1) V 10 Hz 10 Hz ~ 50 kHz (50 ~ 100) kHz  (1 ~ 3) V 10 Hz 10 Hz ~ 50 kHz (50 ~ 100) kHz  (3 ~ 10) V 10 Hz 10 Hz ~ 50 kHz (50 ~ 100) kHz  (10 ~ 30) V 10 Hz 10 Hz ~ 50 kHz (50 ~ 100) kHz  (30 ~ 100) V 10 Hz 10 Hz ~ 50 kHz (50 ~ 100) kHz  (100 ~ 300) V 10 Hz 10 Hz ~ 10 kHz	0.16 mV $1.6 \times 10^{-3}$ $1.8 \times 10^{-3}$  0.63 mV $2.1 \times 10^{-3}$ $2.2 \times 10^{-3}$  1.6 mV $1.7 \times 10^{-3}$ $1.8 \times 10^{-3}$  6.3 mV $2.1 \times 10^{-3}$ $2.1 \times 10^{-3}$  16 mV $1.5 \times 10^{-3}$ $1.6 \times 10^{-3}$  63 mV $2.1 \times 10^{-3}$ $2.6 \times 10^{-3}$  0.15 V $1.5 \times 10^{-3}$ $2.0 \times 10^{-3}$  0.63 V $2.1 \times 10^{-3}$	Multimeter calibrators /HCT-CS-097-40420
Weighting filter Filter(DIN/AUDIO, JIS A CCIR, CCIR/ARM)		1V	1.5 mV	
Frequency Reponse		(-10 ~10) dB 20 Hz ~ 100 kHz	0.016 dB	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Oscilloscopes  DC Voltage	40421	(±)		Oscilloscope calibrators, Multimeter calibrators, RF signal calibrators, Powermeters /HCT-CS-080-40421
		0 V	4.6 μV	
		(0 ~ 1) mV	$3.0 \times 10^{-2}$	
		(1 ~ 2) mV	$1.5 \times 10^{-2}$	
		(2 ~ 3) mV	$1.0 \times 10^{-2}$	
		(3 ~ 4) mV	$7.6 \times 10^{-3}$	
		(4 ~ 5) mV	$6.1 \times 10^{-3}$	
		(5 ~ 6) mV	$5.1 \times 10^{-3}$	
		(6 ~ 7) mV	$4.9 \times 10^{-3}$	
		(7 ~ 8) mV	$4.3 \times 10^{-3}$	
		(8 ~ 9) mV	$3.8 \times 10^{-3}$	
		(9 ~ 10) mV	$3.4 \times 10^{-3}$	
		(10 ~ 15) mV	$2.3 \times 10^{-3}$	
		(15 ~ 25) mV	$1.8 \times 10^{-3}$	
		(20 ~ 25) mV	$1.8 \times 10^{-3}$	
		(25 ~ 30) mV	$1.5 \times 10^{-3}$	
		(30 ~ 35) mV	$1.3 \times 10^{-3}$	
		(35 ~ 40) mV	$1.1 \times 10^{-3}$	
		(40 ~ 45) mV	$9.8 \times 10^{-4}$	
		(45 ~ 50) mV	$8.8 \times 10^{-4}$	
		(50 ~ 60) mV	$7.8 \times 10^{-4}$	
		(60 ~ 70) mV	$1.2 \times 10^{-3}$	
		(70 ~ 80) mV	$1.0 \times 10^{-3}$	
		(80 ~ 90) mV	$9.3 \times 10^{-4}$	
		(90 ~ 100) mV	$8.4 \times 10^{-4}$	
		(100 ~ 150) mV	$5.6 \times 10^{-4}$	
		(150 ~ 200) mV	$4.8 \times 10^{-4}$	
		(200 ~ 250) mV	$6.9 \times 10^{-4}$	
		(250 ~ 300) mV	$5.8 \times 10^{-4}$	
		(300 ~ 350) mV	$5.0 \times 10^{-4}$	
		(350 ~ 400) mV	$4.3 \times 10^{-4}$	
		(400 ~ 450) mV	$3.9 \times 10^{-4}$	
		(450 ~ 500) mV	$3.5 \times 10^{-4}$	
		(0.5 ~ 0.6) V	$3.7 \times 10^{-4}$	
		(0.6 ~ 0.7) V	$8.3 \times 10^{-4}$	
		(0.7 ~ 0.8) V	$7.3 \times 10^{-4}$	
		(0.8 ~ 0.9) V	$6.5 \times 10^{-4}$	
		(0.9 ~ 1) V	$5.8 \times 10^{-4}$	
		(1 ~ 2.5) V	$5.9 \times 10^{-4}$	
		(2.5 ~ 5) V	$2.9 \times 10^{-4}$	
(5 ~ 10) V	$6.7 \times 10^{-4}$			
(10 ~ 25) V	$5.8 \times 10^{-4}$			
(25 ~ 30) V	$4.8 \times 10^{-4}$			
(30 ~ 35) V	$4.1 \times 10^{-4}$			
(35 ~ 40) V	$3.6 \times 10^{-4}$			
(40 ~ 45) V	$3.2 \times 10^{-4}$			
(45 ~ 50) V	$2.9 \times 10^{-4}$			
(50 ~ 60) V	$3.0 \times 10^{-4}$			
(60 ~ 70) V	$4.2 \times 10^{-4}$			
(70 ~ 80) V	$3.7 \times 10^{-4}$			
(80 ~ 90) V	$3.3 \times 10^{-4}$			
(90 ~ 100) V	$3.0 \times 10^{-4}$			
(100 ~ 200) V	$2.9 \times 10^{-4}$			



404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Square wave Voltage	40421	1 kHz		
		1 mV	19 μV	
		(1 ~ 2) mV	$9.3 \times 10^{-3}$	
		(2 ~ 3) mV	$6.2 \times 10^{-3}$	
		(3 ~ 4) mV	$4.6 \times 10^{-3}$	
		(4 ~ 5) mV	$3.7 \times 10^{-3}$	
		(5 ~ 6) mV	$3.1 \times 10^{-3}$	
		(6 ~ 7) mV	$1.2 \times 10^{-2}$	
		(7 ~ 8) mV	$1.0 \times 10^{-2}$	
		(8 ~ 9) mV	$9.0 \times 10^{-3}$	
		(9 ~ 10) mV	$8.1 \times 10^{-3}$	
		(10 ~ 15) mV	$5.4 \times 10^{-3}$	
		(15 ~ 20) mV	$4.1 \times 10^{-3}$	
		(20 ~ 25) mV	$3.2 \times 10^{-3}$	
		(25 ~ 30) mV	$2.7 \times 10^{-3}$	
		(30 ~ 35) mV	$2.3 \times 10^{-3}$	
		(35 ~ 40) mV	$2.0 \times 10^{-3}$	
		(40 ~ 45) mV	$1.8 \times 10^{-3}$	
		(45 ~ 50) mV	$1.6 \times 10^{-3}$	
		(50 ~ 100) mV	$7.1 \times 10^{-3}$	
		(100 ~ 250) mV	$2.8 \times 10^{-3}$	
		(250 ~ 500) mV	$1.4 \times 10^{-3}$	
		(0.5 ~ 1) V	$7.0 \times 10^{-3}$	
		(1 ~ 2.5) V	$2.8 \times 10^{-3}$	
		(2.5 ~ 5) V	$1.4 \times 10^{-3}$	
		(5 ~ 10) V	$7.0 \times 10^{-3}$	
		(10 ~ 25) V	$2.8 \times 10^{-3}$	
		(25 ~ 50) V	$1.4 \times 10^{-3}$	
		(50 ~ 60) V	$1.2 \times 10^{-3}$	
		(60 ~ 70) V	$1.7 \times 10^{-3}$	
		(70 ~ 80) V	$1.5 \times 10^{-3}$	
		(80 ~ 90) V	$1.3 \times 10^{-3}$	
		(90 ~ 100) V	$1.2 \times 10^{-3}$	
(100 ~ 150) V	$1.6 \times 10^{-3}$			
(150 ~ 200) V	$1.2 \times 10^{-3}$			
Bandwidth level	40421	50 kHz ~ 1 MHz		
		60 mV	37 μV	
		(0.06 ~ 3) V	$2.2 \times 10^{-2}$	
		(1 ~ 550) MHz		
		60 mV	2.7 mV	
		(0.06 ~ 3) V	$2.7 \times 10^{-2}$	
		550 MHz ~ 40 GHz		
		60 mV	1.4 mV	
		(60 ~ 600) mV	$1.9 \times 10^{-2}$	
		(0.6 ~ 3) V	$2.5 \times 10^{-2}$	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Time mark	40421	1 ns	8.4 fs	
		(1 ~ 2) ns	$4.2 \times 10^{-6}$	
		(2 ~ 5) ns	$1.7 \times 10^{-6}$	
		(5 ~ 10) ns	$5.8 \times 10^{-6}$	
		(10 ~ 20) ns	$2.9 \times 10^{-6}$	
		(20 ~ 50) ns	$1.2 \times 10^{-6}$	
		(50 ~ 100) ns	$5.8 \times 10^{-6}$	
		(100 ~ 200) ns	$2.9 \times 10^{-6}$	
		(200 ~ 500) ns	$1.2 \times 10^{-6}$	
		(0.5 ~ 1) $\mu$ s	$5.8 \times 10^{-6}$	
		(1 ~ 2) $\mu$ s	$2.9 \times 10^{-6}$	
		(2 ~ 5) $\mu$ s	$1.2 \times 10^{-6}$	
		(5 ~ 10) $\mu$ s	$5.8 \times 10^{-6}$	
		(10 ~ 20) $\mu$ s	$2.9 \times 10^{-6}$	
		(20 ~ 50) $\mu$ s	$1.2 \times 10^{-6}$	
		(50 ~ 100) $\mu$ s	$5.8 \times 10^{-6}$	
		(100 ~ 200) $\mu$ s	$2.9 \times 10^{-6}$	
		(200 ~ 500) $\mu$ s	$1.2 \times 10^{-6}$	
		(0.5 ~ 1) ms	$5.8 \times 10^{-6}$	
		(1 ~ 2) ms	$2.9 \times 10^{-6}$	
		(2 ~ 5) ms	$1.2 \times 10^{-6}$	
		(5 ~ 10) ms	$5.8 \times 10^{-6}$	
		(10 ~ 20) ms	$2.9 \times 10^{-6}$	
		(20 ~ 50) ms	$1.2 \times 10^{-6}$	
		(50 ~ 100) ms	$5.8 \times 10^{-6}$	
		(100 ~ 200) ms	$2.9 \times 10^{-6}$	
		(200 ~ 500) ms	$1.2 \times 10^{-6}$	
		(0.5 ~ 1) s	$5.8 \times 10^{-6}$	
		(1 ~ 2) s	$2.9 \times 10^{-6}$	
		(2 ~ 5) s	$1.2 \times 10^{-6}$	
		(5 ~ 10) s	$5.8 \times 10^{-6}$	
		(10 ~ 20) s	$2.9 \times 10^{-6}$	
Frequency		100 mHz	0.84 $\mu$ Hz	
		(100 ~ 200) mHz	$4.2 \times 10^{-6}$	
		(200 ~ 500) mHz	$1.7 \times 10^{-6}$	
		(0.5 ~ 1) Hz	$5.8 \times 10^{-6}$	
		(1 ~ 2) Hz	$2.9 \times 10^{-6}$	
		(2 ~ 5) Hz	$1.2 \times 10^{-6}$	
		(5 ~ 10) Hz	$5.8 \times 10^{-6}$	
		(10 ~ 20) Hz	$2.9 \times 10^{-6}$	
		(20 ~ 50) Hz	$1.2 \times 10^{-6}$	
		(50 ~ 100) Hz	$5.8 \times 10^{-6}$	
		(100 ~ 200) Hz	$2.9 \times 10^{-6}$	
		(200 ~ 500) Hz	$1.2 \times 10^{-6}$	
		(0.5 ~ 1) kHz	$5.8 \times 10^{-6}$	
		(1 ~ 2) kHz	$2.9 \times 10^{-6}$	
		(2 ~ 5) kHz	$1.2 \times 10^{-6}$	
		(5 ~ 10) kHz	$5.8 \times 10^{-6}$	
		(10 ~ 20) kHz	$2.9 \times 10^{-6}$	
		(20 ~ 50) kHz	$1.2 \times 10^{-6}$	
		(50 ~ 100) kHz	$5.8 \times 10^{-6}$	
		(100 ~ 200) kHz	$2.9 \times 10^{-6}$	
		(200 ~ 500) kHz	$1.2 \times 10^{-6}$	
		(0.5 ~ 1) MHz	$5.8 \times 10^{-6}$	
		(1 ~ 2) MHz	$2.9 \times 10^{-6}$	
		(2 ~ 5) MHz	$1.2 \times 10^{-6}$	
(5 ~ 10) MHz	$5.8 \times 10^{-6}$			
(10 ~ 20) MHz	$2.9 \times 10^{-6}$			
(20 ~ 50) MHz	$1.2 \times 10^{-6}$			
(50 ~ 100) MHz	$5.8 \times 10^{-6}$			
(100 ~ 200) MHz	$2.9 \times 10^{-6}$			

404. 기타 직류 및 저주파측정

측정량/장비	분류번호	교정범위	측정불확도 (신뢰수준 약 95%)	사용표준/측정방법 등
Frequency	40421	(200 ~ 500) MHz	$1.2 \times 10^{-6}$	
		(0.5 ~ 1) GHz	$5.8 \times 10^{-6}$	
AC Voltage		50 Hz ~ 10 kHz		
		10 mV	5.8 $\mu$ V	
		(10 ~ 15) mV	$4.1 \times 10^{-4}$	
		(15 ~ 20) mV	$3.3 \times 10^{-4}$	
		(20 ~ 25) mV	$4.4 \times 10^{-4}$	
		(25 ~ 30) mV	$3.7 \times 10^{-4}$	
		(30 ~ 35) mV	$3.4 \times 10^{-4}$	
		(35 ~ 40) mV	$3.0 \times 10^{-4}$	
		(40 ~ 45) mV	$2.7 \times 10^{-4}$	
		(45 ~ 50) mV	$2.4 \times 10^{-4}$	
		(50 ~ 60) mV	$2.4 \times 10^{-4}$	
		(60 ~ 70) mV	$2.1 \times 10^{-4}$	
		(70 ~ 80) mV	$2.0 \times 10^{-4}$	
		(80 ~ 90) mV	$1.8 \times 10^{-4}$	
		(90 ~ 100) mV	$1.6 \times 10^{-4}$	
		(100 ~ 150) mV	$1.3 \times 10^{-4}$	
		(150 ~ 200) mV	$1.1 \times 10^{-4}$	
		(200 ~ 250) mV	$1.9 \times 10^{-4}$	
		(250 ~ 300) mV	$1.8 \times 10^{-4}$	
		(300 ~ 350) mV	$1.6 \times 10^{-4}$	
		(350 ~ 400) mV	$1.6 \times 10^{-4}$	
		(400 ~ 450) mV	$1.5 \times 10^{-4}$	
		(450 ~ 500) mV	$1.4 \times 10^{-4}$	
		(500 ~ 600) mV	$1.7 \times 10^{-4}$	
		(600 ~ 700) mV	$1.6 \times 10^{-4}$	
		(700 ~ 800) mV	$1.5 \times 10^{-4}$	
		(800 ~ 900) mV	$1.4 \times 10^{-4}$	
		(0.9 ~ 1) V	$1.4 \times 10^{-4}$	
		(1 ~ 1.5) V	$1.2 \times 10^{-4}$	
		(1.5 ~ 2) V	$1.2 \times 10^{-4}$	
		(2 ~ 2.5) V	$1.1 \times 10^{-4}$	
		(2.5 ~ 3) V	$9.7 \times 10^{-5}$	
		(3 ~ 3.5) V	$8.9 \times 10^{-5}$	
	(3.5 ~ 4) V	$8.2 \times 10^{-5}$		
	(4 ~ 4.5) V	$7.7 \times 10^{-5}$		
	(4.5 ~ 5) V	$7.3 \times 10^{-5}$		
	(5 ~ 6) V	$1.2 \times 10^{-4}$		
	(6 ~ 7) V	$1.0 \times 10^{-4}$		
	(7 ~ 8) V	$9.5 \times 10^{-5}$		
	(8 ~ 9) V	$9.3 \times 10^{-5}$		
	(9 ~ 10) V	$8.2 \times 10^{-5}$		
	(10 ~ 15) V	$6.9 \times 10^{-5}$		
	(15 ~ 20) V	$6.2 \times 10^{-5}$		
	(20 ~ 25) V	$1.3 \times 10^{-4}$		
	(25 ~ 30) V	$1.2 \times 10^{-4}$		
	(30 ~ 35) V	$1.1 \times 10^{-4}$		
	(35 ~ 40) V	$9.9 \times 10^{-5}$		
	(40 ~ 45) V	$9.3 \times 10^{-5}$		
	(45 ~ 50) V	$8.9 \times 10^{-5}$		
	(50 ~ 60) V	$1.3 \times 10^{-4}$		
	(60 ~ 70) V	$1.2 \times 10^{-4}$		
	(70 ~ 80) V	$1.1 \times 10^{-4}$		
	(80 ~ 90) V	$1.0 \times 10^{-4}$		
	(90 ~ 100) V	$9.2 \times 10^{-5}$		
Input Resistance		50 $\Omega$	5.8 m $\Omega$	
		75 $\Omega$	5.9 m $\Omega$	
		1 M $\Omega$	0.34 k $\Omega$	
10 MHz Reference out		10 MHz	$5.8 \times 10^{-8}$	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Output Voltage	40421	DC 100 mV (0.1 ~ 1) V (1 ~ 2) V (2 ~ 3) V (3 ~ 4) V (4 ~ 5) V (5 ~ 6) V (6 ~ 7) V (7 ~ 8) V (8 ~ 9) V (9 ~ 10) V (10 ~ 11) V (11 ~ 12) V  1 kHz 100 mV (0.1 ~ 1) V (1 ~ 2) V (2 ~ 3) V (3 ~ 4) V (4 ~ 5) V (5 ~ 6) V (6 ~ 7) V (7 ~ 8) V (8 ~ 9) V (9 ~ 10) V (10 ~ 11) V (11 ~ 12) V	61 μV $6.1 \times 10^{-5}$ $3.6 \times 10^{-5}$ $2.4 \times 10^{-5}$ $1.8 \times 10^{-5}$ $1.4 \times 10^{-5}$ $1.2 \times 10^{-5}$ $1.0 \times 10^{-5}$ $8.9 \times 10^{-6}$ $7.9 \times 10^{-6}$ $7.1 \times 10^{-6}$ $4.8 \times 10^{-5}$ $4.4 \times 10^{-5}$  63 μV $1.1 \times 10^{-4}$ $4.6 \times 10^{-4}$ $3.1 \times 10^{-4}$ $2.3 \times 10^{-4}$ $1.9 \times 10^{-4}$ $1.5 \times 10^{-4}$ $1.3 \times 10^{-4}$ $1.2 \times 10^{-4}$ $1.0 \times 10^{-4}$ $9.3 \times 10^{-5}$ $8.6 \times 10^{-4}$ $7.9 \times 10^{-4}$	
LF phase meters	40422	Synchro / Resolver 0 ° (0 ~ 360) °	0.002 ° 0.002 °	RESOLVER/SYNCHRO SIMULATOR / HCT-CS-217-40422
Random wave generators Frequency Output level	40423	1 Hz ~ 350 MHz  10 mV 20 Hz 20 Hz ~ 20 kHz (20 ~ 100) kHz 100 kHz ~ 1 MHz  (10 ~ 100) mV 20 Hz 20 Hz ~ 20 kHz (20 ~ 100) kHz 100 kHz ~ 1 MHz  (0.1 ~ 1) V 20 Hz 20 Hz ~ 20 kHz (20 ~ 100) kHz 100 kHz ~ 1 MHz	$5.8 \times 10^{-9}$  7.0 μV $1.3 \times 10^{-3}$ $3.2 \times 10^{-3}$ $1.5 \times 10^{-2}$  20 μV $8.1 \times 10^{-4}$ $2.0 \times 10^{-3}$ $2.6 \times 10^{-2}$  0.16 mV $2.9 \times 10^{-4}$ $8.2 \times 10^{-4}$ $2.4 \times 10^{-2}$	Frequency counters Digital multimeters, Spectrum analyzers, Oscilloscopes /HCT-CS-098-40423

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Output level	40423	(1 ~ 10) V 20 Hz 20 Hz ~ 20 kHz (20 ~ 100) kHz 100 kHz ~ 1 MHz  (10 ~ 100) V 20 Hz 20 Hz ~ 20 kHz (20 ~ 100) kHz  (-60 ~ 20) dBm 100 kHz ~ 100 MHz	1.6 mV $5.5 \times 10^{-4}$ $1.8 \times 10^{-3}$ $3.0 \times 10^{-2}$  16 mV $5.5 \times 10^{-4}$ $1.8 \times 10^{-3}$  0.16 dB	
DC Offset		-20 V ~ 0 mV 0 mV 0 mV ~ 20 V	$5.8 \times 10^{-4}$ $5.8 \mu V$ $5.8 \times 10^{-4}$	
Output flatness		(-10 ~ 10) dB 20 Hz ~ 100 kHz 100 kHz ~ 350 MHz	0.016 dB 0.018 dB	
Distortion factor		(-80 ~ 0) dB 20 Hz ~ 80 MHz	1.4 dB	
Output amplitude		20 Hz ~ 1 kHz (0 ~ -60) dB (1 ~ 20) kHz (0 ~ -60) dB (20 ~ 100) kHz (0 ~ -60) dB	0.007 dB 0.009 dB 0.015 dB	
Rise/Fall Time		1 ns (1 ~ 10) ns (10 ~ 100) ns 100 ns ~ 1 s	5.9 ps $1.3 \times 10^{-3}$ $1.2 \times 10^{-3}$ $1.2 \times 10^{-3}$	
AM modulation		(5 ~ 99) %	$1.2 \times 10^{-2}$	
FM modulation		(9 ~ 400) kHz	$1.2 \times 10^{-2}$	
Duty Cycle		(1 ~ 99) %	$5.8 \times 10^{-3}$	
Volt/Current recorders DC Voltage	40424	0 mV (0 ~ 1) mV (1 ~ 2) mV (2 ~ 5) mV (5 ~ 10) mV (10 ~ 20) mV (20 ~ 50) mV (50 ~ 100) mV (100 ~ 200) mV (200 ~ 500) mV 500 mV ~ 1 V (1 ~ 2) V (2 ~ 5) V (5 ~ 10) V (10 ~ 20) V (20 ~ 50) V (50 ~ 100) V (100 ~ 200) V (200 ~ 500) V (500 ~ 1 000) V	0.5 $\mu V$ $5.2 \times 10^{-4}$ $2.7 \times 10^{-4}$ $1.7 \times 10^{-4}$ $8.5 \times 10^{-5}$ $4.6 \times 10^{-5}$ $1.3 \times 10^{-4}$ $6.3 \times 10^{-5}$ $3.3 \times 10^{-5}$ $1.3 \times 10^{-4}$ $6.2 \times 10^{-5}$ $3.2 \times 10^{-5}$ $1.3 \times 10^{-4}$ $6.2 \times 10^{-5}$ $3.1 \times 10^{-5}$ $1.3 \times 10^{-4}$ $6.2 \times 10^{-5}$ $3.2 \times 10^{-5}$ $1.3 \times 10^{-4}$ $6.2 \times 10^{-5}$	Multimeter calibrators /HCT-CS-100-40424

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC Voltage	40424	0 mV (0 ~ -1) mV (-1 ~ -2) mV (-2 ~ -5) mV (-5 ~ -10) mV (-10 ~ -20) mV (-20 ~ -50) mV (-50 ~ -100) mV (-100 ~ -200) mV (-200 ~ -500) mV -500 mV ~ -1 V (-1 ~ -2) V (-2 ~ -5) V (-5 ~ -10) V (-10 ~ -20) V (-20 ~ -50) V (-50 ~ -100) V (-100 ~ -200) V (-200 ~ -500) V (-500 ~ -1 000) V	0.5 μV $5.2 \times 10^{-4}$ $2.7 \times 10^{-4}$ $1.7 \times 10^{-4}$ $8.5 \times 10^{-5}$ $4.6 \times 10^{-5}$ $1.3 \times 10^{-4}$ $6.3 \times 10^{-5}$ $3.3 \times 10^{-5}$ $1.3 \times 10^{-4}$ $6.2 \times 10^{-5}$ $3.2 \times 10^{-5}$ $1.3 \times 10^{-4}$ $6.2 \times 10^{-5}$ $3.1 \times 10^{-5}$ $1.3 \times 10^{-4}$ $6.2 \times 10^{-5}$ $3.2 \times 10^{-5}$ $1.3 \times 10^{-4}$ $6.2 \times 10^{-5}$	
DC Current		(+) 0 mA (0 ~ 1) mA (1 ~ 10) mA (10 ~ 100) mA 100 mA ~ 1 A  (-) 0 mA (0 ~ -1) mA (-1 ~ -10) mA (-10 ~ -100) mA -100 mA ~ -1 A	0.07 μA $8.0 \times 10^{-5}$ $7.8 \times 10^{-5}$ $8.7 \times 10^{-5}$ $1.3 \times 10^{-4}$  0.07 μA $8.0 \times 10^{-5}$ $7.8 \times 10^{-5}$ $8.7 \times 10^{-5}$ $1.3 \times 10^{-4}$	
Relpay test sets	40425	(20 ~ 55) Hz 100 mV 100 mV ~ 10 V (10 ~ 1 000) V  (55 ~ 300) Hz 100 mV 100 mV ~ 100 V (100 ~ 1 000) V  300 Hz ~ 1 kHz 100 mV 100 mV ~ 100 V (100 ~ 1 000) V	21 μV $1.6 \times 10^{-4}$ $1.8 \times 10^{-4}$  16 μV $1.4 \times 10^{-4}$ $1.8 \times 10^{-4}$  16 μV $1.6 \times 10^{-4}$ $1.7 \times 10^{-4}$	Multimeters Current shunts /HCT-CS-218-40425
AC Voltage				
AC Current		20 Hz ~ 1 kHz 10 mA (10 ~ 100) mA 100 mA ~ 1 A (1 ~ 10) A (10 ~ 50) A (50 ~ 100) A	5.3 μA $5.3 \times 10^{-4}$ $9.3 \times 10^{-4}$ $1.3 \times 10^{-3}$ $2.4 \times 10^{-3}$ $2.3 \times 10^{-3}$	
DC Voltage		100 mV (0.1 ~ 1) V (1 ~ 5) V (5 ~ 1 000) V	6.0 μV $5.8 \times 10^{-5}$ $1.2 \times 10^{-4}$ $5.9 \times 10^{-5}$	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC Current	40425	10 mA (10 ~ 100) mA 100 mA ~ 1 A (1 ~ 10) A (10 ~ 20) A (20 ~ 100) A	0.76 $\mu$ A $8.8 \times 10^{-5}$ $2.4 \times 10^{-4}$ $5.0 \times 10^{-4}$ $8.2 \times 10^{-3}$ $7.7 \times 10^{-4}$	
Frequency		50 Hz (50 ~ 60) Hz 60 Hz ~ 1 kHz	8.2 mHz $1.6 \times 10^{-4}$ $1.8 \times 10^{-4}$	
Time interval		1 s (1 ~ 60) s	0.01 s $1.0 \times 10^{-3}$	
LF signal generators	40426			Frequency counters, Digital multimeters, Spectrum analyzers, Oscilloscopes /HCT-CS-101-40426
Frequency		1 Hz ~ 2 MHz	$5.8 \times 10^{-9}$	
Output level		10 mV 20 Hz 20 Hz ~ 20 kHz (20 ~ 100) kHz 100 kHz ~ 1 MHz  (10 ~ 100) mV 20 Hz 20 Hz ~ 20 kHz (20 ~ 100) kHz 100 kHz ~ 1 MHz  (0.1 ~ 1) V 20 Hz 20 Hz ~ 20 kHz (20 ~ 100) kHz 100 kHz ~ 1 MHz  (1 ~ 10) V 20 Hz 20 Hz ~ 20 kHz (20 ~ 100) kHz 100 kHz ~ 1 MHz  (10 ~ 100) V 20 Hz 20 Hz ~ 20 kHz (20 ~ 100) kHz  (-60 ~ 20) dBm 20 Hz ~ 20 kHz (20 ~ 50) kHz (50 ~ 100) kHz 100 kHz ~ 1 MHz	$7.0 \mu$ V $1.3 \times 10^{-3}$ $3.2 \times 10^{-3}$ $1.5 \times 10^{-2}$  20 $\mu$ V $8.1 \times 10^{-4}$ $2.0 \times 10^{-3}$ $2.6 \times 10^{-2}$  0.16 mV $2.9 \times 10^{-4}$ $8.2 \times 10^{-4}$ $2.4 \times 10^{-2}$  1.6 mV $5.5 \times 10^{-4}$ $1.8 \times 10^{-3}$ $3.0 \times 10^{-2}$  16 mV $5.5 \times 10^{-4}$ $1.8 \times 10^{-3}$  0.007 dB 0.008 dB 0.013 dB 0.16 dB	
DC Offset		-20 V ~ 0 mV 0 mV 0 mV ~ 20 V	$5.8 \times 10^{-4}$ $5.8 \mu$ V $5.8 \times 10^{-4}$	
Output flatness		(-10 ~ 10) dB 20 Hz ~ 100 kHz 100 kHz ~ 1 MHz	0.016 dB 0.018 dB	
Distortion factor		(-80 ~ 0) dB 20 Hz ~ 1 MHz	1.4 dB	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Output amplitude	40426	20 Hz ~ 1 kHz (0 ~ -60) dB (1 ~ 20) kHz (0 ~ -60) dB (20 ~ 100) kHz (0 ~ -60) dB	0.007 dB 0.009 dB 0.015 dB	
Rise/Fall time		1 ns (1 ~ 10) ns (10 ~ 100) ns 100 ns ~ 1 s	5.9 ps $1.3 \times 10^{-3}$ $1.2 \times 10^{-3}$ $1.2 \times 10^{-3}$	
AM modulation		(5 ~ 99) %	$1.2 \times 10^{-2}$	
FM modulation		(9 ~ 400) kHz	$1.2 \times 10^{-2}$	
Duty Cycle		(1 ~ 99) %	$5.8 \times 10^{-3}$	
LF spectrum analyzer Input level	40427	27 dBm 10 Hz 10 Hz ~ 20 kHz (20 ~ 100) kHz  (27 ~ 10) dBm 10 Hz 10 Hz ~ 20 kHz 20 kHz ~ 100 kHz (100 ~ 200) kHz  (10 ~ -10) dBm 10 Hz 10 Hz ~ 20 kHz (20 ~ 100) kHz (100 ~ 200) kHz  (-10 ~ -40) dBm 10 Hz 10 Hz ~ 20 kHz (20 ~ 100) kHz (100 ~ 200) kHz  (-40 ~ -50) dBm 10 Hz 10 Hz ~ 20 kHz (20 ~ 100) kHz (100 ~ 200) kHz  10 mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz (100 ~ 200) kHz  (10 ~ 100) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz (100 ~ 200) kHz	0.008 dB 0.007 dB 0.008 dB  0.008 dB 0.007 dB 0.007 dB 0.009 dB  0.008 dB 0.007 dB 0.007 dB 0.009 dB  0.008 dB 0.008 dB 0.013 dB 0.022 dB  0.017 dB 0.016 dB 0.024 dB 0.045 dB  22 $\mu$ V $2.2 \times 10^{-3}$ $2.3 \times 10^{-3}$ $2.3 \times 10^{-3}$  88 $\mu$ V $6.3 \times 10^{-4}$ $8.3 \times 10^{-4}$ $1.2 \times 10^{-3}$	Function generators /HCT-CS-180-40427



404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Input level	40427	(0.1 ~ 1) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz (100 ~ 200) kHz  (1 ~ 10) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz (100 ~ 200) kHz  (10 ~ 30) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	0.69 mV $6.2 \times 10^{-4}$ $6.3 \times 10^{-4}$ $7.8 \times 10^{-4}$  6.9 mV $6.2 \times 10^{-4}$ $6.3 \times 10^{-4}$ $7.3 \times 10^{-4}$  16 mV $2.4 \times 10^{-4}$ $3.9 \times 10^{-4}$	
Input frequency		10 Hz ~ 200 kHz	$6.2 \times 10^{-5}$	
Input impedance		1 MΩ	0.62 kΩ	
Output level(AC)		10 mV 20 Hz 20 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz  (10 ~ 100) mV 20 Hz 20 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz  (0.1 ~ 1) V 20 Hz 20 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz  (1 ~ 10) V 20 Hz 20 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz	8.6 μV $7.1 \times 10^{-4}$ $1.5 \times 10^{-3}$ $3.2 \times 10^{-3}$  64 μV $6.3 \times 10^{-4}$ $7.6 \times 10^{-3}$ $1.3 \times 10^{-3}$  0.63 mV $6.2 \times 10^{-4}$ $6.8 \times 10^{-4}$ $1.0 \times 10^{-3}$  6.3 mV $6.2 \times 10^{-4}$ $6.7 \times 10^{-3}$ $1.0 \times 10^{-3}$	
Output level(AC)		10 mV 10 mV ~ 10 V	6.3 μV $6.2 \times 10^{-5}$	
Sweep generators	40429	1 Hz ~ 21 MHz  10 mV 20 Hz 20 Hz ~ 20 kHz (20 ~ 100) kHz 100 kHz ~ 1 MHz  (10 ~ 100) mV 20 Hz 20 Hz ~ 20 kHz (20 ~ 100) kHz 100 kHz ~ 1 MHz	$5.8 \times 10^{-9}$  7.0 μV $1.3 \times 10^{-3}$ $3.2 \times 10^{-3}$ $1.5 \times 10^{-2}$  20 μV $8.1 \times 10^{-4}$ $2.0 \times 10^{-3}$ $2.6 \times 10^{-2}$	Frequency counters, Digital multimeters, Spectrum analyzers, Oscilloscopes /HCT-CS-102-40429

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Output level	40429	(0.1 ~ 1) V 20 Hz 20 Hz ~ 20 kHz (20 ~ 100) kHz 100 kHz ~ 1 MHz  (1 ~ 10) V 20 Hz 20 Hz ~ 20 kHz (20 ~ 100) kHz 100 kHz ~ 1 MHz  (10 ~ 100) V 20 Hz 20 Hz ~ 20 kHz (20 ~ 100) kHz  -60 dBm ~ 20 dBm 100 kHz ~ 21 MHz	0.16 mV $2.9 \times 10^{-4}$ $8.2 \times 10^{-4}$ $2.4 \times 10^{-2}$  1.6 mV $5.5 \times 10^{-4}$ $1.8 \times 10^{-3}$ $3.0 \times 10^{-2}$  16 mV $5.5 \times 10^{-4}$ $1.8 \times 10^{-3}$  0.16 dB	
DC Offset		-20 V ~ 0 mV 0 mV 0 mV ~ 20 V	$5.8 \times 10^{-4}$ $5.8 \mu V$ $5.8 \times 10^{-4}$	
Output flatness		(-10 ~ 10) dB 20 Hz ~ 100 kHz 100 kHz ~ 21 MHz	0.016 dB 0.018 dB	
Distortion factor		(-80 ~ 0) dB 20 Hz ~ 21 MHz	1.4 dB	
Output amplitude		20 Hz ~ 1 kHz (0 ~ -60) dB (1 ~ 20) kHz (0 ~ -60) dB (20 ~ 100) kHz (0 ~ -60) dB	0.007 dB 0.009 dB 0.015 dB	
Rise/Fall Time		1 ns (1 ~ 10) ns (10 ~ 100) ns 100 ns ~ 1 s	5.9 ps $5.9 \times 10^{-3}$ $1.3 \times 10^{-3}$ $1.2 \times 10^{-3}$	
AM modulation		(5 ~ 99) %	$1.2 \times 10^{-2}$	
FM modulation		(9 ~ 400) kHz	$1.2 \times 10^{-2}$	
Duty Cycle		(1 ~ 99) %	$5.8 \times 10^{-3}$	
Transistor curve tracers DC Voltage (SMU, Base/Emitter/Collector)	40432	(-1 000 ~ -200) V (-200 ~ -100) V (-100 ~ -10) V (-10 ~ -1) V (-1 ~ -0.1) V (-0.1 ~ 0) V 0 V (0 ~ 0.1) V (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 200) V (200 ~ 1 000) V	$5.5 \times 10^{-6}$ $1.1 \times 10^{-5}$ $5.2 \times 10^{-6}$ $3.7 \times 10^{-6}$ $7.0 \times 10^{-6}$ $4.9 \times 10^{-6}$ 0.13 $\mu V$ $4.9 \times 10^{-6}$ $7.0 \times 10^{-6}$ $3.7 \times 10^{-6}$ $5.2 \times 10^{-6}$ $1.1 \times 10^{-5}$ $5.5 \times 10^{-6}$	Multimeter calibrators, Digital multimeters, Electrometers, High resistance meters /HCT-CS-103-40432

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.				
DC Voltage (VSU, Base/Emitter/Collector)	40432	(-1 000 ~ -200) V	$5.5 \times 10^{-6}$					
		(-200 ~ -100) V	$1.1 \times 10^{-5}$					
		(-100 ~ -10) V	$5.2 \times 10^{-6}$					
		(-10 ~ -1) V	$3.7 \times 10^{-6}$					
		(-1 ~ -0.1) V	$7.0 \times 10^{-6}$					
		(-0.1 ~ 0) V	$4.9 \times 10^{-6}$					
		0 V	0.13 $\mu$ V					
		(0 ~ 0.1) V	$4.9 \times 10^{-6}$					
		(0.1 ~ 1) V	$7.0 \times 10^{-6}$					
		(1 ~ 10) V	$3.7 \times 10^{-6}$					
		(10 ~ 100) V	$5.2 \times 10^{-6}$					
		(100 ~ 200) V	$1.1 \times 10^{-5}$					
		(200 ~ 1 000) V	$5.5 \times 10^{-6}$					
		VMU (Base/Emitter/Collector)			(-1 000 ~ -200) V	$5.5 \times 10^{-6}$		
					(-200 ~ -100) V	$1.1 \times 10^{-5}$		
					(-100 ~ -10) V	$5.2 \times 10^{-6}$		
					(-10 ~ -1) V	$3.7 \times 10^{-6}$		
					(-1 ~ -0.1) V	$7.0 \times 10^{-6}$		
					-(0.1 ~ 0) V	$4.9 \times 10^{-6}$		
					0 V	0.13 $\mu$ V		
					(0 ~ 0.1) V	$4.9 \times 10^{-6}$		
					(0.1 ~ 1) V	$7.0 \times 10^{-6}$		
					(1 ~ 10) V	$3.7 \times 10^{-6}$		
					(10 ~ 100) V	$5.2 \times 10^{-6}$		
					(100 ~ 200) V	$1.1 \times 10^{-5}$		
					(200 ~ 1 000) V	$5.5 \times 10^{-6}$		
					DC Current (SMU, Base/Emitter/Collector)		(-50 ~ -20) A	$1.3 \times 10^{-5}$
							(-20 ~ -10) A	$8.3 \times 10^{-6}$
(-10 ~ -2) A	$4.9 \times 10^{-4}$							
(-2 ~ -1) A	$7.0 \times 10^{-4}$							
(-1 ~ -0.1) A	$2.2 \times 10^{-4}$							
(-100 ~ -10) mA	$4.8 \times 10^{-5}$							
(-10 ~ -1) mA	$1.5 \times 10^{-5}$							
(-1 ~ -0.1) mA	$1.3 \times 10^{-5}$							
(-100 ~ -10) $\mu$ A	$1.4 \times 10^{-5}$							
(-10 ~ -1) $\mu$ A	$8.1 \times 10^{-5}$							
(-1 ~ -0.1) $\mu$ A	$7.6 \times 10^{-4}$							
(-100 ~ -10) nA	$2.4 \times 10^{-3}$							
(-10 ~ -1) nA	$2.4 \times 10^{-3}$							
(-1 ~ -0.1) nA	$5.8 \times 10^{-3}$							
(-100 ~ -10) pA	$1.2 \times 10^{-2}$							
(-10 ~ -1) pA	$1.2 \times 10^{-2}$							
-1 pA ~ 0 A	$1.5 \times 10^{-2}$							
0 A	8.1 nA							
0 A ~ 1 pA	$1.5 \times 10^{-2}$							
(1 ~ 10) pA	$1.2 \times 10^{-2}$							
(10 ~ 100) pA	$1.2 \times 10^{-2}$							
(0.1 ~ 1) nA	$5.8 \times 10^{-3}$							
(1 ~ 10) nA	$2.4 \times 10^{-3}$							
(10 ~ 100) nA	$2.4 \times 10^{-3}$							
(0.1 ~ 1) $\mu$ A	$7.6 \times 10^{-4}$							
(1 ~ 10) $\mu$ A	$8.1 \times 10^{-5}$							
(10 ~ 100) $\mu$ A	$1.4 \times 10^{-5}$							
(0.1 ~ 1) mA	$1.3 \times 10^{-5}$							
(1 ~ 10) mA	$1.5 \times 10^{-5}$							
(10 ~ 100) mA	$4.8 \times 10^{-5}$							
(0.1 ~ 1) A	$2.2 \times 10^{-4}$							
(1 ~ 2) A	$7.0 \times 10^{-4}$							
(2 ~ 10) A	$4.9 \times 10^{-4}$							
(10 ~ 20) A	$8.3 \times 10^{-6}$							
(20 ~ 50) A	$1.3 \times 10^{-5}$							

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Waveform analyzers	40433			Multimeter calibrators Digital multimeters /HCT-CS-104-40433
Output frequency		1 Hz ~ 1 MHz	$6.2 \times 10^{-5}$	
Outout level		2 mV		
		20 Hz	7.9 $\mu$ V	
		20 Hz ~ 1 kHz	$3.5 \times 10^{-3}$	
		(1 ~ 20) kHz	$6.0 \times 10^{-3}$	
		(20 ~ 100) kHz	$1.3 \times 10^{-2}$	
		(2 ~ 100) mV		
		20 Hz	20 $\mu$ V	
		20 Hz ~ 1 kHz	$1.8 \times 10^{-4}$	
		(1 ~ 20) kHz	$4.5 \times 10^{-4}$	
		(20 ~ 100) kHz	$1.1 \times 10^{-3}$	
		(0.1 ~ 1) V		
		20 Hz	0.16 mV	
	20 Hz ~ 1 kHz	$1.2 \times 10^{-4}$		
	(1 ~ 20) kHz	$2.9 \times 10^{-4}$		
	(20 ~ 100) kHz	$8.2 \times 10^{-4}$		
	(1 ~ 10) V			
	20 Hz	1.6 mV		
	20 Hz ~ 1 kHz	$1.2 \times 10^{-4}$		
	(1 ~ 20) kHz	$2.9 \times 10^{-4}$		
	(20 ~ 100) kHz	$8.2 \times 10^{-4}$		
	(10 ~ 100) V			
	20 Hz	16 mV		
	20 Hz ~ 1 kHz	$1.7 \times 10^{-4}$		
	(1 ~ 20) kHz	$3.5 \times 10^{-4}$		
	(20 ~ 100) kHz	$8.1 \times 10^{-4}$		
Output DC Offset		-20 V ~ 0 mV	$6.2 \times 10^{-5}$	
		0 mV	6.2 $\mu$ V	
		0 mV ~ 20 V	$6.2 \times 10^{-5}$	
Output flatness		(-10 ~ 10) dB		
		20 Hz ~ 20 kHz	0.006 3 dB	
		(20 ~ 100) kHz	0.009 2 dB	
Output amplitude		20 Hz ~ 1 kHz		
		(-10 ~ -60) dB	0.061 dB	
		(1 ~ 20) kHz		
		(-10 ~ -60) dB	0.11 dB	
		(20 ~ 50) kHz		
		(-10 ~ -60) dB	0.11 dB	
	(50 ~ 100) kHz			
	(-10 ~ -60) dB	0.11 dB		
Output impedance		50 $\Omega$	6.2 m $\Omega$	
		600 $\Omega$	62 m $\Omega$	
Input frequency		1 Hz ~ 1 MHz	$6.2 \times 10^{-5}$	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC Input level	40433	2 mV 10 Hz 10 Hz ~ 20 kHz (20 ~ 50) kHz (50 ~ 100) kHz (2 ~ 100) mV 10 Hz 10 Hz ~ 20 kHz (20 ~ 50) kHz (50 ~ 100) kHz  (0.1 ~ 1) V 10 Hz 10 Hz ~ 20 kHz (20 ~ 50) kHz (50 ~ 100) kHz  (1 ~ 10) V 10 Hz 10 Hz ~ 20 kHz (20 ~ 50) kHz (50 ~ 100) kHz  (10 ~ 100) V 10 Hz 10 Hz ~ 20 kHz (20 ~ 50) kHz (50 ~ 100) kHz  (100 ~ 300) V 10 Hz 10 Hz ~ 10 kHz	7.8 μV $3.8 \times 10^{-3}$ $3.9 \times 10^{-3}$ $4.6 \times 10^{-3}$  43 μV $1.6 \times 10^{-4}$ $2.4 \times 10^{-4}$ $5.7 \times 10^{-4}$  0.14 mV $9.0 \times 10^{-5}$ $1.5 \times 10^{-4}$ $1.5 \times 10^{-4}$  3.4 mV $9.0 \times 10^{-5}$ $1.2 \times 10^{-4}$ $1.4 \times 10^{-4}$  10 mV $1.0 \times 10^{-4}$ $1.3 \times 10^{-4}$ $2.2 \times 10^{-4}$  0.13 V $2.4 \times 10^{-4}$	
DC Input level		1 mV (1 ~ 100) mV (0.1 ~ 100) V (100 ~ 300) V	6.2 μV $6.3 \times 10^{-5}$ $6.2 \times 10^{-5}$ $6.2 \times 10^{-5}$	
Filter(weight,low,high pass, etc.)		400 Hz ~ 80 kHz	$2.1 \times 10^{-4}$	
Distortion factor		20 Hz ~ 20 kHz (-10 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB  20 Hz ~ 20 kHz (0.001 ~ 0.01) % (0.01 ~ 30) %	0.31 dB 0.38 dB 0.56 dB  $5.5 \times 10^{-2}$ $3.1 \times 10^{-2}$	
AC/DC high generator DC Voltage	40434	(+) 1 V 1 ~ 100 V (0.1 ~ 1) kV (1 ~ 2) kV (2 ~ 10) kV (10 ~ 20) kV (20 ~ 50) kV (50 ~ 100) kV	1 mV $6.1 \times 10^{-4}$ $1.1 \times 10^{-5}$ $1.0 \times 10^{-3}$ $6.1 \times 10^{-4}$ $4.1 \times 10^{-4}$ $4.4 \times 10^{-4}$ $4.1 \times 10^{-4}$	High voltage voltmeters /HCT-CS-055-40434

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC Voltage	40434	(-) -1 V 1 ~ 100 V (0.1 ~ 1) kV (1 ~ 2) kV	1 mV $6.1 \times 10^{-4}$ $1.1 \times 10^{-5}$ $1.0 \times 10^{-3}$	
AC Voltage		(2 ~ 10) kV (10 ~ 20) kV (20 ~ 50) kV (50 ~ 100) kV (50 ~ 60) Hz 100 V (0.1 ~ 1) kV (1 ~ 10) kV (10 ~ 20) kV (20 ~ 40) kV (40 ~ 60) kV (60 ~ 75) kV	$6.1 \times 10^{-4}$ $4.1 \times 10^{-4}$ $4.4 \times 10^{-4}$ $4.1 \times 10^{-4}$ 13 mV $1.6 \times 10^{-7}$ $2.4 \times 10^{-3}$ $2.3 \times 10^{-3}$ $2.1 \times 10^{-3}$ $2.0 \times 10^{-3}$ $4.3 \times 10^{-3}$	
AC/DC high voltage probes DC Voltage Ratio	40435	(±) 1 kV less than (1 : 1) 10 mV ~ 1 000 V  (1 ~ 5 : 1) 100 mV ~ 1 000 V  (5 ~ 10 : 1) 100 mV ~ 1 000 V  (10 ~ 50 : 1) (1 ~ 1 000) V  (50 ~ 100 : 1) (10 ~ 1 000) V  (100 ~ 500 : 1) (10 ~ 1 000) V  (500 ~ 1 000 : 1) (100 ~ 1 000) V  1 kV 이상 (100 : 1) (1 ~ 5) kV  (100 ~ 1 000 : 1) (1 ~ 100) kV  (1 000 ~ 10 000 : 1) (1 ~ 100) kV	$3.9 \times 10^{-5}$  $2.2 \times 10^{-4}$  $1.8 \times 10^{-4}$  $1.7 \times 10^{-3}$  $2.2 \times 10^{-3}$  $5.3 \times 10^{-2}$  0.20 %  0.053 %  0.53 %  5.4 %	High voltage sources /HCT-CS-056-40435
DC Voltage(SCOPE PROBE)		(±) 1 V 1 V ~ 1 kV 1 kV ~ 20 kV 20 kV ~ 40 kV	0.1 mV $1.0 \times 10^{-4}$ $1.4 \times 10^{-3}$ $1.3 \times 10^{-3}$	
AC Voltage		50 Hz ~ 1 kHz 1 V (1 ~ 10) V	0.09 mV $1.3 \times 10^{-4}$	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC Voltage	40435	(1 ~ 100) V 100 V ~ 1 kV  (50 ~ 60) Hz 1 kV (1 ~ 10) kV  (10 ~ 20) kV (20 ~ 60) kV (60 ~ 70) kV	$1.5 \times 10^{-4}$ $6.2 \times 10^{-4}$  0.07 kV $1.5 \times 10^{-2}$  $1.4 \times 10^{-2}$ $1.3 \times 10^{-2}$ $1.4 \times 10^{-2}$	
Resistance		1 Ω (1 ~ 10) Ω (0.01 ~ 10) kΩ (10 ~ 100) kΩ (0.1 ~ 10) MΩ (10 ~ 100) MΩ (0.1 ~ 1) GΩ	0.58 mΩ $5.8 \times 10^{-4}$ $5.8 \times 10^{-4}$ $5.9 \times 10^{-5}$ $5.8 \times 10^{-4}$ $1.7 \times 10^{-4}$ $1.9 \times 10^{-3}$	
Capacitance		(1 kHz) 1 pF (1 ~ 10) pF (10 ~ 100) pF (0.1 ~ 1) nF (1 ~ 10) nF	5.9 fF $9.3 \times 10^{-4}$ $6.2 \times 10^{-4}$ $5.9 \times 10^{-3}$ $8.5 \times 10^{-4}$	
Logic analyzers Input voltage	40436	100 mV (0.1 ~ 1) V (1 ~ 2) V (2 ~ 3) V (3 ~ 10) V  -100 mV (-0.1 ~ -1) V (-1 ~ -2) V (-2 ~ -3) V (-3 ~ -10) V	$6.3 \mu V$ $1.3 \times 10^{-4}$ $6.2 \times 10^{-5}$ $3.2 \times 10^{-5}$ $8.8 \times 10^{-5}$  $6.3 \mu V$ $1.3 \times 10^{-4}$ $6.2 \times 10^{-5}$ $3.2 \times 10^{-5}$ $8.8 \times 10^{-5}$	Multimeter calibrators /HCT-CS-201-40436
Telephone testers Frequency AC Amplitude	40437	1 Hz ~ 1 MHz  10 mV 20 Hz 20 Hz ~ 1 kHz (1 ~ 20) kHz (20 ~ 100) kHz  (10 ~ 100) mV 20 Hz 20 Hz ~ 1 kHz (1 ~ 20) kHz (20 ~ 100) kHz  (0.1 ~ 1) V 20 Hz 20 Hz ~ 1 kHz (1 ~ 20) kHz (20 ~ 100) kHz	$6.2 \times 10^{-7}$  10 μV $9.1 \times 10^{-4}$ $1.5 \times 10^{-3}$ $3.3 \times 10^{-3}$  20 μV $1.8 \times 10^{-4}$ $4.5 \times 10^{-4}$ $1.1 \times 10^{-3}$  0.16 mV $1.4 \times 10^{-4}$ $2.9 \times 10^{-4}$ $8.5 \times 10^{-4}$	Frequency cutters, Digital multimeters /HCT-CS-127-40437

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC Amplitude	40437	(1 ~ 10) V 20 Hz 20 Hz ~ 1 kHz (1 ~ 20) kHz (20 ~ 100) kHz  (10 ~ 100) V 20 Hz 20 Hz ~ 1 kHz (1 ~ 20) kHz (20 ~ 100) kHz  (100 ~ 500) V 20 Hz 20 Hz ~ 1 kHz  (20 ~ -10) dBm 20 Hz 20 Hz ~ 20 kHz (20 ~ 100) kHz  (-10 ~ -40) dBm 20 Hz 20 Hz ~ 20 kHz (20 ~ 100) kHz	1.6 mV 1.4×10 <sup>-4</sup> 2.9×10 <sup>-4</sup> 8.2×10 <sup>-4</sup>  1.6 mV 1.4×10 <sup>-4</sup> 2.9×10 <sup>-4</sup> 8.2×10 <sup>-4</sup>  91 mV 1.9×10 <sup>-4</sup>  0.006 2 dB 0.006 3 dB 0.010 dB  0.006 1 dB 0.007 0 dB 0.011 dB	
Loop Current		1 mA (1 ~ 100) mA (0.1 ~ 1) A	0.62 μA 6.2×10 <sup>-4</sup> 6.2×10 <sup>-4</sup>	
DC Voltage		10 mV 10 mV ~ 100 V (100 ~ 500) V	6.2 μV 6.2×10 <sup>-4</sup> 1.3×10 <sup>-4</sup>	
Dial Level		(-39 ~ 10) dBm	0.58 dB	
Resistance		50 Ω (50 ~ 1 000) Ω	6.2 mΩ 6.2×10 <sup>-4</sup>	
Video signal analyzers SQUARE WAVE level	40438	50 mV (50 ~ 100) mV (100 ~ 200) mV (200 ~ 300) mV (300 ~ 400) mV (400 ~ 500) mV (500 ~ 600) mV (600 ~ 1 000) mV  50 mV (50 ~ 100) mV (100 ~ 200) mV (200 ~ 300) mV (300 ~ 400) mV (400 ~ 500) mV (500 ~ 600) mV (600 ~ 700) mV (700 ~ 1 000) mV	0.11 mV 1.5×10 <sup>-3</sup> 1.4×10 <sup>-3</sup> 1.3×10 <sup>-3</sup> 1.9×10 <sup>-3</sup> 1.7×10 <sup>-3</sup> 1.5×10 <sup>-3</sup> 1.5×10 <sup>-3</sup>  1.4 mV 2.1×10 <sup>-2</sup> 2.7×10 <sup>-2</sup> 2.2×10 <sup>-2</sup> 2.5×10 <sup>-2</sup> 2.2×10 <sup>-2</sup> 2.1×10 <sup>-2</sup> 3.4×10 <sup>-2</sup> 3.1×10 <sup>-2</sup>	Video signal generators / HCT-CS-130-40438
BURST Frequency		(3 ~ 5) MHz	4.8×10 <sup>-7</sup>	



404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Vector scopes, Video signal monitors Color Bar Level(chrominance)	40438	50 mV	3.4 mV	
		(50 ~ 100) mV	$3.4 \times 10^{-2}$	
		(100 ~ 200) mV	$2.4 \times 10^{-2}$	
		(200 ~ 300) mV	$1.6 \times 10^{-2}$	
		(300 ~ 400) mV	$1.3 \times 10^{-2}$	
		(400 ~ 800) mV	$1.2 \times 10^{-2}$	
		(800 ~ 1 000) mV	$9.8 \times 10^{-3}$	
Color Bar phase		0 ° ~ 360 °	0.7 °	
Frequency		50 Hz ~ 10 MHz	$5.8 \times 10^{-5}$	
Vertical Level		50 mV	2.1 mV	
		(50 ~ 100) mV	$2.1 \times 10^{-2}$	
		(100 ~ 200) mV	$1.4 \times 10^{-2}$	
		(200 ~ 300) mV	$8.4 \times 10^{-3}$	
		(300 ~ 400) mV	$6.2 \times 10^{-3}$	
		(400 ~ 500) mV	$4.9 \times 10^{-3}$	
		(500 ~ 600) mV	$4.1 \times 10^{-3}$	
		(600 ~ 700) mV	$3.3 \times 10^{-3}$	
		(700 ~ 800) mV	$3.1 \times 10^{-3}$	
		(800 ~ 900) mV	$2.8 \times 10^{-3}$	
		(900 ~ 1 000) mV	$3.2 \times 10^{-3}$	
Vertical Level(Response)		(50 kHz ~ 10 MHz)		
		50 mV	2.5 mV	
		(0 ~ 100) mV	$2.9 \times 10^{-2}$	
		(100 ~ 200) mV	$3.0 \times 10^{-2}$	
		(200 ~ 300) mV	$2.3 \times 10^{-2}$	
		(300 ~ 400) mV	$2.6 \times 10^{-2}$	
		(400 ~ 500) mV	$2.3 \times 10^{-2}$	
		(500 ~ 600) mV	$2.1 \times 10^{-2}$	
		(600 ~ 700) mV	$3.5 \times 10^{-2}$	
		(700 ~ 800) mV	$3.1 \times 10^{-2}$	
		(800 ~ 900) mV	$2.9 \times 10^{-2}$	
		(900 ~ 1 000) mV	$2.7 \times 10^{-2}$	

405. Low frequency electronic & magnetic field

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Flux meters	40503	0.1 mWb (0.1 ~ 1) mWb 1 mWb ~ 10 Wb	0.6 $\mu$ Wb $1.5 \times 10^{-3}$ $8.2 \times 10^{-4}$	Volt-second generator / HCT-CS-257-40503
Flux sources	40504	0.1 mWb (0.1 ~ 1) mWb (1 ~ 10) mWb (10 ~ 100) mWb (0.1 ~ 10) Wb	12 nWb $1.2 \times 10^{-4}$ $1.4 \times 10^{-5}$ $6.1 \times 10^{-6}$ $1.2 \times 10^{-5}$	DMM, Counter, Scope / HCT-CS-258-40504
Magnetometers	40508	0 mT (0 ~ 1) mT (1 ~ 25) mT (40 ~ 150) mT (150 ~ 1 000) mT (1 ~ 1.9) T	2.2 $\mu$ T $6.7 \times 10^{-3}$ $3.7 \times 10^{-3}$ $7.3 \times 10^{-4}$ $7.2 \times 10^{-4}$ $7.9 \times 10^{-4}$	Helmholtz coil, Electro magnet NMR teslameter / HCT-CS-259-40508
Reference/standard magnets	40510	5 mT (5 ~ 10) mT (10 ~ 100) mT (0.1 ~ 1) T (1 ~ 2) T	31 $\mu$ T $3.2 \times 10^{-3}$ $3.7 \times 10^{-3}$ $1.5 \times 10^{-3}$ $1.2 \times 10^{-3}$	Helmholtz coil, Electro magnet Gauss meter / HCT-CS-260-40510



406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Attenuation	40602	30 dB ~ 40 dB 40 dB ~ 50 dB 50 dB ~ 60 dB 60 dB ~ 70 dB 70 dB ~ 80 dB 80 dB ~ 90 dB 90 dB ~ 100 dB 100 dB ~ 110 dB 110 dB ~ 120 dB  (26.5 GHz ~ 50 GHz) 0 dB ~ 10 dB 10 dB ~ 20 dB 20 dB ~ 30 dB 30 dB ~ 40 dB 40 dB ~ 50 dB 50 dB ~ 60 dB	0.05 dB 0.06 dB 0.06 dB 0.07 dB 0.08 dB 0.08 dB 0.09 dB 0.09 dB 0.10 dB  0.21 dB 0.23 dB 0.29 dB 0.30 dB 0.47 dB 1.2 dB	
Waveguide attenuators Attenuation	40603	(40 GHz ~ 75 GHz) 0 dB ~ 10 dB 10 dB ~ 20 dB 20 dB ~ 30 dB 30 dB ~ 40 dB 40 dB ~ 50 dB 50 dB ~ 60 dB  (75 GHz ~ 110 GHz) 0 dB ~ 10 dB 10 dB ~ 20 dB 20 dB ~ 30 dB 30 dB ~ 40 dB 40 dB ~ 50 dB 50 dB ~ 60 dB	0.34 dB 0.34 dB 0.34 dB 0.34 dB 0.34 dB 0.36 dB  0.34 dB 0.34 dB 0.34 dB 0.34 dB 0.34 dB 0.36 dB	Network analyzers, Calibration kits / HCT-CS-343-40603
Burst pulse generators Output Voltage	40605	50 Ω (±) 10 V (10 ~ 20) V (20 ~ 50) V (50 ~ 200) V (200 ~ 500) V (0.5 ~ 1) kV (1 ~ 2) kV (2 ~ 2.5) kV (2.5 ~ 3) kV (3 ~ 4) kV	0.39 V $3.8 \times 10^{-2}$ $3.3 \times 10^{-2}$ $3.8 \times 10^{-2}$ $3.3 \times 10^{-2}$ $3.8 \times 10^{-2}$ $3.1 \times 10^{-2}$ $2.8 \times 10^{-2}$ $2.6 \times 10^{-2}$ $3.8 \times 10^{-2}$	Attenuators, Oscilloscopes / HCT-CS-109-40605

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Output Voltage	40605	1 kΩ (±) 10 V (10 ~ 40) V (40 ~ 100) V (100 ~ 400) V (0.4 ~ 1) kV (1 ~ 2) kV (2 ~ 4) kV (4 ~ 5) kV (5 ~ 6) kV (6 ~ 8) kV	0.36 V $4.2 \times 10^{-2}$ $3.8 \times 10^{-2}$ $4.2 \times 10^{-2}$ $3.8 \times 10^{-2}$ $4.2 \times 10^{-2}$ $3.6 \times 10^{-2}$ $3.4 \times 10^{-2}$ $3.3 \times 10^{-2}$ $3.1 \times 10^{-2}$	
Delta time measurement (rise/fall/duration/period/ repetition rate/burst duration)		1.0 ns (1.0 ~ 2.0) ns (2.0 ~ 5.0) ns (5.0 ~ 10.0) ns (10 ~ 20) ns (20 ~ 50) ns (50 ~ 100) ns (100 ~ 200) ns (200 ~ 500) ns (0.5 ~ 1.0) μs (1.0 ~ 2.0) μs (2.0 ~ 5.0) μs (5.0 ~ 10.0) μs (10 ~ 20) μs (20 ~ 50) μs (50 ~ 100) μs (100 ~ 200) μs (200 ~ 500) μs (0.5 ~ 1) ms (1 ~ 2) ms (2 ~ 5) ms (5 ~ 10) ms (10 ~ 20) ms (20 ~ 50) ms (50 ~ 100) ms (100 ~ 200) ms (200 ~ 500) ms (0.5 ~ 1.0) s (1.0 ~ 2.0) s (2.0 ~ 5.0) s	0.014 ns $7.0 \times 10^{-3}$ $2.8 \times 10^{-3}$ $1.5 \times 10^{-3}$ $7.5 \times 10^{-4}$ $3.1 \times 10^{-4}$ $6.0 \times 10^{-4}$ $6.7 \times 10^{-4}$ $2.7 \times 10^{-4}$ $1.2 \times 10^{-3}$ $5.8 \times 10^{-4}$ $2.3 \times 10^{-4}$ $5.9 \times 10^{-4}$ $3.1 \times 10^{-4}$ $1.3 \times 10^{-4}$ $8.3 \times 10^{-4}$ $4.2 \times 10^{-4}$ $6.1 \times 10^{-4}$ $2.8 \times 10^{-2}$ $3.5 \times 10^{-2}$ $2.3 \times 10^{-4}$ $5.9 \times 10^{-4}$ $3.0 \times 10^{-4}$ $1.6 \times 10^{-4}$ $5.8 \times 10^{-4}$ $6.7 \times 10^{-4}$ $2.7 \times 10^{-4}$ $2.1 \times 10^{-3}$ $1.0 \times 10^{-3}$ $4.2 \times 10^{-4}$	
Frequency measurement		2.5 kHz (2.5 ~ 5) kHz (5 ~ 10) kHz (10 ~ 100) kHz (0.1 ~ 1) MHz (1 ~ 3) MHz (3 ~ 10) MHz (10 ~ 30) MHz (30 ~ 100) MHz	1.6 Hz $3.3 \times 10^{-4}$ $8.8 \times 10^{-4}$ $6.6 \times 10^{-4}$ $1.2 \times 10^{-3}$ $3.2 \times 10^{-4}$ $6.6 \times 10^{-4}$ $3.6 \times 10^{-4}$ $1.5 \times 10^{-3}$	
Attenuator calibrators Attenuation measurement accuracy	40606	0 dB ~ 10 dB 10 dB ~ 20 dB 20 dB ~ 30 dB 30 dB ~ 40 dB 40 dB ~ 50 dB 50 dB ~ 60 dB 60 dB ~ 70 dB	0.027 dB 0.029 dB 0.032 dB 0.038 dB 0.043 dB 0.043 dB 0.048 dB	Standard attenuators / HCT-CS-175-40606

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Attenuation measurement accuracy	40606	70 dB ~ 80 dB 80 dB ~ 90 dB 90 dB ~ 100 dB 100 dB ~ 110 dB 110 dB ~ 120 dB	0.054 dB 0.060 dB 0.066 dB 0.069 dB 0.074 dB	
RF power meter calibrators Power range	40607	3 μW 10 μW 30 μW 100 μW 300 μW 1 mW 3 mW 10 mW 30 mW 100 mW	0.27 nW 0.44 nW 1.8 nW 2.9 nW 15 nW 0.02 μW 0.10 μW 0.18 μW 0.45 μW 2.5 μW	Digital multimeter / HCT-CS-166-40607
EMC transducers; current probes, absorbing clamps etc. EMC transducers Transfer impedance  Reflection coefficient  Absorbing clamps Insertion loss	40608	5 Hz ~ 400 MHz 400 MHz ~ 3 GHz  5 Hz ~ 3 GHz  30 MHz ~ 1 GHz	0.54 dB 1.1 dB  5.9×10 <sup>-3</sup>  1.8 dB	Network analyzers, Calibration kits / HCT-CS-167-40608 / HCT-CS-198-40608
Coaxial directional couplers /splitters  Coupling factor	40610	(5 Hz ~ 9 kHz) 0 dB ~ 10 dB 10 dB ~ 20 dB 20 dB ~ 30 dB 30 dB ~ 40 dB 40 dB ~ 50 dB 50 dB ~ 60 dB  (9 kHz ~ 26.5 GHz) 0 dB ~ 10 dB 10 dB ~ 20 dB 20 dB ~ 30 dB 30 dB ~ 40 dB 40 dB ~ 50 dB 50 dB ~ 60 dB 60 dB ~ 70 dB 70 dB ~ 80 dB 80 dB ~ 90 dB 90 dB ~ 100 dB 100 dB ~ 110 dB 110 dB ~ 120 dB  (26.5 GHz ~ 40 GHz) 0 dB ~ 10 dB 10 dB ~ 20 dB 20 dB ~ 30 dB 30 dB ~ 40 dB 40 dB ~ 50 dB 50 dB ~ 60 dB	0.11 dB 0.13 dB 0.15 dB 0.17 dB 0.21 dB 0.30 dB  0.04 dB 0.04 dB 0.05 dB 0.05 dB 0.06 dB 0.06 dB 0.07 dB 0.08 dB 0.08 dB 0.09 dB 0.09 dB 0.10 dB  0.21 dB 0.23 dB 0.29 dB 0.30 dB 0.47 dB 1.2 dB	Network analyzers, Calibration kits / HCT-CS-110-40610

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Waveguide standard mismatches Coupling factor	40611	(40 GHz ~ 75 GHz) 0 dB ~ 10 dB 10 dB ~ 20 dB  20 dB ~ 30 dB 30 dB ~ 40 dB 40 dB ~ 50 dB 50 dB ~ 60 dB  (75 GHz ~ 110 GHz) 0 dB ~ 10 dB 10 dB ~ 20 dB 20 dB ~ 30 dB 30 dB ~ 40 dB 40 dB ~ 50 dB 50 dB ~ 60 dB	0.34 dB 0.34 dB  0.34 dB 0.34 dB 0.34 dB 0.36 dB  0.34 dB 0.34 dB 0.34 dB 0.34 dB 0.34 dB 0.36 dB	Network analyzers, Calibration kits / HCT-CS-349-40611
Electrostatic discharge generators Current(I <sub>p</sub> )  T1 30 ns, 65 ns  T2 60 ns, 130 ns  T3 180 ns, 400 ns	40613	(±) (6.7 ~ 7.5) A (7.5 ~ 15) A (15 ~ 22.5) A (22.5 ~ 30) A (30 ~ 45.0) A (45.0 ~ 56.3) A (56.3 ~ 75) A (75 ~ 93.8) A (93.8 ~ 150) A  (±) (2 ~ 4) A (4 ~ 8) A (8 ~ 12) A (12 ~ 16) A (16 ~ 24) A (24 ~ 30) A (30 ~ 40) A (40 ~ 50) A (50 ~ 80) A  (±) (1 ~ 2) A (2 ~ 4) A (4 ~ 6) A (6 ~ 8) A (8 ~ 12) A (12 ~ 15) A (15 ~ 20) A (20 ~ 25) A (25 ~ 40) A  (±) (0.3 ~ 0.55) A (0.55 ~ 1.10) A (1.10 ~ 1.65) A (1.65 ~ 2.20) A (2.20 ~ 3.30) A (3.30 ~ 4.13) A (4.13 ~ 5.50) A (5.50 ~ 6.88) A (6.88 ~ 10.6) A	$3.0 \times 10^{-2}$ $2.1 \times 10^{-2}$ $2.7 \times 10^{-2}$ $2.3 \times 10^{-2}$ $2.7 \times 10^{-2}$ $2.5 \times 10^{-2}$ $2.3 \times 10^{-2}$ $2.2 \times 10^{-2}$ $2.5 \times 10^{-2}$  $4.9 \times 10^{-2}$ $2.9 \times 10^{-2}$ $4.2 \times 10^{-2}$ $3.4 \times 10^{-2}$ $4.3 \times 10^{-2}$ $3.6 \times 10^{-2}$ $3.0 \times 10^{-2}$ $2.7 \times 10^{-2}$ $3.5 \times 10^{-2}$  $9.4 \times 10^{-2}$ $4.9 \times 10^{-2}$ $7.9 \times 10^{-2}$ $6.0 \times 10^{-2}$ $8.1 \times 10^{-2}$ $6.4 \times 10^{-2}$ $5.0 \times 10^{-2}$ $4.1 \times 10^{-2}$ $6.2 \times 10^{-2}$  $6.1 \times 10^{-2}$ $3.4 \times 10^{-2}$ $5.5 \times 10^{-2}$ $4.3 \times 10^{-2}$ $7.8 \times 10^{-2}$ $6.8 \times 10^{-2}$ $5.3 \times 10^{-2}$ $4.4 \times 10^{-2}$ $4.5 \times 10^{-2}$	Electrostatic discharge measurement system, Oscilloscope calibrators / HCT-CS-111-40613

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
T4 360 ns, 800 ns	40613	(±) (0.1 ~ 0.30) A (0.30 ~ 0.60) A	$1.1 \times 10^{-1}$ $5.7 \times 10^{-2}$	
Rise/Fall Time		(0.60 ~ 0.90) A (0.90 ~ 1.20) A (1.20 ~ 1.80) A (1.80 ~ 2.25) A (2.25 ~ 3.00) A (3.00 ~ 3.75) A (3.75 ~ 5.90) A	$9.7 \times 10^{-2}$ $7.4 \times 10^{-2}$ $1.4 \times 10^{-1}$ $1.2 \times 10^{-1}$ $9.2 \times 10^{-2}$ $7.4 \times 10^{-2}$ $7.6 \times 10^{-2}$	
Voltage		(0.5 ~ 1) ns	$3.7 \times 10^{-2}$	
		(±) (0.1 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 2) kV (2 ~ 4) kV (4 ~ 6) kV (6 ~ 8) kV (8 ~ 10) kV (10 ~ 12) kV (12 ~ 14) kV (14 ~ 16) kV (16 ~ 18) kV (18 ~ 20) kV (20 ~ 25) kV (25 ~ 30) kV	$1.5 \times 10^{-2}$ $9.0 \times 10^{-3}$ $2.4 \times 10^{-2}$ $1.2 \times 10^{-2}$ $1.1 \times 10^{-2}$ $8.6 \times 10^{-3}$ $7.8 \times 10^{-3}$ $8.1 \times 10^{-3}$ $8.4 \times 10^{-3}$ $7.4 \times 10^{-3}$ $7.1 \times 10^{-3}$ $6.9 \times 10^{-3}$ $7.9 \times 10^{-3}$ $6.6 \times 10^{-3}$	
Semiconductor ESD Peak Current (HBM)		(±) (0.15 ~ 0.17) A (0.17 ~ 0.33) A (0.33 ~ 1.33) A (1.33 ~ 5.23) A	$4.9 \times 10^{-2}$ $3.9 \times 10^{-2}$ $3.6 \times 10^{-2}$ $3.5 \times 10^{-2}$	
Semiconductor ESD Peak Current (MM)		(±) (1.5 ~ 7.0) A (7 ~ 16) A	$3.7 \times 10^{-2}$ $3.6 \times 10^{-2}$	
Semiconductor ESD Rise/Fall Time		(1 ~ 11) ns	$1.8 \times 10^{-2}$	
Semiconductor ESD Decay Time		(100 ~ 200) ns	$3.5 \times 10^{-3}$	
Semi ESD Peak Voltage		(±) (0.01 ~ 1) kV (1 ~ 2) kV (2 ~ 4) kV (4 ~ 8) kV	$3.5 \times 10^{-2}$ $6.1 \times 10^{-2}$ $4.4 \times 10^{-2}$ $3.8 \times 10^{-2}$	
EMC receivers	40614	80 kHz ~ 100 MHz	$5.8 \times 10^{-11}$	Calibration pulse generators, Frequency standards, Power sensors, Standard attenuators, RF signal generators, Network analyzers / HCT-CS-112-40614
Reference frequency		9 kHz ~ 1 GHz 1 GHz ~ 3 GHz 3 GHz ~ 20 GHz 20 GHz ~ 40 GHz 40 GHz ~ 50 GHz	$3.8 \times 10^{-3}$ $5.3 \times 10^{-3}$ $9.3 \times 10^{-3}$ $1.2 \times 10^{-2}$ $5.9 \times 10^{-2}$	
Input impedance (Reflection coefficient)				



406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Sinewave voltage accuracy	40614	10 Hz ~ 2 GHz	0.04 dB	
		2 GHz ~ 12 GHz	0.06 dB	
		12 GHz ~ 40 GHz	0.08 dB	
		40 GHz ~ 50 GHz	0.25 dB	
Pulse response		9 kHz ~ 40 GHz	0.25 dB	
Repetition frequency response		9 kHz ~ 1 GHz	0.10 dB	
Overall selectivity		9 kHz ~ 40 GHz	0.08 dB	
IF rejection ratio		9 kHz ~ 40 GHz	0.31 dB	
Image frequency response		9 kHz ~ 40 GHz	0.31 dB	
Other spurious response	9 kHz ~ 40 GHz	0.31 dB		
Random noise	9 kHz ~ 40 GHz	0.07 dB		
Resolution bandwidth		10 Hz ~ 20 MHz	$7.4 \times 10^{-4}$	
RF filters	40615	Cutoff frequency	9 kHz ~ 26.5 GHz	Network analyzers, Calibration kits / HCT-CS-113-40615
		(9 kHz ~ 1 GHz)	$6.4 \times 10^{-7}$	
Insert loss		0 dB ~ 10 dB	0.11 dB	
		10 dB ~ 20 dB	0.12 dB	
		20 dB ~ 30 dB	0.14 dB	
		30 dB ~ 40 dB	0.17 dB	
		40 dB ~ 50 dB	0.21 dB	
		50 dB ~ 60 dB	0.30 dB	
		60 dB ~ 70 dB	0.54 dB	
		70 dB ~ 80 dB	1.3 dB	
		80 dB ~ 100 dB	3.3 dB	
		(1 GHz ~ 18 GHz)		
		0 dB ~ 10 dB	0.11 dB	
		10 dB ~ 20 dB	0.12 dB	
		20 dB ~ 30 dB	0.13 dB	
		30 dB ~ 40 dB	0.15 dB	
		40 dB ~ 50 dB	0.20 dB	
		50 dB ~ 60 dB	0.34 dB	
		60 dB ~ 70 dB	0.72 dB	
		70 dB ~ 80 dB	1.9 dB	
		80 dB ~ 100 dB	4.7 dB	
		(18 GHz ~ 26.5 GHz)		
		0 dB ~ 10 dB	0.21 dB	
		10 dB ~ 20 dB	0.23 dB	
		20 dB ~ 30 dB	0.24 dB	
		30 dB ~ 40 dB	0.27 dB	
		40 dB ~ 50 dB	0.35 dB	
		50 dB ~ 60 dB	0.59 dB	
		60 dB ~ 70 dB	1.3 dB	
		70 dB ~ 80 dB	3.2 dB	
		80 dB ~ 100 dB	7.6 dB	
		(26.5 GHz ~ 40 GHz)		
		0 dB ~ 10 dB	0.21 dB	
		10 dB ~ 20 dB	0.23 dB	
		20 dB ~ 30 dB	0.24 dB	
		30 dB ~ 40 dB	0.29 dB	
	40 dB ~ 50 dB	0.47 dB		
	50 dB ~ 60 dB	1.2 dB		
	60 dB ~ 70 dB	3.1 dB		



406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Mobile communication test sets	40621			Frequency standards, Power sensors, Measuring receivers, RF spectrum analyzers /HCT-CS-115-40621
Output frequency		1 mHz ~ 46 GHz	$5.8 \times 10^{-11}$	
Output level		(-30 dBm ~ 20 dBm)		
		9 kHz ~ 100 MHz	0.05 dB	
		100 MHz ~ 1 GHz	0.07 dB	
		1 GHz ~ 8 GHz	0.08 dB	
		8 GHz ~ 12 GHz	0.09 dB	
		12 GHz ~ 26 GHz	0.12 dB	
		26 GHz ~ 40 GHz	0.15 dB	
		40 GHz ~ 50 GHz	0.21 dB	
Absolute TRFL accuracy		(9 kHz ~ 8 GHz)		
		0 dBm ~ 30 dBm	0.15 dB	
		-40 dBm ~ 0 dBm	0.16 dB	
		-80 dBm ~ -40 dBm	0.18 dB	
		-120 dBm ~ -80 dBm	0.20 dB	
		-140 dBm ~ -120 dBm	0.21 dB	
		(8 GHz ~ 18 GHz)		
		0 dBm ~ 30 dBm	0.20 dB	
		-40 dBm ~ 0 dBm	0.20 dB	
		-80 dBm ~ -40 dBm	0.22 dB	
	-120 dBm ~ -80 dBm	0.24 dB		
	-140 dBm ~ -120 dBm	0.25 dB		
	(18 GHz ~ 26.5 GHz)			
	0 dBm ~ 30 dBm	0.27 dB		
	-40 dBm ~ 0 dBm	0.27 dB		
	-80 dBm ~ -40 dBm	0.29 dB		
	-120 dBm ~ -80 dBm	0.31 dB		
	-140 dBm ~ -120 dBm	0.32 dB		
Relative TRFL accuracy	(9 kHz ~ 18 GHz)			
	0 dBm ~ 30 dBm	0.05 dB		
	-40 dBm ~ 0 dBm	0.05 dB		
	-80 dBm ~ -40 dBm	0.08 dB		
	-120 dBm ~ -80 dBm	0.09 dB		
	-140 dBm ~ -120 dBm	0.10 dB		
	(18 GHz ~ 26.5 GHz)			
	0 dBm ~ 30 dBm	0.05 dB		
	-40 dBm ~ 0 dBm	0.05 dB		
	-80 dBm ~ -40 dBm	0.08 dB		
	-120 dBm ~ -80 dBm	0.09 dB		
	-140 dBm ~ -120 dBm	0.11 dB		
Output amplitude modulation	(CW 100 kHz ~ 26.5 GHz, Rate 10 Hz ~ 100 kHz)			
	(1 ~ 100) %	$1.2 \times 10^{-2}$		
Output frequency modulation	(CW 100 kHz ~ 26.5 GHz, Rate 10 Hz ~ 100 kHz)			
	0.1 Hz ~ 5 MHz	$1.2 \times 10^{-2}$		
Output phase modulation	(CW 100 kHz ~ 26.5 GHz, Rate 10 Hz ~ 100 kHz)			
	(0.1 ~ 10) krad	$1.2 \times 10^{-2}$		
Output phase distortion	100 kHz ~ 26.5 GHz	$3.0 \times 10^{-2}$		
Output harmonics	9 kHz ~ 10 GHz	1.4 dB		
	10 GHz ~ 26.5 GHz	1.7 dB		

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Output AC Voltage	40621	(10 Hz ~ 25 kHz) 10 mV ~ 100 V	$7.4 \times 10^{-4}$		
Output DC voltage		10 mV ~ 100 V	$5.8 \times 10^{-5}$		
Input frequency		1 mHz ~ 18 GHz	$5.8 \times 10^{-11}$		
Input voltage		(-120 dBm ~ 20 dBm)			
		9 kHz ~ 100 MHz	0.05 dB		
		100 MHz ~ 1 GHz	0.07 dB		
		1 GHz ~ 8 GHz	0.08 dB		
		8 GHz ~ 12 GHz	0.09 dB		
		12 GHz ~ 18 GHz	0.12 dB		
		18 GHz ~ 50 GHz	0.12 dB		
Input level linearity		(9 kHz ~ 26.5 GHz)			
		-10 dBm ~ 30 dBm	0.034 dB		
		-20 dBm ~ -10 dBm	0.040 dB		
		-30 dBm ~ -20 dBm	0.046 dB		
		-40 dBm ~ -30 dBm	0.052 dB		
		-50 dBm ~ -40 dBm	0.058 dB		
		-60 dBm ~ -50 dBm	0.064 dB		
		-70 dBm ~ -60 dBm	0.070 dB		
		-80 dBm ~ -70 dBm	0.076 dB		
	-90 dBm ~ -80 dBm	0.080 dB			
	-100 dBm ~ -90 dBm	0.086 dB			
Input amplitude modulation	100 kHz ~ 26.5 GHz	$1.2 \times 10^{-2}$			
	Input frequency modulation	100 kHz ~ 26.5 GHz	$1.2 \times 10^{-2}$		
		Input phase modulation	100 kHz ~ 26.5 GHz	$1.2 \times 10^{-2}$	
Input modulation distortion	100 kHz ~ 26.5 GHz	$3.0 \times 10^{-2}$			
Input harmonics	9 kHz ~ 10 GHz	1.4 dB			
	10 GHz ~ 18 GHz	1.7 dB			
Input AC voltage	(10 Hz ~ 25 kHz) 10 mV ~ 100 V	$7.4 \times 10^{-4}$			
Input DC voltage	10 mV ~ 100 V	$7.3 \times 10^{-5}$			
Modulation meters	40622	Frequency	1 mHz ~ 26.5 GHz	$5.8 \times 10^{-11}$	Measuring receivers, AM/FM test source / HCT-CS-116-40622
Amplitude Modulation		(CW 100 kHz ~ 26.5 GHz, Rate 10 Hz ~ 100 kHz) (1 ~ 100) %	$1.2 \times 10^{-2}$		
Frequency Modulation		(CW 100 kHz ~ 26.5 GHz, Rate 10 Hz ~ 100 kHz) 0.1 Hz ~ 5 MHz	$1.2 \times 10^{-2}$		
Phase Modulation		(CW 100 kHz ~ 26.5 GHz, Rate 10 Hz ~ 100 kHz) (0.1 ~ 10) krad	$1.2 \times 10^{-2}$		
Audio RMS Accuracy		(20 Hz ~ 50 kHz) 100 mV ~ 5 V	$1.2 \times 10^{-3}$		

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Reference Power	40622	(50 MHz)		
		1 mW	$8.0 \times 10^{-3}$	
Zero Set		0.000 $\mu$ W	0.001 $\mu$ W	
		0.00 $\mu$ W	0.01 $\mu$ W	
		0.0 $\mu$ W	0.1 $\mu$ W	
		0.000 mW	0.001 mW	
		0.00 mW	0.01 mW	
Range-to-Range Error		10 $\mu$ W ~ 100 mW	$1.3 \times 10^{-3}$	
Tuned RF Level		(0 ~ 10) dB	0.027 dB	
		(10 ~ 20) dB	0.029 dB	
		(20 ~ 30) dB	0.032 dB	
		(30 ~ 40) dB	0.038 dB	
		(40 ~ 50) dB	0.043 dB	
	(50 ~ 60) dB	0.043 dB		
	(60 ~ 70) dB	0.048 dB		
	(70 ~ 80) dB	0.054 dB		
	(80 ~ 90) dB	0.060 dB		
	(90 ~ 100) dB	0.066 dB		
	(100 ~ 110) dB	0.069 dB		
	(110 ~ 120) dB	0.074 dB		
Network analyzers	40623			Calibration kit Frequency standards, Standard attenuators, Power sensors, Standard mismatches / HCT-CS-117-40623
Output frequency		1 mHz ~ 46 GHz	$5.8 \times 10^{-11}$	
Output level accuracy		(-30 dBm ~ 20 dBm)		
		5 Hz ~ 100 MHz	0.06 dB	
		100 MHz ~ 1 GHz	0.07 dB	
		1 GHz ~ 8 GHz	0.08 dB	
		8 GHz ~ 12 GHz	0.09 dB	
		12 GHz ~ 18 GHz	0.12 dB	
		18 GHz ~ 26 GHz	0.12 dB	
		26 GHz ~ 33 GHz	0.14 dB	
		33 GHz ~ 40 GHz	0.15 dB	
		40 GHz ~ 50 GHz	0.15 dB	
		50 GHz ~ 75 GHz	0.24 dB	
		75 GHz ~ 110 GHz	0.28 dB	
Absolute TRFL accuracy		(9 kHz ~ 8 GHz)		
		0 dBm ~ 30 dBm	0.15 dB	
		-40 dBm ~ 0 dBm	0.16 dB	
		-80 dBm ~ -40 dBm	0.18 dB	
		-120 dBm ~ -80 dBm	0.20 dB	
		-140 dBm ~ -120 dBm	0.21 dB	
		(8 GHz ~ 18 GHz)		
		0 dBm ~ 30 dBm	0.20 dB	
		-40 dBm ~ 0 dBm	0.20 dB	
		-80 dBm ~ -40 dBm	0.22 dB	
		-120 dBm ~ -80 dBm	0.24 dB	
		-140 dBm ~ -120 dBm	0.25 dB	
		(18 GHz ~ 26.5 GHz)		
		0 dBm ~ 30 dBm	0.27 dB	
		-40 dBm ~ 0 dBm	0.27 dB	
		-80 dBm ~ -40 dBm	0.29 dB	
		-120 dBm ~ -80 dBm	0.31 dB	
		-140 dBm ~ -120 dBm	0.32 dB	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Absolute TRFL accuracy	40623	(26.5 GHz ~ 40 GHz) -30 dBm ~ 20 dBm	0.27 dB	
Output level linearity		(40 GHz ~ 50 GHz) -30 dBm ~ 20 dBm	0.31 dB	
		(9 kHz ~ 26.5 GHz) 0 dBm ~ 10 dBm	0.034 dB	
		-10 dBm ~ 0 dBm	0.034 dB	
		-20 dBm ~ -10 dBm	0.040 dB	
		-30 dBm ~ -20 dBm	0.046 dB	
		-40 dBm ~ -30 dBm	0.052 dB	
		-50 dBm ~ -40 dBm	0.058 dB	
		-60 dBm ~ -50 dBm	0.064 dB	
		-70 dBm ~ -60 dBm	0.070 dB	
		-80 dBm ~ -70 dBm	0.076 dB	
-90 dBm ~ -80 dBm		0.080 dB		
-100 dBm ~ -90 dBm		0.086 dB		
-110 dBm ~ -100 dBm		0.092 dB		
-120 dBm ~ -110 dBm	0.098 dB			
Harmonics	(26.5 GHz ~ 40 GHz) -30 dBm ~ 20 dBm	0.024 dB		
	(40 GHz ~ 50 GHz) -30 dBm ~ 20 dBm	0.050 dB		
Magnitude dynamic accuracy	20 Hz ~ 20 GHz 20 GHz ~ 40 GHz	1.4 dB 1.7 dB		
Mismatch measurement accuracy	0 dB ~ 120 dB	0.029 dB		
Input impedance	9 kHz ~ 1 GHz 1 GHz ~ 18 GHz	$4.8 \times 10^{-3}$ $1.0 \times 10^{-2}$		
	9 kHz ~ 1 GHz 1 GHz ~ 18 GHz	$4.8 \times 10^{-3}$ $1.0 \times 10^{-2}$		
Noise figure meters	40624			Noise standards RF signal generators, Noise sources / HCT-CS-118-40624
Output frequency	1 mHz ~ 18 GHz	$5.8 \times 10^{-11}$		
Input impedance	9 kHz ~ 1 GHz	$0.9 \times 10^{-2}$		
	1 GHz ~ 3 GHz	$1.2 \times 10^{-2}$		
	3 GHz ~ 18 GHz	$1.9 \times 10^{-2}$		
Output DC voltage	0 V 0.1 V ~ 30 V	10 $\mu$ V $1.1 \times 10^{-6}$		
	10 MHz ~ 18 GHz	0.35 dB		
Noise generators	40625			RF spectrum generators / HCT-CS-177-40625
Output frequency	1 mHz ~ 18 GHz	$5.8 \times 10^{-11}$		
Output level	(-120 dBm ~ 30 dBm) 9 kHz ~ 3 GHz	0.51 dB		
	3 GHz ~ 6.6 GHz	1.8 dB		
	6.6 GHz ~ 18 GHz	2.4 dB		

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Noise impulse simulators Output Voltage	40626	(±) 10 V (10 ~ 20) V (20 ~ 50) V (50 ~ 200) V (200 ~ 250) V (250 ~ 500) V (0.5 ~ 1) kV (1 ~ 2) kV (2 ~ 2.5) kV (2.5 ~ 3) kV (3 ~ 4) kV	0.39 V $3.8 \times 10^{-2}$ $3.3 \times 10^{-2}$ $3.8 \times 10^{-2}$ $3.4 \times 10^{-2}$ $3.3 \times 10^{-2}$ $3.8 \times 10^{-2}$ $3.1 \times 10^{-2}$ $2.8 \times 10^{-2}$ $2.7 \times 10^{-2}$ $3.8 \times 10^{-2}$	High voltage probes Oscilloscopes /HCT-CS-119-40626
Delta time measurement (rise/fall/duration/period/ repetition rate/burst duration)		0.1 ns (0.1 ~ 1.0) ns (1.0 ~ 2.0) ns (2.0 ~ 5.0) ns (5.0 ~ 10.0) ns (10 ~ 20) ns (20 ~ 50) ns (50 ~ 100) ns (100 ~ 200) ns (200 ~ 500) ns (0.5 ~ 1.0) μs (1.0 ~ 2.0) μs (2.0 ~ 5.0) μs (5.0 ~ 10.0) μs (10 ~ 20) μs (20 ~ 50) μs (50 ~ 100) μs (100 ~ 200) μs (200 ~ 500) μs (0.5 ~ 1) ms (1 ~ 2) ms (2 ~ 5) ms (5 ~ 10) ms (10 ~ 20) ms (20 ~ 50) ms (50 ~ 100) ms (100 ~ 200) ms (200 ~ 500) ms (0.5 ~ 1.0) s (1.0 ~ 2.0) s (2.0 ~ 5.0) s	0.014 ns $1.4 \times 10^{-2}$ $7.0 \times 10^{-3}$ $2.8 \times 10^{-3}$ $1.6 \times 10^{-3}$ $8.0 \times 10^{-4}$ $3.2 \times 10^{-4}$ $6.0 \times 10^{-4}$ $7.0 \times 10^{-4}$ $2.8 \times 10^{-4}$ $1.2 \times 10^{-3}$ $5.8 \times 10^{-4}$ $2.3 \times 10^{-4}$ $5.9 \times 10^{-4}$ $3.1 \times 10^{-4}$ $1.3 \times 10^{-4}$ $8.4 \times 10^{-4}$ $4.2 \times 10^{-4}$ $6.1 \times 10^{-4}$ $2.8 \times 10^{-2}$ $3.5 \times 10^{-2}$ $2.3 \times 10^{-4}$ $5.9 \times 10^{-4}$ $3.0 \times 10^{-4}$ $1.6 \times 10^{-4}$ $5.8 \times 10^{-4}$ $2.9 \times 10^{-4}$ $1.2 \times 10^{-4}$ $2.1 \times 10^{-3}$ $1.0 \times 10^{-3}$ $4.2 \times 10^{-4}$	
RF power meters RF power meters Output frequency Output levels Instrument accuracy Input level accuracy	40635	1 MHz ~ 18 GHz  (10 MHz ~ 300 MHz) 1 μW ~ 100 mW  3 μW ~ 100 mW  (9 kHz ~ 18 GHz) -80 dBm ~ 20 dBm	$5.8 \times 10^{-11}$   $5.1 \times 10^{-3}$  $4.4 \times 10^{-3}$  0.15 dB	Range calibrators, Power sensors /HCT-CS-120-40635

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Input voltage  RF high power meters Calibration factor		(DC) 0 V ~ 400 V  (10 kHz ~ 220 MHz) 0.01 W ~ 2.5 kW  (200 MHz ~ 1 GHz) 0.01 W ~ 100 W  (1 GHz ~ 4.2 GHz) 0.01 W ~ 10 W	$5.8 \times 10^{-5}$  $1.5 \times 10^{-2}$  $2.9 \times 10^{-2}$  $3.5 \times 10^{-2}$	RF calorimeters /HCT-CS-162-40635
Diode power sensors Calibration factor	40636	(1 $\mu$ W ~ 100 mW) 9 kHz ~ 1 GHz 1 GHz ~ 10 GHz 10 GHz ~ 18 GHz 18 GHz ~ 26 GHz 26 GHz ~ 34 GHz 34 GHz ~ 38 GHz 38 GHz ~ 43 GHz 43 GHz ~ 50 GHz  (0 ~ 1) 9 kHz ~ 1 GHz 1 GHz ~ 3 GHz 3 GHz ~ 20 GHz 20 GHz ~ 40 GHz 40 GHz ~ 50 GHz	$1.5 \times 10^{-2}$ $1.6 \times 10^{-2}$ $2.1 \times 10^{-2}$ $2.1 \times 10^{-2}$ $2.5 \times 10^{-2}$ $3.0 \times 10^{-2}$ $3.3 \times 10^{-2}$ $3.6 \times 10^{-2}$  $3.8 \times 10^{-3}$ $5.3 \times 10^{-3}$ $9.3 \times 10^{-3}$ $1.2 \times 10^{-2}$ $5.9 \times 10^{-2}$	Coaxial thermistor mounts Power sensors /HCT-CS-121-40636



406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Thermocouple power sensors Calibration factor	40637	(1 μW ~ 100 mW)		Coaxial thermistor mounts Power sensors /HCT-CS-122-40637
		9 kHz ~ 1 GHz	$1.5 \times 10^{-2}$	
		1 GHz ~ 10 GHz	$1.6 \times 10^{-2}$	
		10 GHz ~ 18 GHz	$2.1 \times 10^{-2}$	
		18 GHz ~ 26 GHz	$2.1 \times 10^{-2}$	
		26 GHz ~ 34 GHz	$2.5 \times 10^{-2}$	
		34 GHz ~ 38 GHz	$3.0 \times 10^{-2}$	
		38 GHz ~ 43 GHz	$3.3 \times 10^{-2}$	
		43 GHz ~ 50 GHz	$3.6 \times 10^{-2}$	
Reflection coefficient		(0 ~ 1)		
		9 kHz ~ 1 GHz	$3.8 \times 10^{-3}$	
		1 GHz ~ 3 GHz	$5.3 \times 10^{-3}$	
		3 GHz ~ 20 GHz	$9.3 \times 10^{-3}$	
		20 GHz ~ 40 GHz	$1.2 \times 10^{-2}$	
	40 GHz ~ 50 GHz	$5.9 \times 10^{-2}$		
Pulse generators	40638			Frequency cutters, Oscilloscopes /HCT-CS-123-40646
Frequency		1 Hz ~ 10 GHz	$6.1 \times 10^{-9}$	
Period		300 ps ~ 1 s	$6.1 \times 10^{-9}$	
Delay		1 s ~ 100 ns	$1.2 \times 10^{-3}$	
		(100 ~ 10) ns	$1.3 \times 10^{-3}$	
		(10 ~ 1) ns	$5.9 \times 10^{-3}$	
Double Pulse		1 s ~ 100 ns	$1.2 \times 10^{-3}$	
		(100 ~ 10) ns	$1.3 \times 10^{-3}$	
		(10 ~ 1) ns	$5.9 \times 10^{-3}$	
Width		1 s ~ 100 ns	$1.2 \times 10^{-3}$	
		(100 ~ 10) ns	$1.3 \times 10^{-3}$	
		(10 ~ 1) ns	$5.9 \times 10^{-3}$	
Transition Time		1 s ~ 100 ns	$1.2 \times 10^{-3}$	
		(100 ~ 10) ns	$1.3 \times 10^{-3}$	
		(10 ~ 1) ns	$5.9 \times 10^{-3}$	
DC Level		10 mV	6.2 μV	
		10 mV ~ 100 V	$6.2 \times 10^{-4}$	
Output Level		10 mV		
		20 Hz ~ 1 kHz	9.4 μV	
		(1 ~ 20) kHz	9.4 μV	
	(20 ~ 100) kHz	9.4 μV		
	(10 ~ 100) mV			
	20 Hz ~ 1 kHz	$6.4 \times 10^{-4}$		
	(1 ~ 20) kHz	$7.6 \times 10^{-4}$		
	(20 ~ 100) kHz	$1.1 \times 10^{-3}$		
	(100 mV ~ 1 V)			
	20 Hz ~ 20 kHz	$6.4 \times 10^{-4}$		
	(20 ~ 50) kHz	$6.7 \times 10^{-4}$		
	(50 ~ 100) kHz	$6.7 \times 10^{-4}$		
(1 ~ 10) V				
20 Hz ~ 20 kHz	$6.4 \times 10^{-4}$			
(20 ~ 50) kHz	$6.7 \times 10^{-4}$			
(50 ~ 100) kHz	$6.7 \times 10^{-4}$			

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Output Level	40638	(10 ~ 100) V 20 Hz ~ 20 kHz (20 ~ 50) kHz (50 ~ 100) kHz  (100 ~ 300) V 20 Hz ~ 1 kHz	$6.4 \times 10^{-4}$ $6.7 \times 10^{-4}$ $6.7 \times 10^{-4}$  $3.1 \times 10^{-4}$	
Radar test sets Output frequency Output level Harmonics Output modulation signal level Output amplitude modulation Output frequency modulation Output modulation distortion Phase DDM SDM VOR Purse width Input frequency Input level	40639	1 mHz ~ 18 GHz  (9 kHz ~ 18 GHz) 10 dBm ~ 30 dBm -30 dBm ~ 10 dBm -60 dBm ~ -30 dBm -100 dBm ~ -60 dBm -120 dBm ~ -100 dBm  (9 kHz ~ 5 GHz) -100 dBc ~ 0 dBc (5 GHz ~ 18 GHz) -100 dBc ~ 0 dBc  (9 kHz ~ 18 GHz) -100 dBc ~ 0 dBc  (CW 9 kHz ~ 18 GHz, Rate 10 Hz ~ 100 kHz) 0 % ~ 100 %  (CW 9 kHz ~ 18 GHz, Rate 10 Hz ~ 100 kHz) 0 kHz ~ 800 kHz  (9 kHz ~ 18 GHz) 0 % ~ 100 %  (9 kHz ~ 18 GHz) 0 ° ~ 360 °  100 kHz ~ 1.36 GHz  100 kHz ~ 1.36 GHz  100 kHz ~ 1.36 GHz  1 ns ~ 10 ms  9 kHz ~ 18 GHz  (100 kHz ~ 1.36 GHz) 1 mW ~ 100 W	$5.8 \times 10^{-11}$  0.12 dB 0.12 dB 0.13 dB 0.15 dB 0.30 dB  1.2 dB 1.5 dB  1.3 dB  $1.7 \times 10^{-2}$  $1.2 \times 10^{-2}$  $1.2 \times 10^{-2}$  $1.2 \times 10^{-2}$  $3.0 \times 10^{-2}$ $3.0 \times 10^{-2}$ $3.0 \times 10^{-2}$  $2.3 \times 10^{-2}$  $5.8 \times 10^{-10}$  $1.9 \times 10^{-2}$	VOR/ILS signal calibrators, Frequency standards, Power sensors / HCT-CS-168-40639(RADAR) / HCT-CS-204-40639(SART) / HCT-CS-207-40639(AIS) / HCT-CS-209-40639(GMDSS) / HCT-CS-214-40639(EPIRB)
RF signal generators Output frequency Absolute output level	40640	1 mHz ~ 46 GHz  (-30 dBm ~ 20 dBm) 5 Hz ~ 100 MHz 100 MHz ~ 1 GHz 1 GHz ~ 8 GHz 8 GHz ~ 12 GHz 12 GHz ~ 18 GHz 18 GHz ~ 26 GHz	$5.8 \times 10^{-11}$  0.05 dB 0.07 dB 0.08 dB 0.09 dB 0.12 dB 0.11 dB	Measuring receivers, Power sensors, Frequency standards, RF spectrum analyzers / HCT-CS-124-40640

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Absolute output level	40640	26 GHz ~ 33 GHz	0.13 dB		
		33 GHz ~ 40 GHz	0.14 dB		
		40 GHz ~ 50 GHz	0.16 dB		
		50 GHz ~ 75 GHz	0.21 dB		
		75 GHz ~ 110 GHz	0.28 dB		
Absolute TRFL accuracy		(9 kHz ~ 8 GHz)			
		0 dBm ~ 30 dBm	0.15 dB		
		-40 dBm ~ 0 dBm	0.16 dB		
		-80 dBm ~ -40 dBm	0.18 dB		
		-120 dBm ~ -80 dBm	0.20 dB		
		-140 dBm ~ -120 dBm	0.21 dB		
		(8 GHz ~ 18 GHz)			
		0 dBm ~ 30 dBm	0.20 dB		
	-40 dBm ~ 0 dBm	0.20 dB			
	-80 dBm ~ -40 dBm	0.22 dB			
	-120 dBm ~ -80 dBm	0.24 dB			
	-140 dBm ~ -120 dBm	0.25 dB			
	(18 GHz ~ 26.5 GHz)				
	0 dBm ~ 30 dBm	0.27 dB			
	-40 dBm ~ 0 dBm	0.27 dB			
	-80 dBm ~ -40 dBm	0.29 dB			
	-120 dBm ~ -80 dBm	0.31 dB			
	-140 dBm ~ -120 dBm	0.32 dB			
Relative TRFL accuracy	(9 kHz ~ 26.5 GHz)				
	0 dBm ~ 30 dBm	0.05 dB			
	-40 dBm ~ 0 dBm	0.05 dB			
	-80 dBm ~ -40 dBm	0.08 dB			
	-120 dBm ~ -80 dBm	0.09 dB			
	-140 dBm ~ -120 dBm	0.11 dB			
Output amplitude modulation	(CW 100 kHz ~ 26.5 GHz, Rate 10 Hz ~ 100 kHz)				
	(1 ~ 100) %		$1.2 \times 10^{-2}$		
Output frequency modulation	(CW 100 kHz ~ 26.5 GHz, Rate 10 Hz ~ 100 kHz)				
	0.1 Hz ~ 5 MHz		$1.2 \times 10^{-2}$		
Output phase modulation	(CW 100 kHz ~ 26.5 GHz, Rate 10 Hz ~ 100 kHz)				
	(0.1 ~ 10) krad		$1.2 \times 10^{-2}$		
Output modulation distortion	100 kHz ~ 26.5 GHz		$3.0 \times 10^{-2}$		
Harmonics	20 Hz ~ 20 GHz		1.4 dB		
	20 GHz ~ 40 GHz		1.7 dB		
Pulse modulation	1 μs ~ 1 s		$1.2 \times 10^{-3}$		
RF spectrum analyzers	40641			Power sensors, Frequency standards, RF signal generators, Standard attenuators / HCT-CS-125-40641	
Reference frequency		10 MHz ~ 1 GHz	$5.8 \times 10^{-11}$		
Reference level		(10 MHz ~ 1 GHz) -30 dBm ~ 10 dBm	0.07 dB		
Frequency readout		5 Hz ~ 110 GHz	$9.6 \times 10^{-4} \cdot \text{SPAN}$		
Frequency counter	5 Hz ~ 110 GHz		0.1 Hz		

406. Radio frequency measuremet

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Frequency span	40641	5 Hz ~ 110 GHz	$1.4 \times 10^{-3} \cdot \text{SPAN}$		
Resolution bandwidth		1 Hz ~ 100 MHz	$2.2 \times 10^{-3} \cdot \text{RBW}$		
Resolution bandwidth selectivity		1 Hz ~ 100 MHz	$4.0 \times 10^{-3} \cdot \text{RBW}$		
Resolution bandwidth switching error		1 Hz ~ 100 MHz	0.004 dB		
Input attenuator accuracy		0 dB ~ 100 dB	0.08 dB		
Scale fidelity		0 dB ~ 100 dB	0.08 dB		
Reference level accuracy		0 dB ~ 100 dB	0.06 dB		
Frequency response		5 Hz ~ 4 GHz	0.09 dB		
		4 GHz ~ 18 GHz	0.15 dB		
		18 GHz ~ 26.5 GHz	0.19 dB		
		26.5 GHz ~ 40 GHz	0.21 dB		
		40 GHz ~ 110 GHz	0.35 dB		
Average noise level		5 Hz ~ 3 GHz	0.58 dB		
		3 GHz ~ 12 GHz	1.0 dB		
	12 GHz ~ 18 GHz	1.4 dB			
	18 GHz ~ 40 GHz	1.7 dB			
	40 GHz ~ 50 GHz	2.0 dB			
Sideband noise level	-30 kHz ~ 30 kHz	1.7 dB			
Input level	(1 kHz ~ 100 kHz) -60 dBV ~ 30 dBV	0.18 dB			
Conversion factor	18 GHz ~ 110 GHz	0.82 dB			
RF speed guns	40642	(5 ~ 2 000) m/s	0.03 m/s	Frequency standards / HCT-CS-278-40642	
Speed					
Surge generators	40643	(±)	0.11 V	High voltage probes /HCT-CS-126-40643	
Voltage output			2 V		$1.1 \times 10^{-2}$
			(2 ~ 10) V		$7.6 \times 10^{-3}$
			(10 ~ 20) V		$4.4 \times 10^{-3}$
			(20 ~ 50) V		$4.2 \times 10^{-3}$
			(50 ~ 100) V		$4.7 \times 10^{-3}$
			(100 ~ 200) V		$1.7 \times 10^{-3}$
			(200 ~ 500) V		$4.0 \times 10^{-3}$
			(500 ~ 1000 V)		$5.0 \times 10^{-2}$
			(1 ~ 2) kV		$2.8 \times 10^{-2}$
			(2 ~ 4) kV		$2.0 \times 10^{-2}$
			(4 ~ 6) kV		$1.6 \times 10^{-2}$
			(6 ~ 8) kV		$1.4 \times 10^{-2}$
			(8 ~ 10) kV		$1.2 \times 10^{-2}$
			(10 ~ 12) kV		$1.1 \times 10^{-2}$
			(12 ~ 15) kV		$9.8 \times 10^{-3}$
			(15 ~ 18) kV		$8.8 \times 10^{-3}$
			(18 ~ 20) kV		
			Current output		1 A
(1 ~ 2) A					$1.4 \times 10^{-2}$
(2 ~ 5) A					$7.2 \times 10^{-3}$
(5 ~ 10) A					$6.2 \times 10^{-3}$
(10 ~ 20) A					$6.0 \times 10^{-3}$
(20 ~ 50) A					$5.5 \times 10^{-3}$
(50 ~ 100) A	$4.2 \times 10^{-3}$				

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Current output	40643	(100 ~ 200) A (200 ~ 500) A (500 ~ 1 000) A (1 000 ~ 2 000) A (2 000 ~ 3 000) A (3 000 ~ 5 000) A (5 000 ~ 7 000) A (7 000 ~ 10 000) A (10 000 ~ 20 000) A (20 000 ~ 50 000) A (50 000 ~ 100 000) A	$6.0 \times 10^{-3}$ $5.5 \times 10^{-3}$ $4.2 \times 10^{-4}$ $6.3 \times 10^{-3}$ $9.6 \times 10^{-3}$ $5.7 \times 10^{-3}$ $6.3 \times 10^{-3}$ $4.4 \times 10^{-3}$ $6.1 \times 10^{-3}$ $2.5 \times 10^{-3}$ $1.3 \times 10^{-3}$	
Delta time measurement (rise/fall/duration/period/ repetition rate/burst duration)		0.2 ns (0.2 ~ 1) ns (1 ~ 2) ns (2 ~ 5) ns (5 ~ 10) ns (10 ~ 20) ns (20 ~ 50) ns (50 ~ 100) ns (100 ~ 200) ns (200 ~ 500) ns (0.5 ~ 1) $\mu$ s (1 ~ 2) $\mu$ s (2 ~ 5) $\mu$ s (5 ~ 10) $\mu$ s (1 ~ 20) $\mu$ s (20 ~ 50) $\mu$ s (50 ~ 100) $\mu$ s (100 ~ 200) $\mu$ s (200 ~ 500) $\mu$ s (0.5 ~ 1) ms (1 ~ 2) ms (2 ~ 5) ms (5 ~ 10) ms (10 ~ 20) ms (20 ~ 50) ms (50 ~ 100) ms (100 ~ 200) ms (200 ~ 500) ms (0.5 ~ 1) s (1 ~ 2) s (2 ~ 5) s (5 ~ 10) s	0.015 ns $1.5 \times 10^{-2}$ $7.5 \times 10^{-3}$ $3.0 \times 10^{-3}$ $6.0 \times 10^{-3}$ $3.0 \times 10^{-3}$ $1.2 \times 10^{-3}$ $5.9 \times 10^{-3}$ $2.9 \times 10^{-3}$ $1.2 \times 10^{-3}$ $5.9 \times 10^{-3}$ $2.9 \times 10^{-3}$ $1.2 \times 10^{-3}$ $5.9 \times 10^{-3}$ $2.9 \times 10^{-3}$ $1.2 \times 10^{-3}$ $5.9 \times 10^{-3}$ $2.9 \times 10^{-3}$ $1.2 \times 10^{-2}$ $1.2 \times 10^{-2}$ $5.9 \times 10^{-3}$ $2.9 \times 10^{-3}$ $1.2 \times 10^{-3}$ $5.9 \times 10^{-3}$ $2.9 \times 10^{-3}$ $1.2 \times 10^{-3}$ $5.9 \times 10^{-3}$ $2.9 \times 10^{-3}$ $1.2 \times 10^{-3}$ $5.9 \times 10^{-3}$	
Frequency		0.1 Hz (0.1 ~ 1) Hz 1 Hz ~ 10 MHz	5.8 mHz $5.9 \times 10^{-6}$ $1.2 \times 10^{-6}$	
RF terminations Reflection coefficients	40645	(0 ~ 1) 5 Hz ~ 9 kHz 9 kHz ~ 1 GHz 1 GHz ~ 18 GHz 18 GHz ~ 40 GHz 40 GHz ~ 50 GHz 50 GHz ~ 75 GHz 75 GHz ~ 110 GHz	$4.4 \times 10^{-3}$ $4.8 \times 10^{-3}$ $1.0 \times 10^{-2}$ $1.3 \times 10^{-2}$ $1.4 \times 10^{-2}$ $2.1 \times 10^{-2}$ $2.4 \times 10^{-2}$	Network analyzers, Calibration kits / HCT-CS-128-40645

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Coaxial thermistor mounts Calibration factor	40646	(1 $\mu$ W ~ 100 mW) 10 MHz ~ 1 GHz 1 GHz ~ 10 GHz 10 GHz ~ 18 GHz	$1.4 \times 10^{-2}$ $1.6 \times 10^{-2}$ $2.1 \times 10^{-2}$	Coaxial thermistor mounts / HCT-CS-129-40646
Reflection coefficient		(0 ~ 1) 10 MHz ~ 1 GHz 1 GHz ~ 3 GHz 3 GHz ~ 18 GHz	$3.8 \times 10^{-3}$ $5.3 \times 10^{-3}$ $9.3 \times 10^{-3}$	
Transmission trouble testers Pulse width	40648	1 ns ~ 100 $\mu$ s	$1.4 \times 10^{-2}$	Frequency counters, Oscilloscopes, Artifacts / HCT-CS-261-40648
Pulse amplitude		1 mV ~ 20 V	$6.3 \times 10^{-2}$	
Pulse rate		1 ns ~ 100 $\mu$ s	$5.8 \times 10^{-11}$	
Pulse reflection delay time		1 ns ~ 200 $\mu$ s	$1.5 \times 10^{-2}$	
Impedance		0 $\Omega$ 0.1 $\Omega$ ~ 500 $\Omega$	1.2 m $\Omega$ $1.0 \times 10^{-4}$	
Insertion loss		1 MHz ~ 2.5 GHz	0.32 dB	
Return loss		1 MHz ~ 2.5 GHz	0.51 dB	
RF voltmeters Voltage	40650	(DC) 0 V ~ 400 V	$5.8 \times 10^{-5}$	Meter calibrators, Power sensors / HCT-CS-133-40650
		(DC ~ 100 kHz) 0.1 mV ~ 10 V	$1.6 \times 10^{-4}$	
		(100 kHz ~ 1 GHz) -120 dBm ~ 20 dBm	0.15 dB	
Vector voltmeters Voltage	40651	(DC) 0 V ~ 400 V	$5.8 \times 10^{-5}$	Meter calibrators, Power sensors / HCT-CS-173-40651
		(DC ~ 100 kHz) 0.1 mV ~ 10 V	$1.6 \times 10^{-4}$	
		(100 kHz ~ 1 GHz) -120 dBm ~ 20 dBm	0.15 dB	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Field strength meters	40652			Power sensors, Frequency standards / HCT-CS-200-40652
Frequency		9 kHz ~ 18 GHz	$5.8 \times 10^{-11}$	
Frequency response		9 kHz ~ 4 GHz 4 GHz ~ 18 GHz	0.09 dB 0.15 dB	
Amplitude modulation		100 kHz ~ 18 GHz	$1.2 \times 10^{-2}$	
Frequency modulation	100 kHz ~ 18 GHz	$1.2 \times 10^{-2}$		
AM/FM test sources	40653			Measuring receivers / HCT-CS-250-40653
Output frequency		1 MHz ~ 1 GHz	$6.4 \times 10^{-11}$	
Vestigial FM		50 Hz ~ 3 kHz	$2.0 \times 10^{-2}$	
Vestigial AM		50 Hz ~ 3 kHz	$2.0 \times 10^{-2}$	
Distortion factor	12.5 kHz ~ 400 kHz	$4.0 \times 10^{-4}$		
Dip simulators	40654			Digital multimeters, Oscilloscopes, High voltage probes /HCT-CS-202-40654
Output voltage		1 V	0.65 mV	
DC Output voltage		(1 ~ 10) V	$6.5 \times 10^{-4}$	
		(10 ~ 50) V	$8.2 \times 10^{-5}$	
		(50 ~ 100) V	$1.6 \times 10^{-4}$	
		(100 ~ 150) V	$9.2 \times 10^{-5}$	
		(150 ~ 200) V	$2.4 \times 10^{-2}$	
		(200 ~ 250) V	$1.9 \times 10^{-2}$	
		(250 ~ 300) V	$1.6 \times 10^{-2}$	
		(300 ~ 400) V	$1.4 \times 10^{-2}$	
AC Output voltage		(50 ~ 60) Hz		
		50 V	0.30 V	
		(50 ~ 100) V	$3.4 \times 10^{-3}$	
		(100 ~ 150) V	$3.1 \times 10^{-3}$	
		(150 ~ 200) V	$2.4 \times 10^{-3}$	
		(200 ~ 250) V	$2.1 \times 10^{-3}$	
		(250 ~ 300) V	$1.8 \times 10^{-3}$	
		(300 ~ 400) V	$1.5 \times 10^{-3}$	
Line frequency		(50 ~ 60) Hz	$3.5 \times 10^{-4}$	
Dip & Up Voltage	(0 ~ 12) V			
DC Voltage	0 %			
	0 V	0.36 V		
	(0 ~ 40) %			
	(0 ~ 4.8) V	$7.9 \times 10^{-2}$		
	(40 ~ 70) %			
	(4.8 ~ 8.4) V	$5.0 \times 10^{-2}$		
	(70 ~ 80) %			
	(8.4 ~ 9.6) V	$4.4 \times 10^{-2}$		
	(80 ~ 120) %			
	(9.6 ~ 14.4) V	$3.5 \times 10^{-2}$		

406. Radio frequency measuremet

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Dip & Up Voltage DC Voltage	40654	(12 ~ 25) V		
		0 %		
		0 V	0.36 V	
		(0 ~ 40) %		
		(0 ~ 10) V	4.3×10 <sup>-2</sup>	
		(40 ~ 70) %		
		(10 ~ 17.5) V	3.1×10 <sup>-2</sup>	
		(70 ~ 80) %		
		(17.5 ~ 20) V	2.9×10 <sup>-2</sup>	
		(80 ~ 120) %		
		(20 ~ 30) V	2.6×10 <sup>-2</sup>	
		(25 ~ 50) V		
		0 %		
		0 V	0.36 V	
		(0 ~ 40) %		
		(0 ~ 20) V	2.9×10 <sup>-2</sup>	
		(40 ~ 70) %		
		(20 ~ 35) V	2.5×10 <sup>-2</sup>	
		(70 ~ 80) %		
		(35 ~ 40) V	2.5×10 <sup>-2</sup>	
		(80 ~ 120) %		
		(40 ~ 60) V	2.4×10 <sup>-2</sup>	
		(50 ~ 100) V		
		0 %		
		0 V	0.37 V	
		(0 ~ 40) %		
		(0 ~ 40) V	2.6×10 <sup>-2</sup>	
		(40 ~ 70) %		
		(40 ~ 70) V	2.4×10 <sup>-2</sup>	
		(70 ~ 80) %		
		(70 ~ 80) V	2.4×10 <sup>-2</sup>	
		(80 ~ 120) %		
		(80 ~ 120) V	2.3×10 <sup>-2</sup>	
		(100 ~ 200) V		
		0 %		
		0 V	0.37 V	
(0 ~ 40) %				
(0 ~ 80) V	2.8×10 <sup>-2</sup>			
(40 ~ 70) %				
(80 ~ 140) V	2.4×10 <sup>-2</sup>			
(70 ~ 80) %				
(140 ~ 160) V	2.4×10 <sup>-2</sup>			
(80 ~ 120) %				
(160 ~ 240) V	2.4×10 <sup>-2</sup>			
(200 ~ 300) V				
0 %				
0 V	0.35 V			
(0 ~ 40) %				
(0 ~ 120) V	2.8×10 <sup>-2</sup>			
(40 ~ 70) %				
(120 ~ 210) V	2.5×10 <sup>-2</sup>			
(70 ~ 80) %				
(210 ~ 240) V	2.5×10 <sup>-2</sup>			
(80 ~ 120) %				
(240 ~ 360) V	2.4×10 <sup>-2</sup>			



406. Radio frequency measuremet

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Dip & Up Voltage DC Voltage	40654	(300 ~ 400) V		
		0 %		
		0 V	0.37 V	
		(0 ~ 40) %		
		(0 ~ 160) V	$2.6 \times 10^{-2}$	
		(40 ~ 70) %		
		(160 ~ 280) V	$2.4 \times 10^{-2}$	
		(70 ~ 80) %		
		(280 ~ 320) V	$2.4 \times 10^{-2}$	
		(80 ~ 120) %		
		(320 ~ 480) V	$2.3 \times 10^{-2}$	
		(100 ~ 110) V, (50 ~ 60) Hz		
		0 %		
		0 V	0.37 V	
		(0 ~ 40) %		
		(0 ~ 44) V	$3.3 \times 10^{-2}$	
		(40 ~ 70) %		
		(44 ~ 77) V	$2.7 \times 10^{-2}$	
		(70 ~ 80) %		
		(77 ~ 88) V	$2.6 \times 10^{-2}$	
		(80 ~ 120) %		
		(88 ~ 132) V	$2.4 \times 10^{-2}$	
		(110 ~ 120) V, (50 ~ 60) Hz		
		0 %		
		0 V	0.37 V	
		(0 ~ 40) %		
		(0 ~ 48) V	$3.2 \times 10^{-2}$	
		(40 ~ 70) %		
		(48 ~ 84) V	$2.6 \times 10^{-2}$	
		(70 ~ 80) %		
		(84 ~ 96) V	$2.6 \times 10^{-2}$	
		(80 ~ 120) %		
		(96 ~ 144) V	$2.4 \times 10^{-2}$	
		(120 ~ 220) V, (50 ~ 60) Hz		
		0 %		
		0 V	0.38 V	
		(0 ~ 40) %		
		(0 ~ 88) V	$3.2 \times 10^{-2}$	
		(40 ~ 70) %		
		(88 ~ 154) V	$2.6 \times 10^{-2}$	
		(70 ~ 80) %		
		(154 ~ 176) V	$2.6 \times 10^{-2}$	
		(80 ~ 120) %		
		(176 ~ 264) V	$2.4 \times 10^{-2}$	
		(220 ~ 230) V, (50 ~ 60) Hz		
		0 %		
		0 V	0.38 V	
		(0 ~ 40) %		
(0 ~ 92) V	$3.2 \times 10^{-2}$			
(40 ~ 70) %				
(92 ~ 161) V	$2.6 \times 10^{-2}$			
(70 ~ 80) %				
(161 ~ 184) V	$2.5 \times 10^{-2}$			
(80 ~ 120) %				
(184 ~ 276) V	$2.4 \times 10^{-2}$			



407. Field strength & antenna

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Microwave leakage monitors  Power Density	40701	2.45 GHz (0.01 ~ 3) mW/cm <sup>2</sup>	0.16	Transfer standard probes / HCT-CS-310-40701
Probes  E-field probes           H-field probes Frequency response           Linearity	40702	5 kHz ~ 200 MHz (1 ~ 800) V/m  200 MHz ~ 1 GHz (1 ~ 300) V/m  (1 ~ 18) GHz (1 ~ 200) V/m  (18 ~ 40) GHz (1 ~ 200) V/m  10 Hz ~ 400 kHz (0.16 ~ 40) A/m  400 kHz ~ 220 MHz (0.02 ~ 2.97) A/m  220 MHz ~ 1 GHz (0.02 ~ 1.48) A/m  (50 ~ 60) Hz (0.16 ~ 400) A/m	0.13  0.13  0.13  0.14  0.06  0.14  0.16  0.04	Transfer probes / HCT-CS-262-40702           H-field probes / HCT-CS-311-40702
Dipole antennas  SAR E-field probe Conversion factor  Dipole antenna Antenna factor VSWR Radiation pattern  Bioconical pattern Antenna factor VSWR  Log periodic antenna Antenna factor VSWR	40703	800 MHz ~ 6 GHz  20 MHz ~ 18 GHz 20 MHz ~ 18 GHz 700 MHz ~ 18 GHz  20 MHz ~ 18 GHz (18 GHz ~ 40 GHz) 20 MHz ~ 18 GHz (18 ~ 40) GHz  20 MHz ~ 18 GHz (18 ~ 40) GHz 20 MHz ~ 6 GHz (6 ~ 40) GHz	1.3×10 <sup>-1</sup>  1.1 dB 0.02 1.4 dB  1.2 dB 1.5 dB 0.02 0.24  1.2 dB 1.4 dB 0.02 0.24	SAR calibration system / HCT-CS-106-40703  Network analyzers / HCT-CS-263-40703  Network analyzers / HCT-CS-272-40703  Network analyzers / HCT-CS-273-40703
Loop antennas  Antenna factor	40704	10 Hz ~ 30 MHz 30 MHz ~ 400 MHz	1.3 dB 1.5 dB	Standard loop antennas / HCT-CS-237-40704
Monopole antennas  Antenna factor	40705	1 kHz ~ 30 MHz	1.4 dB	Network analyzers / HCT-CS-238-40705

407. Field strength & antenna

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Horn antennas	40707	200 MHz ~ 18 GHz (18 ~ 40) GHz (40 ~ 110) GHz	0.9 dB	Network analyzers / HCT-CS-264-40707
Antenna factor			1.4 dB	
			1.2 dB	
VSWR		200 MHz ~ 40 GHz (40 ~ 110) GHz	0.02	
		0.03		
Radiation pattern		700 MHz ~ 18 GHz (18 ~ 40 GHz)	1.4 dB	
			1.4 dB	

501. Contact thermometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Temperature generators: ovensm furnaces, isothermal liquid baths, ice-point baths, dry-block calibrators dry-block calibrators Ice-point baths Isothermal liquid baths Furnaces Temperature controlled chambers/ovens	50101	(-80 ~ 500) °C 0 °C (-80 ~ 500) °C (250 ~ 1 100) °C (1 100 ~ 1 600) °C (-80 ~ 250) °C (250 ~ 400) °C	0.05 °C 0.02 °C 0.03 °C 0.9 °C 2.6 °C 0.5 °C 1.0 °C	Standard thermometers,  /HCT-CS-203-50101  /HCT-CS-210-50101  /HCT-CS-211-50101  /HCT-CS-212-50101  /HCT-CS-134-50101
Temperature inducators/recorders/ controllers, temperature calibrators Temperature indicators/recorders /controllers (With Sensor) Thermoeletric Type Resistance Type (Without Sensor) Thermoelectric Type Resistance Type	50102	(-80 ~ 250) °C (250 ~ 500) °C (500 ~ 1 100) °C (-80 ~ 250) °C (250 ~ 500) °C (-80 ~ 1 100) °C (-80 ~ 500) °C	0.4 °C 0.7 °C 1.7 °C 0.03 °C 0.06 °C 0.04 °C 0.03 °C	Standard thermometers,  /HCT-CS-135-50102  /HCT-CS-274-50102  /HCT-CS-137-50102 /HCT-CS-139-50102
Glass thermometers; liquid-in-glass, Beckmann liquid-in-glass	50103	(-80 ~ 250) °C	0.04 °C	Standard thermometers, / HCT-CS-147-50103
Resistance thermometers; SPRT, IPRT, thermistors, etc. IPRT	50104	(-80 ~ 500) °C	0.04 °C	Standard thermometers, / HCT-CS-148-50104
Thermal expansion thermometers; bimetal, gas or liquid type bimetal	50105	(-80 ~ 250) °C (250 ~ 500) °C	0.6 °C 1.4 °C	Standard thermometers, / HCT-CS-149-50105
Thermomecoules: noble metal, base metal, pure metal, special type, etc. Jewelry thermocouple Nonmmetal thermocouple	50106	(0 ~ 250) °C (250 ~ 500) °C (500 ~ 1 100) °C (-80 ~ 250) °C (250 ~ 500) °C (500 ~ 1 100) °C	0.5 °C 0.4 °C 1.0 °C 0.4 °C 0.7 °C 1.8 °C	Standard thermometers, Standard thermocouples /HCT-CS-152-50106  /HCT-CS-151-50106
Temperature transducers	50107	(-80 ~ 250) °C	0.11 °C	Standard thermometers / HCT-CS-170-50107

502. Non contact thermometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Standard radiation thermometers	50204	(-20 ~ 0) °C (0 ~ 50) °C (50 ~ 200) °C (200 ~ 800) °C (800 ~ 1 000) °C	1.0 °C 0.3 °C 0.4 °C 1.4 °C 1.8 °C	Standard radiation thermometers, Blackbody sources HCT-CS-222-50204
Thermal image apparatus	50205	(-20 ~ 0) °C (0 ~ 50) °C (50 ~ 200) °C (200 ~ 800) °C (800 ~ 1 000) °C	1.0 °C 0.3 °C 0.4 °C 1.4 °C 1.8 °C	Standard radiation thermometers, Blackbody sources HCT-CS-286-50205
Blackbody furnaces	50206	(-20 ~ 0) °C (0 ~ 200) °C (200 ~ 400) °C (400 ~ 500) °C (500 ~ 800) °C (800 ~ 1 000) °C	1.0 °C 0.4 °C 0.8 °C 1.0 °C 1.4 °C 1.8 °C	Standard radiation thermometers HCT-CS-333-50206

503. Humidity

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Dew-point hygrometers; chilled mirror, alumina thin film, etc. Alumina thinfilm	50301	(-20 ~ 47) °C D.P.	0.7 °C D.P.	Automatic dewpoint hygrometers /HCT-CS-154-50301
Relative humidity hygrometers; polimer thinfilm, etc.  Hair (Humidity) (Termometry)  Polimer thinfilm (Humidity) (Termometry)	50302	(30 ~ 90) % R.H. (-20 ~ 50) °C  (5 ~ 98) % R.H. (-40 ~ 85) °C	2.9 % R.H. 0.5 °C  2.1 % R.H. 0.4 °C	Audomatic dewpoint hygrometers, Standard thermometers / HCT-CS-153-50302 / HCT-CS-156-50302
Temperature humidity recoders; Hygrothermograph, etc.  (Humidity) (Termometry)	50304	(30 ~ 90) % R.H. (-20 ~ 50) °C	2.8 % R.H. 0.8 °C	Audomatic dewpoint hygrometers / HCT-CS-157-50304
Transducers; dew-point/ relative humidity  Relative humidity	50305	(5 ~ 95) % R.H.	2.6 % R.H.	Audomatic dewpoint hygrometers / HCT-CS-171-50305
Humidity generators; two-pressure, two-temperature, flow mixing humidity generator, constant temperature and humidity chamber, etc.  Flow mixing humidity generator  Constant temperature and humidity chamber Humidity  Thermometry	50306	(5 ~ 98) % R.H.  (10 ~ 98) % R.H.  (-80 ~ 250) °C	2.0 % R.H.  2.5 % R.H. 0.5 °C	Audomatic dewpoint hygrometers / HCT-CS-213-50306  Audomatic dewpoint hygrometers / HCT-CS-182-50306 Temperature indicators / HCT-CS-182-50306

601. Sound in air

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Sound calibrators; Multifunction calibrator	60102	31.5 Hz (31.5 ~ 63) Hz (63 ~ 8 000) Hz (8 000 ~ 12 500) Hz	0.14 dB 0.12 dB 0.11 dB 0.13 dB	Autostic calibrators /HCT-CS-196-60102
Pistonphone, Sound calibraor		250 Hz 1 000 Hz	0.11 dB 0.11 dB	Autostic calibrators /HCT-CS-196-60102
Microphones Pistonphone	60104	250 Hz	0.14 dB	Pistonphone / HCT-CS-194-60104
3-port coupler		20 Hz (20 ~ 25) Hz (25 ~ 31.5) Hz (31.5 ~ 50) Hz (50 ~ 63) Hz (63 ~ 1 250) Hz (1 250 ~ 6 300) Hz (6 300 ~ 8 000) Hz (8 000 ~ 10 000) Hz (10 000 ~ 12 500) Hz (12 500 ~ 16 000) Hz	0.16 dB 0.14 dB 0.13 dB 0.12 dB 0.10 dB 0.09 dB 0.10 dB 0.24 dB 0.26 dB 0.27 dB 0.36 dB	3-port Coupler, Microphone / HCT-CS-293-60104
Sound level meters Multifunction calibrator	60106	(63 ~ 4 000) Hz (4 000 ~ 8 000) Hz	0.3 dB 0.4 dB	ACOUSTIC CALIBRATOR / HCT-CS-158-60106
3-port coupler		125 Hz (125 ~ 2 500) Hz (2 500 ~ 8 000) Hz	0.4 dB 0.2 dB 0.3 dB	3-port Coupler / HCT-CS-172-60106



603. Vibration

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Vibration calibrators Vibration calibrator	60301	(20 ~ 1 250) Hz	$1.9 \times 10^{-2}$	Standard accelerometer / HCT-CS-219-60301
Vibration transducers Vibration transducer	60302	1 Hz	$1.9 \times 10^{-2}$	Standard accelerometers / HCT-CS-220-60302
		(1 ~ 5) Hz	$2.0 \times 10^{-2}$	
		(5 ~ 8) Hz	$1.9 \times 10^{-2}$	
		(8 ~ 20) Hz	$1.2 \times 10^{-2}$	
		(20 ~ 630) Hz	$1.1 \times 10^{-2}$	
		(630 ~ 1 250) Hz	$1.2 \times 10^{-2}$	
		(1 250 ~ 2 500) Hz	$1.7 \times 10^{-2}$	
		(2 500 ~ 5 000) Hz	$2.1 \times 10^{-2}$	
		(5 000 ~ 10 000) Hz	$3.0 \times 10^{-2}$	
		(10 000 ~ 15 000) Hz	$4.1 \times 10^{-2}$	
		(15 000 ~ 20 000) Hz	$5.3 \times 10^{-2}$	
Vibration transducer(Shock)		(0.1 ~ 11) ms		Standard accelerometers / HCT-CS-291-60302
		200 m/s <sup>2</sup>	$3.6 \times 10^{-2}$	
		(200 ~ 500) m/s <sup>2</sup>	$3.2 \times 10^{-2}$	
	(500 ~ 1 000) m/s <sup>2</sup>	$3.1 \times 10^{-2}$		
	(1 000 ~ 5 000) m/s <sup>2</sup>	$3.6 \times 10^{-2}$		
	(5 000 ~ 20 000) m/s <sup>2</sup>	$3.3 \times 10^{-2}$		
	(20 000 ~ 100 000) m/s <sup>2</sup>	$4.0 \times 10^{-2}$		

603. Vibration

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Vibration measuring instruments	60303			
Vibration measuring instrument				
Acceleration		10 Hz	$1.7 \times 10^{-2}$	Standard accelerometers / HCT-CS-221-60303
		(10 ~ 40) Hz	$1.8 \times 10^{-2}$	
		(40 ~ 100) Hz	$1.7 \times 10^{-2}$	
		(100 ~ 630) Hz	$1.8 \times 10^{-2}$	
		(630 ~ 1 250) Hz	$1.9 \times 10^{-2}$	
		(1 250 ~ 2 500) Hz	$2.1 \times 10^{-2}$	
Velocity		(10 ~ 40) Hz	$1.8 \times 10^{-2}$	
		(40 ~ 160) Hz	$1.7 \times 10^{-2}$	
	(160 ~ 630) Hz	$1.8 \times 10^{-2}$		
	(630 ~ 1 250) Hz	$2.1 \times 10^{-2}$		
	(1 250 ~ 2 500) Hz	$2.7 \times 10^{-2}$		
Displacement	(10 ~ 160) Hz	$1.6 \times 10^{-2}$		
	(160 ~ 315) Hz	$2.2 \times 10^{-2}$		
Vibration measuring instrument(Shock)		200 m/s <sup>2</sup>	$4.9 \times 10^{-2}$	Standard accelerometers / HCT-CS-292-60303
	(200 ~ 500) m/s <sup>2</sup>	$3.5 \times 10^{-2}$		
	(500 ~ 1 000) m/s <sup>2</sup>	$3.3 \times 10^{-2}$		
	(1 000 ~ 1 500) m/s <sup>2</sup>	$3.2 \times 10^{-2}$		
	(1 500 ~ 2 000) m/s <sup>2</sup>	$3.7 \times 10^{-2}$		

701. Photometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Illuminance meters	70101	0.5 lx (0.5 ~ 10) lx (10 ~ 2 000) lx (2 000 ~ 11 000) lx	$3.4 \times 10^{-2}$ $2.9 \times 10^{-2}$ $2.8 \times 10^{-2}$ $2.9 \times 10^{-2}$	Reference Illuminance meters /HCT-CS-159-70101
Luminance meters Luminance	70102	(2 ~ 10) cd/m <sup>2</sup> (10 ~ 100) cd/m <sup>2</sup> (100 ~ 1 000) cd/m <sup>2</sup> (1 000 ~ 13 000) cd/m <sup>2</sup>	$2.4 \times 10^{-2}$ $1.7 \times 10^{-2}$ $1.6 \times 10^{-2}$ $1.8 \times 10^{-2}$	Luminance standard sources /HCT-CS-316-70102
Total luminous flux meters Total luminous flux	70103	(70 ~ 20 000) lm	$2.3 \times 10^{-2}$	Total luminous flux standard lamp / HCT-CS-296-70103
Luminous intensity meters Luminous intensity	70104	(8.22 ~ 2 950) cd	$3.3 \times 10^{-2}$	Luminous intensity standard lamp / HCT-CS-297-70104

702. Property of detectors & sources

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color Temperature Meters Color Temperature	70202	(2 715 ~ 3 058) K	22 K	Color Temperature Standard lamps / HCT-CS-298-70202
Color temperature standard lamps Color temperature	70203	2 856 K	22 K	Color temperature standard lamps, Spectral irradiance meters / HCT-CS-299-70203
Calorimeters; source color Chromaticity coordinates (CIE 1931)  Luminance  Illuminance	70204	x,y : (0.01 ~ 0.9) Tungsten light sources 2 856 K x y Red x y Green x y Blue x y White x y  (2 ~ 10) cd/m <sup>2</sup> (10 ~ 100) cd/m <sup>2</sup> (100 ~ 1 000) cd/m <sup>2</sup> (1 000 ~ 13 000) cd/m <sup>2</sup>  1 lx (1 ~ 2 000) lx	0.003 0.003  0.005 0.004  0.006 0.005  0.004 0.004  2.4×10 <sup>-2</sup> 1.7×10 <sup>-2</sup> 1.6×10 <sup>-2</sup> 1.8×10 <sup>-2</sup>  3.0×10 <sup>-2</sup> 2.8×10 <sup>-2</sup>	Luminance standard sources, Color temperature standard lamps, Color filters / HCT-CS-317-70204
Total luminance flux standard lamps Total luminance flux	70209	(70 ~ 20 000) lm	3.6×10 <sup>-2</sup>	Total luminous flux standard lamps, Total luminous flux meters / HCT-CS-300-70209
Display color analyzers; luminance, , chromaticity, white balance, etc. Chromaticity( x, y ) Chromaticity coordinates (CIE 1931)  Luminance	70213	x,y : (0.01 ~ 0.9) 텡스텐광원 2 856 K x y Red x y Green x y Blue x y White x y  (2 ~ 10) cd/m <sup>2</sup> (10 ~ 100) cd/m <sup>2</sup> (100 ~ 1 000) cd/m <sup>2</sup> (1 000 ~ 13 000) cd/m <sup>2</sup>	0.003 0.003  0.005 0.004  0.006 0.005  0.004 0.004  2.4×10 <sup>-2</sup> 1.7×10 <sup>-2</sup> 1.6×10 <sup>-2</sup> 1.8×10 <sup>-2</sup>	Luminance standard sources, Color filters / HCT-CS-318-70213

702. Property of detectors & sources

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Luminous intensity standard lamps  Luminous intensity	70214	(8.22 ~ 2 950) cd	$3.3 \times 10^{-2}$	Luminous intensity standard lamps, Luminous intensity meters / HCT-CS-301-70214
Spectral irradiance standard lamps Spectral irradiance           Illuminance  Color temperature  Chromaticity coordinates (CIE 1931)	70215	250 nm (250 ~ 255) nm (255 ~ 265) nm (265 ~ 275) nm (275 ~ 285) nm (285 ~ 295) nm (295 ~ 305) nm (305 ~ 340) nm (340 ~ 370) nm (370 ~ 400) nm (400 ~ 475) nm (475 ~ 1 020) nm  (6 833 ~ 7 224 ) lx  (3 014 ~ 3 061) K  x (0.431 ~ 0.437) y (0.401 ~ 0.407)	$7.6 \times 10^{-2}$ $6.6 \times 10^{-2}$ $6.2 \times 10^{-2}$ $5.7 \times 10^{-2}$ $5.3 \times 10^{-2}$ $5.0 \times 10^{-2}$ $4.4 \times 10^{-2}$ $4.0 \times 10^{-2}$ $3.5 \times 10^{-2}$ $2.9 \times 10^{-2}$ $2.6 \times 10^{-2}$ $2.2 \times 10^{-2}$  $2.8 \times 10^{-2}$  22 K  0.003 0.003	Spectral irradiance standard lamps, Spectral irradiance meters / HCT-CS-302-70215
Total spectral radiant flux standard lamps Total spectral radiant      Total luminous flux  Color temperature  Chromaticity coordinates (CIE 1931)	70216	350 nm (350 ~ 365) nm (365 ~ 380) nm (380 ~ 410) nm (410 ~ 480) nm (480 ~ 850) nm  (2 130 ~ 2 208) nm  (2 715 ~ 2 758) K  x (0.454 ~ 0.460) y (0.407 ~ 0.413)	$6.2 \times 10^{-2}$ $5.1 \times 10^{-2}$ $4.3 \times 10^{-2}$ $3.6 \times 10^{-2}$ $2.9 \times 10^{-2}$ $2.6 \times 10^{-2}$  $2.6 \times 10^{-2}$  22 K  0.003 0.003	Total spectral radiant flux standard lamps Total spectral radiant flux meters / HCT-CS-303-70216
Luminance standard sources Luminance  Chromaticity coordinates (CIE 1931)	70217	(2 ~ 13 000) cd/m <sup>2</sup>  x,y : (0.01 ~ 0.9) 텡스텐광원 2 856 K x y Red x y Green x y Blue x y White x y	$2.4 \times 10^{-2}$  0.003 0.003  0.005 0.004  0.006 0.005  0.004 0.004  0.004 0.004	Spectral radiance meters, Colorimeters; source color / HCT-CS-319-70217

702. Property of detectors & sources

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Spectral radiance standard sources Spectral radiance	70218	380 nm (380 ~ 395) nm (395 ~ 410) nm (410 ~ 430) nm (430 ~ 450) nm (450 ~ 475) nm (475 ~ 505) nm (505 ~ 1 005) nm (1 005 ~ 1 040) nm	$4.3 \times 10^{-2}$ $3.8 \times 10^{-2}$ $3.4 \times 10^{-2}$ $3.1 \times 10^{-2}$ $2.9 \times 10^{-2}$ $2.5 \times 10^{-2}$ $2.2 \times 10^{-2}$ $2.0 \times 10^{-2}$ $2.1 \times 10^{-2}$	Spectral radiance standard sources, Spectral radiance meters / HCT-CS-320-70218
UV irradiance meters	70219	365 nm $60 \mu\text{W}/\text{cm}^2 \sim 200 \text{mW}/\text{cm}^2$  405 nm $60 \mu\text{W}/\text{cm}^2 \sim 70 \text{mW}/\text{cm}^2$	$4.8 \times 10^{-2}$   $4.8 \times 10^{-2}$	UV Sensor /HCT-CS-159-70219
Spectral irradiance meters Wavelength  Spectral irradiance   Illuminance  Color temperature  Chromaticity coordinates (CIE 1931)	70220	(350 ~ 850) nm  250 nm (250 ~ 255) nm (255 ~ 275) nm (275 ~ 300) nm (300 ~ 340) nm (340 ~ 455) nm (455 ~ 565) nm (565 ~ 1 020) nm  (6 833 ~ 7 224 ) lx  (3 014 ~ 3 061) K  x (0.431 ~ 0.437) y (0.401 ~ 0.407)	0.51 nm  $6.7 \times 10^{-2}$ $6.1 \times 10^{-2}$ $5.2 \times 10^{-2}$ $4.4 \times 10^{-2}$ $3.7 \times 10^{-2}$ $2.9 \times 10^{-2}$ $2.2 \times 10^{-2}$ $2.0 \times 10^{-2}$  $2.9 \times 10^{-2}$  22 K  0.003 0.003	Spectral irradiance standard lamps / HCT-CS-304-70220
Total spectral radiant flux meters Wavelength  Total spectral radiant flux   Total luminous flux meters  Color temperature  Chromaticity coordinates (CIE 1931)	70221	(350 ~ 850) nm  350 nm (350 ~ 355) nm (355 ~ 370) nm (370 ~ 390) nm (390 ~ 425) nm (425 ~ 460) nm (460 ~ 850) nm  (2 130 ~ 2 208) lm  (2 715 ~ 2 876) K  x (0.454 ~ 0.460) y (0.407 ~ 0.413)	0.51 nm  $5.0 \times 10^{-2}$ $4.0 \times 10^{-2}$ $3.7 \times 10^{-2}$ $3.1 \times 10^{-2}$ $2.4 \times 10^{-2}$ $2.0 \times 10^{-2}$ $1.9 \times 10^{-2}$  $2.3 \times 10^{-2}$  22 K  0.004 0.004	Total spectral radiant flux standard lamps / HCT-CS-305-70221

702. Property of detectors & sources

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.		
Spectral radiance meters	70222	(400 ~ 765) nm	0.51 nm	Spectral radiance meters, Luminance standard sources / HCT-CS-321-70222		
Wavelength						
Spectral radiance					(380) nm	$4.1 \times 10^{-2}$
					(380 ~ 395) nm	$3.6 \times 10^{-2}$
					(395 ~ 410) nm	$3.3 \times 10^{-2}$
					(410 ~ 430) nm	$3.0 \times 10^{-2}$
					(430 ~ 445) nm	$2.7 \times 10^{-2}$
					(445 ~ 465) nm	$2.4 \times 10^{-2}$
					(465 ~ 485) nm	$2.1 \times 10^{-2}$
					(485 ~ 1 040) nm	$1.8 \times 10^{-2}$
Luminance	(2 ~ 10) cd/m <sup>2</sup>	$2.4 \times 10^{-2}$				
	(10 ~ 100) cd/m <sup>2</sup>	$1.7 \times 10^{-2}$				
	(100 ~ 1 000) cd/m <sup>2</sup>	$1.6 \times 10^{-2}$				
	(1 000 ~ 13 000) cd/m <sup>2</sup>	$1.8 \times 10^{-2}$				
Color temperature	(2 841 ~ 2 881) K	22 K				
Chromaticity coordinates (CIE 1931)	x : (0.447 ~ 0.451)	0.003				
	y : (0.409 ~ 0.413)	0.003				
Spectral radiant intensity meters	70223	(350 ~ 850) nm	0.51 nm	Luminous intensity standard lamps / HCT-CS-306-70223		
Wavelength						
Spectral radiant intensity					250 nm	$6.7 \times 10^{-2}$
					(250 ~ 255) nm	$6.1 \times 10^{-2}$
					(255 ~ 275) nm	$5.2 \times 10^{-2}$
					(275 ~ 300) nm	$4.4 \times 10^{-2}$
					(300 ~ 340) nm	$3.7 \times 10^{-2}$
					(340 ~ 455) nm	$2.9 \times 10^{-2}$
					(455 ~ 565) nm	$2.2 \times 10^{-2}$
					(565 ~ 1 020) nm	$2.0 \times 10^{-2}$

703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters;object color (Including Specular Component Standard Illuminant A, C, D65 Standard Observe : 2° , 10° )	70301			Color standard tiles / HCT-CS-354-70301
White	X		$0.81 \times 10^{-2}$	
	Y		$0.81 \times 10^{-2}$	
	Z		$0.82 \times 10^{-2}$	
M. Grey	X		$0.83 \times 10^{-2}$	
	Y		$0.82 \times 10^{-2}$	
	Z		$0.85 \times 10^{-2}$	
D. Grey	X		$0.93 \times 10^{-2}$	
	Y		$0.88 \times 10^{-2}$	
	Z		$0.93 \times 10^{-2}$	
Red	X		$0.98 \times 10^{-2}$	
	Y		$0.93 \times 10^{-2}$	
	Z		$1.08 \times 10^{-2}$	
Orange	X		$0.86 \times 10^{-2}$	
	Y		$0.91 \times 10^{-2}$	
	Z		$0.92 \times 10^{-2}$	
Yellow	X		$0.82 \times 10^{-2}$	
	Y		$0.84 \times 10^{-2}$	
	Z		$1.19 \times 10^{-2}$	
Green	X		$0.93 \times 10^{-2}$	
	Y		$0.95 \times 10^{-2}$	
	Z		$1.03 \times 10^{-2}$	
D. Blue	X		$0.99 \times 10^{-2}$	
	Y		$0.94 \times 10^{-2}$	
	Z		$1.15 \times 10^{-2}$	
(Excluding Specular Component Standard Illuminant A, C, D65 Standard Observe : 2° , 10° )				
White	X		$0.81 \times 10^{-2}$	
	Y		$0.81 \times 10^{-2}$	
	Z		$0.82 \times 10^{-2}$	
M. Grey	X		$0.83 \times 10^{-2}$	
	Y		$0.83 \times 10^{-2}$	
	Z		$0.88 \times 10^{-2}$	
D. Grey	X		$1.01 \times 10^{-2}$	
	Y		$0.99 \times 10^{-2}$	
	Z		$1.12 \times 10^{-2}$	
Red	X		$1.05 \times 10^{-2}$	
	Y		$1.04 \times 10^{-2}$	
	Z		$1.43 \times 10^{-2}$	
Orange	X		$0.87 \times 10^{-2}$	
	Y		$0.92 \times 10^{-2}$	
	Z		$1.04 \times 10^{-2}$	
Yellow	X		$0.82 \times 10^{-2}$	
	Y		$0.84 \times 10^{-2}$	
	Z		$1.42 \times 10^{-2}$	
Green	X		$0.90 \times 10^{-2}$	
	Y		$0.88 \times 10^{-2}$	
	Z		$1.13 \times 10^{-2}$	



703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
D. Blue	70301	X	$1.35 \times 10^{-2}$	
		Y	$1.49 \times 10^{-2}$	
		Z	$1.45 \times 10^{-2}$	
(Including Specular Component Standard Illuminant A, C, D65 Standard Observe : $2^\circ$ , $10^\circ$ )				
White		x	0.000 5	
		y	0.000 5	
M. Grey		x	0.000 5	
		y	0.000 5	
D. Grey		x	0.000 5	
		y	0.000 5	
Red		x	0.000 7	
		y	0.000 5	
Orange		x	0.000 5	
		y	0.000 5	
Yellow		x	0.000 5	
		y	0.000 5	
Green		x	0.000 5	
		y	0.000 5	
D. Blue		x	0.000 5	
		y	0.000 5	
(Excluding Specular Component Standard Illuminant A, C, D65 Standard Observe : $2^\circ$ , $10^\circ$ )				
White		x	0.000 5	
		y	0.000 5	
M. Grey		x	0.000 5	
		y	0.000 5	
D. Grey		x	0.000 5	
		y	0.000 5	
Red		x	0.000 6	
		y	0.000 5	
Orange		x	0.000 5	
		y	0.000 5	
Yellow		x	0.000 5	
		y	0.000 5	
Green		x	0.000 5	
		y	0.000 5	
D. Blue		x	0.000 7	
		y	0.000 7	
Reflectance meters	70321	(1.332 97 ~ 1.490 78)nD	0.000 06 nD	Reflectance meters / HCT-CS-355-70321

704. Fiber optics

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Broadband light sources Wavelength output  Optical power output	70402	1 310 nm 1 550 nm  1 310 nm (-60 ~ 0) dBm  1 550 nm (-60 ~ 0) dBm	$5.4 \times 10^{-7}$ $4.6 \times 10^{-7}$  0.08 dB  0.08 dB	Wavelength meters, Optical power meters / HCT-CS-266-70402
Optical attenuators Optical attenuation	70410	1 310 nm, 1 550 nm (-60 ~ 0) dB	0.04 dB	Optical power meters / HCT-CS-267-70410
Optical loss testers Wavelength output  Optical power input  Linearity measure	70413	1 310 nm 1 550 nm  1 310 nm (-60 ~ 0) dBm  1 550 nm (-60 ~ 0) dBm  1 310 nm, 1 550 nm (-60 ~ 0) dB	$5.4 \times 10^{-7}$ $4.6 \times 10^{-7}$  0.08 dB  0.08 dB  0.04 dB	Wavelength meters, Optical power meters / HCT-CS-280-70413
Optical multimeters Optical power input  Linearity measure	70415	1 310 nm (-60 ~ 0) dBm  1 550 nm (-60 ~ 0) dBm  1 310 nm, 1 550 nm (-60 ~ 0) dB	0.08 dB  0.08 dB  0.04 dB	Optical power meters / HCT-CS-268-70415
Optical spectrum analyzers Wavelength output  Resolution measure  Optical power output  Linearity measure	70417	1 310 nm, 1 550 nm  1 310 nm, 1 550 nm RBW (0.1 ~ 1) nm  1 310 nm (-60 ~ 0) dBm  1 550 nm (-60 ~ 0) dBm  1 310 nm, 1 550 nm (-60 ~ 0) dB	0.084 nm  0.084 nm  0.08 dB  0.08 dB  0.04 dB	Wavelength meters, Optical power meters / HCT-CS-269-70417
Optical time domain reflectometers; OTDR  Wavelength  Length	70418	1 310 nm 1 550 nm  (1 310 nm) 3 km 13 km	0.36 nm 0.36 nm  0.1 m 0.34 m	Optical spectrum analyzers, Standard CRM / HCT-CS-270-70418

704. Fiber optics

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Length	70418	(1 550 nm)		
		3 km	0.1 m	
		13 km	0.34 m	
Return loss		(1 310 nm)		
		3 km	0.07 dB	
		13 km	0.18 dB	
		(1 550 nm)		
		3 km	0.07 dB	
		13 km	0.07 dB	
ASE light sources	70430			Wavelength meters, Optical power meters / HCT-CS-281-70430
Wavelength output		1 310 nm	$5.4 \times 10^{-7}$	
		1 550 nm	$4.6 \times 10^{-7}$	
Optical power output		1 310 nm (-60 ~ 0) dBm	0.08 dB	
		1 550 nm (-60 ~ 0) dBm	0.08 dB	
Optical power stabilized lasers and LDs	70433			Wavelength meters, Optical power meters / HCT-CS-271-70433
Wavelength output		1 310 nm	$5.4 \times 10^{-7}$	
		1 550 nm	$4.6 \times 10^{-7}$	
Optical power output		1 310 nm (-60 ~ 0) dBm	0.08 dB	
		1 550 nm (-60 ~ 0) dBm	0.08 dB	

901. Chemical analysis

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Breath Alcohol Analyzer Dry process	90101	(0.000 ~ 0.030) % BAC	$3.2 \times 10^{-2}$	CRM / HCT-CS-358-90101
		(0.030 ~ 0.080) % BAC	$2.1 \times 10^{-2}$	
		(0.080 ~ 0.450) % BAC	$1.0 \times 10^{-2}$	
Wet process		(0.000 ~ 0.020) % BAC	$4.0 \times 10^{-2}$	CRM / HCT-CS-358-90101
	(0.020 ~ 0.030) % BAC	$2.8 \times 10^{-2}$		
	(0.030 ~ 0.080) % BAC	$1.6 \times 10^{-2}$		
	(0.080 ~ 0.400) % BAC	$1.3 \times 10^{-2}$		
Environmental air quality monitoring instruments	90102			CRM / HCT-CS-346-90102
Oxygen(O <sub>2</sub> )		0 cmol/mol ~ 22.0 cmol/mol	$3.0 \times 10^{-2}$	
Carbon monoxide(CO)		0 μmol/mol ~ 105 μmol/mol	$2.6 \times 10^{-2}$	
Sulfer dioxide(SO <sub>2</sub> )		0 μmol/mol ~ 110 μmol/mol	$4.9 \times 10^{-2}$	
Nitrogen monoxide(NO)	0 μmol/mol ~ 110 μmol/mol	$2.0 \times 10^{-2}$		
Gas analyzers	90103			CRM / HCT-CS-164-90103
Oxygen(O <sub>2</sub> )		0 cmol/mol ~ 22.0 cmol/mol	$3.0 \times 10^{-2}$	
Carbon monoxide(CO)		0 μmol/mol ~ 105 μmol/mol	$2.6 \times 10^{-2}$	
Methane(CH <sub>4</sub> )		0 cmol/mol ~ 2.2 cmol/mol	$2.2 \times 10^{-2}$	
Carbon dioxide(CO <sub>2</sub> )		0 cmol/mol ~ 10.5 cmol/mol	$2.2 \times 10^{-2}$	
Hydrogen sulfide(H <sub>2</sub> S)		0 μmol/mol ~ 53 μmol/mol	$3.5 \times 10^{-2}$	
Sulfer dioxide(SO <sub>2</sub> )		0 μmol/mol ~ 110 μmol/mol	$4.9 \times 10^{-2}$	
Hydrogen chloride(HCl)		0 μmol/mol ~ 53 μmol/mol	$5.0 \times 10^{-2}$	
Nitrogen monoxide(NO)		0 μmol/mol ~ 110 μmol/mol	$2.1 \times 10^{-2}$	
Hydrogen(H <sub>2</sub> )	0 cmol/mol ~ 2.2 cmol/mol	$4.7 \times 10^{-2}$		
Exhaust gas test instruments	90104			CRM / HCT-CS-347-90104
Oxygen(O <sub>2</sub> )		0.3 cmol/mol ~ 1.0 cmol/mol	$3.0 \times 10^{-2}$	
Carbon monoxide(CO)		0.3 cmol/mol ~ 5.0 cmol/mol	$2.2 \times 10^{-2}$	
Carbon dioxide(CO <sub>2</sub> )		5.0 cmol/mol ~ 10.5 cmol/mol	$2.2 \times 10^{-2}$	
Nitrogen monoxide(NO)	500 μmol/mol ~ 1 000 μmol/mol	$2.0 \times 10^{-2}$		